Electronic Detection of DNA Adsorption and Hybridization

Ulrich Bockelmann

Laboratoire de Nanobiophysique
UMR Gulliver CNRS-ESPCI 7083
10 rue Vauquelin, 75005 Paris (France)
`ulrich.bockelmann@espci.fr`

Abstract. We consider a transistor based approach for DNA detection. Immobilization of DNA at different positions of poly(L-lysine) coated field effect transistor arrays is achieved by deposition with a microspotting device or specific hybridization between complementary oligonucleotide sequences. The current voltage characteristics of the transistors are measured with the sample surface immersed in aqueous solution. The chapter provides a brief overview of our experimental technique and of its applications to the detection of DNA adsorption and hybridization.

1 Introduction

The large majority of nucleic acid analysis techniques presently rely on detection schemes which involve fluorescence, radioactivity or enzyme based labelling [1]. In many cases the labelling steps are expensive, time-consuming and error-prone and the corresponding equipment is expensive. These disadvantages are major motivations for research on alternative detection methods [2–4].

After early work by P. Bergveld [5], several papers addressed the detection of biomolecules binding on field effect transistor (FET) structures [6–8]. DNA is a strong negatively charged polyelectrolyte in aqueous solutions. Capacitive impedance and FET detection of DNA have been achieved over the last years [9–14].

We use a differential technique based on dc measurements with arrays of integrated FET devices [12, 15, 16]. Our approach allows for efficient miniaturisation and parallelisation through on-chip integration of the detection and readout circuits [17]. It involves adsorption of the biomolecules to part of the array (achieved by local deposition of sub-nanoliter volumes of solutions with a microspotting device) and measurement of the transistors (with the surface of the microchip covered by an aqueous solution).

The chapter is organized as follows. The experimental techniques are presented in Sect. 2. Electronic detection of DNA immobilized by direct adsorption on poly(L-lysine) coated FET arrays is considered in Sect. 3. Sect. 4 is devoted to the detection of specific hybridization between DNA oligonucleotides. Salt effects are briefly considered in Sect. 5. Sect. 6 contains the concluding part of the chapter.

Please use the following format when citing this chapter:

Bockelmann, U., 2007, in IFIP International Federation for Information Processing, Volume 249, VLSI-SoC: Research Trends in VLSI and Systems on Chip, eds. De Micheli, G., Mir, S., Reis, R., (Boston: Springer), pp. 55–67

2 Experimental techniques

2.1 Fabrication of the FET arrays

Arrays of silicon p-channel field effect transistors have been used in this work. The samples were fabricated in the clean room facility of the Max Planck Institut für Biochemie, Martinsried, Germany. The process is based on standard silicon microtechnology and is described in [18, 19]. The active areas of the individual transistors carry a 10 nm thick SiO_2 oxide without metal gate. Typically, 96 FETs are linearly arranged with a period of 20 μm , the active surface area of an individual FET amounts to 36 or 100 μm^2 . The drain of each transistor is individually connected, while the whole array shares a common source contact. On the silicon chip these connexions are realized by boron p^+ implantation. The chips are mounted on ceramics sockets, wire bonded and finally a plastic well is glued on the surface to protect the wiring during the subsequent manipulations where the surface is incubated with liquids.

2.2 Surface preparation

Prior to DNA immobilization, we perform the following global treatment of the SiO_2 surface. Incubation in sulfochromic acid, rinsing with H_2O , incubation in a NaOH /ethanol solution, H_2O rinsing and drying with air. Afterwards, the FET array is incubated in a poly(L-lysine) dilution, again followed by H_2O rinsing and drying with air. Immobilization of DNA by a poly(L-lysine) layer is widely used in the field of DNA microarrays, although the resulting attachment is less stable than the one obtained by covalent immobilization strategies. DNA immobilization with poly(L-lysine) is relatively simple and allows us to use a FET array several times. In fact, we can clean the SiO_2 surface after use and repeat DNA immobilization and detection.

2.3 Local deposition of DNA

A piezo spotting device equipped with a custom microscopic imaging system is used to deposit DNA solutions on the FET array, as shown in Figure 1. Typically spot diameters of 50-200 μm are used, corresponding to volumina in the 0.1-1 nl range.

2.4 Electronic measurements

For the electronic measurements the surface of the chip is immersed in aqueous electrolyte solution, together with a reference electrode. We measure the drain current I_D of each transistor as a function of a dc voltage U_{SD} applied between source and drain and a dc voltage U_{SE} applied between source and the reference electrode. These 2D characteristics are recorded at room temperature. An analog amplifier circuit is used to bias the FETs and to measure the drain current I_D , as shown in Figure 2. The individual FETs of the array are multiplexed by a

Fig. 1. Microscopic image of a transistor array with two local deposits. The spots have diameters of about $200\ \mu\text{m}$ and each one covers about 10 FETs of the linear array. The active regions of the individual transistors appear as white squares in the image. The gray lines in the bottom part are the individual drain connexions. The homogeneous region in the upper part is the common source contact.

switch unit, the latter being computer controlled by a digital I/O board. Two dc voltages $\tilde{U}_{SE}$, $\tilde{U}_{SD}$ are generated by 16 bit D/A converters of a multifunction I/O board and are lowpass filtered in the input stage of the amplifier. The drain current is converted to a voltage that is measured by a 16 bit A/D converter of the I/O board. The acquisition of the bidimensional characteristics $I_D(\tilde{U}_{SD}, \tilde{U}_S)$ for all FETs is computer controlled.

Depending on the design of the FET array device the implanted on-chip connexions to the individual drains and the common source may lead to non-negligible serial resistances. In this case, we consider the lateral geometry of the connexions and derive the serial resistances of the drain lead R_D and the source lead R_S for each FET. The transformation

$$\begin{aligned} U_{SE} &= \tilde{U}_{SE} - R_S I_D \\ U_{SD} &= \tilde{U}_{SD} - (R_D + R_S) I_D \end{aligned}$$

finally gives the intrinsic characteristics $I_D(U_{SD}, U_{SE})$. To study shifts in U_{SE} we convert the $I_D(U_{SD}, U_{SE})$ characteristics, to $U_{SE}(I_D, U_{SD})$ by numerical interpolation.

2.5 Microfluorescence

For comparison with the electronic signals we use microscopic fluorescence measurements. The corresponding optical setup is schematically presented in Figure 3. The beam of a red (632.8 nm) HeNe cw laser or a green (532 nm) solid state cw laser is expanded, introduced in the illumination path of an upright microscope and focused with a microscope objective to a spot of adjustable diameter ($0.5\text{-}10\ \mu\text{m}$). Fluorescence is collected by the same objective and passes two

Fig. 3. Dual color fluorescence setup for imaging the transistor array surfaces.

is presented. The adsorption of DNA gives rise to negative shifts ΔU_{SE} , while for the transistors below the reference spot no significant shift relative to the baseline is measured.

Fig. 4. Electronic detection of DNA adsorbed on a poly(L-lysine) coated array of 96 FETs linearly arranged with a period of 20 μm . Four local deposits of a DNA oligonucleotide solution are performed and a H_2O spot is added for comparison. The positions of these spots are indicated by arrows. Each data point correspond to one transistor of the array (bottom axis). The quantity ΔU_{SE} gives the shift of the current/voltage characteristics at a given (I_D , U_{SD}) working point, observed between two measurements. The first measurement is done prior to the deposition, the second one immediately afterwards, in both cases the sample surface was covered by an aqueous electrolyte solution (0.1 mM KNO_3). The DNA spots induced negative shifts ΔU_{SE} of about 60 mV, while no corresponding signal is observed on the reference spot.

In an earlier publication [12], we presented a comprehensive investigation of the electronic detection of poly(L-lysine) and DNA. Adsorption of DNA on poly(L-lysine) coated SiO_2 induces negative shifts in U_{SE} , while adsorption of poly(L-lysine) on SiO_2 leads to positive shifts. This can be attributed to the fact that we use p-channel FET arrays and that DNA carries negative charge in aqueous solution around pH 7, opposite to the positively charged poly(L-lysine). The electronic signals induced by adsorption of these charged polymers were studied as a function of electrolyte salt and polymer concentrations and an analytical model which accounts for screening of the adsorbed charge by mobile ions was

developed. The microfluorescence setup described above was used in combination with Cy3 and/or Cy5 modified oligonucleotides to quantify the local DNA adsorption. Non-modified, Cy3-modified and Cy5-modified oligonucleotides are electronically detected without noticeable differences. It is also possible to detect the adsorption of double stranded DNA. This was shown with double stranded molecules synthesized in-vitro by PCR (polymerase chain reaction).

In another study, we showed that the FET based measurement is compatible with enzymatic technology and complex DNA samples [15]. We combined the electronic detection with an allele-specific polymerase chain reaction and thus detected a single-basepair mutation in genomic DNA. The approach was applied to test human DNA for the 35delG mutation, a frequent mutation related to prelingual nonsyndromic deafness.

4 Detecting DNA hybridization

Hybridization involves the specific interaction between complementary base sequences and the transition from two single stranded to one double stranded molecule. In general, the single stranded molecule carries a different effective charge than the double stranded one [20, 21]. Therefore hybridization between a surface bound probe molecule and a target molecule from solution can change the electrostatic potential of the SiO_2 /electrolyte interface and it is conceivable to electronically detect this specific interaction with our FET arrays. Specific recognition of target sequences by hybridization to surface bound probe DNA is the basis of microarray based DNA analysis, a rapidly developing technology with broad applications in research and diagnostics.

Fig. 5. Detection of hybridization. Oligonucleotide probes with two different base sequences are spotted to the left (transistors 1-31) and right (transistors 54-96) parts of an array of 96 transistors. Each transistor exhibits an active surface of $100 \mu\text{m}^2$. The first hybridization is done with Cy3 labeled target oligonucleotides, matching the probes on the right part of the array. Afterwards, the second hybridization is done with Cy5 labeled targets matching the probes on the left part of the array.

Fig. 6. Detection of the first hybridization. Both electronic (left, $-\Delta U_S$) and fluorescence (right) signals are higher on the right part of the array, where specific hybridization is expected.

Fig. 7. Detection of the second hybridization. This time electronic and fluorescence signals are higher at the left, as expected.

Figures 5, 6 and 7 illustrate an experiment showing electronic and fluorescence detection of hybridization between DNA oligonucleotides. After initial cleaning we coated the surface of our planar device with poly(L-lysine). Using a micropipette we then deposited two different oligonucleotide solutions ($0.3 \mu\text{l}$ of $1 \mu\text{M}$ solutions of 20 mer oligonucleotides in 20 mM KCl) to the left and right parts of the FET array (transistors 1-31 and 54-96, respectively). Incubation was done for 15 min at room temperature, before H_2O rinsing and drying with air.

For hybridization, the sample is incubated for 5 min at room temperature in 20 mM KCl buffer containing target oligonucleotides to a concentration of 100 nM . The hybridization is stopped by rinsing with fresh 20 mM KCl buffer (without target DNA and using a micropipette). The left part of figure 6 shows the difference ΔU_S between two electronic measurements at $[\text{KCl}]=20 \text{ mM}$, done immediately before and after hybridization 1. The base sequence of the target molecules is complementary to the sequence of the probes deposited on the right part of the array; the targets are Cy3 labeled. In hybridization 2 (figure 7), performed afterwards on the same sample, targets are Cy5 labeled and match the probes on the left. Hybridization 1 leads to a stronger negative shift in U_S on the right (8.3 compared to 1 mV in average), while for hybridization 2 we observe a stronger negative shift on the left (8.1 compared to 4.8 mV). This is consistent with specific hybridization in both directions.

The specificity of both hybridizations is subsequently confirmed by fluorescence imaging. For this purpose the sample is rinsed with H_2O and dried. Under excitation at 532 nm (figure 6, right) we observe 4 times more fluorescence on the right part than on the left part of the array. This way we measure the fluorescence arising from the Cy3 labeled targets used in the first hybridization (targets matching the probes on the right part of the array). Under excitation at 633 nm (figure 7) we observe 5 times more fluorescence on the left part than on the right part of the array. This way we measure the fluorescence arising from the Cy5 labeled targets used in hybridization 2 (targets match probes on the left part of the array). Each fluorescence data point corresponds to the average of 3 measurements taken across the active surface of a transistor with a spatial resolution of about 500 nm . The fluorescence measurements are in qualitative agreement with the electronic signals and the signal-to-noise ratios observed with the two different techniques are not very different.

Residual non-specific interactions are expected under our experimental conditions and might explain the signals observed even on the non-matching probes. In particular, here we didn't introduce a blocker step to neutralize the positive charges in the poly(L-lysine) layer that still remain after the deposition of the DNA probes.

The hybridization conditions used in the present experiment (salt, temperature, duration, oligonucleotide length, surface concentrations of probes and hybridized targets) are similar to the conditions used by Fritz et al [10] in ac measurements of oligonucleotide hybridization on capacitive Si/SiO₂ devices and our

differences in the average shifts ΔU_S (7.3 mV for hybridization 1 and 3.3 mV for hybridization 2) are close to their result (3 mV).

It is possible to use different salt concentrations for hybridization and detection in order to optimise both the specificity of the sequence recognition (preferentially high salt) and the sensitivity of the field effect detection (preferentially low salt to reduce screening by mobile ions). We have shown experimentally that the differential signals of hybridization can be enhanced significantly (from about 3 mV to 20 mV) by changing the salt concentration between hybridization and detection [16].

5 Salt dependence of the electronic signals

Descriptions of the silica surface in aqueous environment [23] and of its modifications induced by biomolecule binding are complex, combining surface chemistry, statistical mechanics and electrostatics. In these systems the concentration of the mobile ions in the electrolyte is an important parameter, controlling the biomolecule interaction with the surface as well as the sensitivity of the electronic detection.

Towards a quantitative understanding of the FET based electronic detection, we performed combined experimental and theoretical studies of the salt dependencies of the electrostatic potentials at the silica/biomolecule/electrolyte interface. This section only gives a glimpse at our study of the salt dependence of a bare silica surface in contact with a KCl electrolyte. The interested reader is referred to the original references [12, 22]. The topic is related to the modelisation of pH measurement by field effect devices [24, 25] and to recent publications on the mechanisms of field effect detection of surface bound biomolecules [26–28].

To prepare the experimental study of the salt dependence of the silica/electrolyte interface, we prepare the FET array surface according to Sect. 2.2 without poly(L-lysine) coating. In figure 8, a measurement of the salt dependence of U_{SE} at fixed I_D, U_{SD} working point is shown. Starting at 5×10^{-6} M, a series of increasing concentration has been obtained by successively adding concentrated KCl to the electrolyte. The experimental data are compared to two different theoretical descriptions. In the first description (constant charge model, dashed line) we assume an effective interface charge σ_{eff} which does not depend on the salt concentration [KCl] of the electrolyte. In the second case (salt-dependent charge model, solid line) the dependence of σ_{eff} on [KCl] is taken into account considering the chemical equilibrium of the ionisable sites at the silica surface. This introduces two fitting parameters, the surface density of ionisable site N_S and the pH value. We find that, although the constant charge model can describe the global trend of the measured salt dependence, the salt-dependent charge model provides a better description of the experimental data.

Fig. 8. Average shifts ΔU_{SE} of an array of 29 transistors as a function of the salt concentration of the buffer solution. Zero voltage shift corresponds to the reference measurement performed with a concentration $[KCl]$ of 5×10^{-6} M. The experimental data (squares) are compared to two different theoretical descriptions as explained in the text.

6 Concluding remarks

The electronic detection is sensitive to charged molecules in general. Two different concepts have been used to obtain sequence specific transistor based DNA detection.

In a first approach, point-mutations have been detected in human DNA by allele-specific PCR followed by electronic detection of the double stranded PCR product [15]. The PCR and the electronic detection are separated by a purification that retains PCR product and substrat DNA, but eliminates nucleotides, primers, proteines and salt. As very few substrat DNA is used, the PCR product represents the only species susceptible to induce a significant signal in the electronic detection. Comparative analysis of two different samples on the FET array provides reproducible mutation detection, the overall specificity is simply that of the allele-specific PCR. A combination of FET based DNA detection and sequence specific extension of surface bound oligonucleotides has been reported by Sakata and Miyahara [29].

The second concept consists in obtaining a sequence specific binding on the surface of the FET device by hybridization. Electronic detection of hybridization with probe DNA covalently bound on the surface of FET devices has been reported by several groups [9, 13, 14]. In the work reviewed in Sect. 4, we immobilize DNA probes on poly(L-lysine) and hybridize oligonucleotide targets at low salt, without preceeding blocker step. The low salt hybridization on a

charge-compensated surface has initially been introduced to obtain faster hybridization [30]. It has been shown that under these conditions a single base mismatch can be detected in 12 mer oligonucleotides [10], although more standard hybridization protocols use higher concentrations of monovalent salt (50 mM–1M) to improve stringency [31]. For FET array based electronic detection it has been shown experimentally, that the differential signals of hybridization can be enhanced significantly by changing the salt concentration between hybridization and detection [16].

We use devices with multiple transistor structures to detect DNA. Silicon FET arrays can be made with an individual structure size of a few μm^2 . Using silicon CMOS technology, Infineon fabricated 2D sensor arrays with 16384 FET structures and integrated readout electronics on a chip surface of 1 mm^2 [17]. As the signals from the individual active FET structures do not decrease with sensor surface and because the dc measurement does not require sophisticated on-chip electronics, this suggests that an electronic dc detection based on a two-dimensional transistor array even integrated with on-chip multiplexing would not limit the number of different DNA probes that could be used in parallel. In the field of DNA microarrays, robotic spotting of probe molecules typically gives spot diameters of about $100\text{ }\mu\text{m}$, while smaller diameters down to about $5\text{ }\mu\text{m}$ are achieved by direct on-surface oligonucleotide synthesis.

The planar SiO_2 surface of the FET arrays exhibits sufficiently uniform wetting properties to allow for local spotting and microscopic fluorescence measurements, common in the field of DNA microarrays. In this geometry, integrations with microfluidics approaches and even sophisticated “lab on chip” devices are conceivable and are also potentially interesting in terms of applications.

Acknowledgments

The contributions of my collaborators C. Gentil, F. Pouthas, D. Côte, G. Philipin and C. Guiducci are gratefully acknowledged. We thank M. Ulbrich, M. Völker, G. Zeck, B. Straub and P. Fromherz for stimulating discussions and help in the preparation of the transistor arrays.

References

1. Csako G (2006) Clin Chim Acta 363:6–31
2. Wang J, Nielsen P E, Jiang M, Cai X, Fernandes J R, Grant D H, Ozsoz M, Beglieter A, Mowat M (1997) Anal. Chem. 69:5200–5202
3. Steemers F J, Ferguson J A, Walt D R (2000) Nat. Biotech. 18:91–94.
4. McKendry R, Zhang J, Arntz Y, Strunz T, Hegner M, Lang H P, Baller M K, Certa U, Meyer E, Güntherodt H J, Gerber C (2002) Proc. Natl. Acad. Sci. USA 99:9783–9788
5. Bergveld P (1972) IEEE Trans. Biomed. Eng. BME-19:342–351
6. Bergveld P (1996) Sens. Actuators A 56:65–73
7. Kharitonov A B, Wassermann J, Katz E, Willner I (2001) J. Phys. Chem. B 105:4205–4213

8. Cui Y, Wei Q, Park H, Lieber C M (2001) *Science* 293:1289–1292
9. Souteyrand E, Cloarec J P, Martin J R, Wilson C, Lawrence I, Mikkelsen S, Lawrence M F (1997) *J. Phys. Chem. B* 101:2980–2985
10. Fritz J, Cooper E B, Gaudet S, Sorger P K, Manalis S R (2002) *Proc. Natl. Acad. Sci. USA* 99:14142–14146
11. Hahm J, Lieber C M (2004) *Nano Lett.* 4:51–54 (2004).
12. Pouthas F, Gentil C, Côte D, Zeck G, Straub B, Bockelmann U (2004) *Phys. Rev. E* 70:031906
13. Uslu F, Ingebrand S, Mayer D, Böcker-Meffert S, Odenthal M, Offenhäuser A (2004) *Biosens. Bioelectron.* 19:1723–1731
14. Sakata T, Miyahara Y (2005) *ChemBioChem* 6:703–710
15. Pouthas F, Gentil C, Côte D, Bockelmann U (2004) *Appl. Phys. Lett.* 84:1594–1596
16. Gentil C, Philippin G, Bockelmann U (2007) *Phys. Rev. E* 75:011926
17. Eversmann B, Jenkner M, Hofmann F, Paulus C, Brederlow R, Holzapfl B, Fromherz P, Merz M, Brenner M, Schreiter M, Gabl R, Plehnert K, Steinhäuser M, Eckstein G, Schmitt-Landsiedel D, Thewes R (2003) *IEEE J. Solid-St. Circ.* 38:2306–2317
18. Kiessling V, Müller B, Fromherz P (2000) *Langmuir* 16:3517–3521
19. Zeck G, Fromherz P (2001) *Proc. Natl. Acad. Sci. USA* 98:10457–10462
20. Manning G S (1978) *Quart. Rev. Biophys.* 11:179–246
21. Pack G R, Wong L, Lamm G (1999) *Biopolymers* 49:575–590
22. Gentil C, Côte D, Bockelmann U (2006) *Phys. Stat. Sol. A* 203:3412–3416
23. van Hal R E G, Eijkel J C T, Bergveld P (1996) *Adv. Coll. Interf. Sci.* 69:31–62
24. Siu W M, Cobbald R S C (1979) *IEEE Trans. El. Dev.* 26:1805–1815
25. Fung C D, Cheung P W, Ko W H (1986) *IEEE Trans. El. Dev.* 33:8–18
26. Landheer D, Aers G, McKinnon W R, Deen M J, Ranuarez J C (2005) *J. Appl. Phys.* 98:044701
27. Neff P A, Wunderlich B K, Lud S Q, Bausch A R (2006) *Phys. Stat. Sol. A* 203:3417–3427
28. Poghosian A, Cherstvy A, Ingebrandt S, Offenhäuser A, Schoning M J (2005) *Sens. Actuators B* 111:470–480
29. Sakata T, Miyahara Y (2006) *Angew. Chem. Int. Ed.* 45:2225–2228
30. Belosludtsev Y, Belosludtsev I, Iverson B, Lemeshko S, Wiese R, Hogan M, Powdrill T (2001) *Biochem. Biophys. Res. Commun.* 282:1263–1267
31. Sambrook J, Russel D W (2000), *Molecular Cloning, A Laboratory Manual*. Cold Spring Harbor Laboratory Press, New York