ERRATUM

Buccal Micronucleus Cytome Assay

Philip Thomas and Michael Fenech

Vladimir V. Didenko (ed.), *DNA Damage Detection In Situ, Ex Vivo, and In Vivo: Methods and Protocols*, Methods in Molecular Biology, vol. 682, DOI 10.1007/978-1-60327-409-8, pp. 235–248, © Springer Science+Business Media, LLC 2011

DOI 10.1007/978-1-60327-409-8_22

The publisher regrets that in chapter 17, page 239, the Buccal Reagent Setup contains an error in the proportions of salts required to prepare the buccal cell buffer. Reagent Setup description:

BUCCAL REAGENT SETUP

Buccal cell buffer To prepare 1 liter of buccal buffer, weigh 1.6 g Tris–HCl, 1.2 g EDTA and 37.2 g of sodium chloride, and dissolve in 600 ml of Milli-Q water. Thoroughly dissolve the salts and adjust the volume to 1,000 ml. Adjust pH to 7.0 and autoclave at 121 °C for 30 min. The buffer will last for up to 3 months when stored at room temperature.

The correct version is:

BUCCAL REAGENT SETUP

Buccal cell buffer To prepare 1 liter of buccal buffer, weigh 1.6 g Tris–HCl, 38.0 g EDTA and 1.2 g of sodium chloride, and dissolve in 600 ml of Milli-Q water. Thoroughly dissolve the salts and adjust the volume to 1,000 ml. Adjust pH to 7.0 and autoclave at 121 °C for 30 min. The buffer will last for up to 3 months when stored at room temperature.

The online version of the original chapter can be found at
http://dx.doi.org/10.1007/978-1-60327-409-8_17