Biomechanics of Bone Cement Augmentation with Compression Hip Screw System for the Treatment of Intertrochanteric Fractures

S.J. Lee¹, B.J. Kim¹, S.Y. Kwon², and G.R. Tack³

¹Department of Biomedical Engineering, Inje University, 607 Obang-Dong, Kyung-nam, Korea
{sj1, kbj93}@bme.inje.ac.kr

²Department of Orthopaedic Surgery, Youido St' Mary's Hospital, Catholic University, Seoul, Korea

sykwon@cmc.cuk.ac.kr

³Dept. of Biomedical Engineering, Konkuk University, 322 Danwol-Dong, Chungju, Chungbuk, Korea
grtack@kku.ac.kr

Abstract. Recently, cement augmentation technique was introduced to enhance screw fixation within the femoral head. This study investigates biomechanical efficacies of cement augmentation technique. Finite element models of the femur with sliding hip screw assemblies were constructed with and without bone cement augmentation. Appropriate contact conditions with varying friction coefficients were assigned to simulate the fracture planes and other interfacial regions. With cement augmentation, 80% reduction in stresses was found in the cancellous bone, suggesting reduced possibility of cancellous bone fracture and screw cut-outs. The peak von Mises stress within the cement mantle was about 1/3 of its fatigue strength, which suggested the longevity of the cement mantle and less likelihood of osteolysis due to cement debris. Micromotions at the hip screw interfaces were also down from 0.275mm to 0.008mm, an indication for strong fixation after the surgery.

1 Introduction

Recent findings estimate that intertrochanteric femoral fractures (IFF) occur in more than 200,000 patients each year in the United States alone, with reported mortality rates ranging from 15% to 20% [1]. Most IFFs occur in patients over 70 years of age and they are likely to increase each year as the population of these age group increases. Traffic accidents contribute significant increase in patient numbers as well.

One of the most commonly used surgical treatments for IFF is using internal fixation devices such as sliding compression hip screws with side plate assemblies. These devices are considered to be safe with minimum amount of drilling of the cancellous bone in the femoral head and neck region. They allow early weight-bearing and limit interfacial movement on the fracture plane by providing strong compressive forces.

Clinical studies show, however, that superior cutting-out through the femoral head by the sliding hip screw is one of the major complications [2,3]. Improper screw placement has been cited as one of the major contributing factors. Frequent non-union

on the fracture plane often requires revision surgery that can compound more surgical difficulties [2,4]. For patients with severe osteoporosis in which the primary compression trabecular structures have diminished in considerable proportions can be more prone to this type of complication and may add further surgical challenges. Therefore, it is very important to maintain strong fixation of the sliding hip screws within the femur with minimum amount of sliding motion for optimal bony-healing.

Biomechanical studies have been done to elucidate the effects of biomechanical factors such as the placement angle of the hip screws, positioning within the femoral head and neck, and the effect of anatomical reduction [5,6]. Recently, augmenting of hip screws with polymethylmethacrylate (PMMA) has been suggested to provide more secure fixation of the screws within the femoral region [3,4]. Clinical study by Cheng et al (1989) found that cement augmentation may provide initial stability but can lead to late complications if not properly used [7]. Bartucci et al (1985) advocated limiting the use of PMMA on the proximal side to prevent intrusion of the cement into the fracture plane [4]. Using cadaveric femurs Choueka et al (1995) investigated biomechanical effects of using various types of sliding hip screws and/or dome plunger in terms of load-bearing and fixation strength [2].

This study was designed to investigate the biomechanical feasibility of the bone cement augmentation technique. The changes in stress distributions due to cement augmentation were studied. For this purpose, finite element models of the operated femur with sliding hip screw assemblies were constructed with and without bone cement augmentation. Close attentions were given to the peak von Mises stress within the cement mantle region to assess the likelihood of cement failure that might lead to osteolysis due to cement debris. In addition, changes in micromotion at the interfacial junctions such as screw-cement-surrounding cancellous bone and fracture plane were assessed to study the fixation strength that is critical for post-op bony healing.

2 Material Methods

A three-dimensional finite element model of the femur was constructed using geometric data acquired from computed tomography (CT) scans (Fig. 1). Intertrochanteric fracture was simulated by assigning a fracture plane from the greater trochanter to the lesser trochanter at 30 degrees with respect to the long axis of the femur (Fig. 2). Based on this hip fracture model, two types of surgical model were generated. One is the non-cemented model (NC) in which only the hip screw assembly was added to the femur and the other is the cement-augmented model (CA) where the cement mantle was added to surround the hip screw (Fig. 3).

The dimension of the hip screw assembly as well as screw profiles (screw threads, pitches, inner and outer diameters) were based on the Richard compression hip screw (Osteo, Swiss). The distance from the apex of the femoral head to the tip of the screw known as TAD (Tip-Apex Distance) was set at 20-24mm near the center of the femoral head according to suggestion by Baumgaertner et al [8]. The average thickness of the cement mantle was determined as 12.5mm distributed uniformly over the hip screw based on the clinical experience of one of the authors (SYK).

Fig. 1. Geometric data acquired from CT scans

The material properties for the anatomical and implants were based on the findings from literatures (Table I, [9]). To simulate the interfacial conditions, appropriate friction coefficients (μ) were assigned at the fracture plane ($\mu=0.5$), the interface between screw and surrounding cancellous bone ($\mu=0.5$), and between the screw and the cement ($\mu=0.3$). The cement-bone interface was assumed to be rigidly and the side plate assembly was rigidly fixed to the lateral aspect of the femur. The sources for the friction coefficient are also listed Table I. 2014N of compressive loading in a cubic cosine distribution was simulated on the top of the femoral head at angles of 12 and 26 degrees in sagittal and transverse planes, respectively (Fig. 4). This loading condition is designed to simulate the after heel strike phase in a gait cycle in which the highest compressive load is applied to the femur. It was assumed that the distal part of the model was fixed in all directions. Eight-noded 3-D brick elements were used for the entire model except for the interfacial regions where the 3-D contact elements were used.

Fig. 2. Construction of a finite element model

3 Results

Strain data from three different locations on the cortical surface of the femur were compared for the validity check (Table II). Here, two sets of results were in a very close agreement with each other, thereby confirming the validity of our model.

Results showed that cement augmentation had resulted in decrease in stresses in the hip screw and in the bone, both cortical and cancellous (Table III), an indication of favorable stress transfer due to the addition of bone cement. Highest decrease in stresses was noted with cement augmentation at the cancellous bone region (reduction of 80%), which makes the further fracture of the femur far less likely in cement-augmented (CA) case than in non-cemented (NC).

Fig. 3. FE models, (A) non-cemented(NC) (B) cement-augmented(CA)

Fig. 4. Loading & boundary conditions of a finite element model

Table 1. Mechanical properties of the proximal femur for finite element modeling

		Elastic modulus (MPa)	Possion's ratio (v)	Coefficient of friction
Cortical bone	Subchondral bone	2,000	0.32	
	Pure cortical bone	14,000	0.32	
Cancellous bone	Femoral head	550	0.32	
	Proximal region	411	0.32	
	Distal region	345	0.32	
Bone cement		2,200	0.23	
Compression hip screw		200,000	0.30	
Cement-screw				0.3 ^[13]
Cancellous bone – screw				0.5 ^[14]
Fracture plane				0.5 ^[15]

In fact, the peak von Mises stresses (PVMS) of 27MPa was aspsessed at the screw-bone junction in the cancellous bone with NC. This exceeded its yield strength of 22MPa suggesting the likelihood of micro-fracture in this region. On the other hand, with CA the corresponding PVMS decreased to 5MPa, making it far less susceptible to fracture. The PVMS at the screw was considerably lower than its yield strength. For example, the PVMS at the hip screw were only 322MPa, slightly above one third of its yield strength of 860MPa (stainless steel). The peak von Mises stress at the cement region was about 9MPa that was far less than 1/3 of its reported shear strength of 30MPa and endurance limit of 28MPa at one million cycles [11].

The micromotions at the screw-bone (NC) and the screw-cement (CA) interfaces were 0.275mm and 0.008mm, respectively. The high level of micromotion in NC coupled with high PVMS that exceeds the yield strength may all contribute the eventual cut-out of the screw because the microscale failure of the cancellous bone and increased micromotion may be initiated

Table 2. Validation results

Method	Experimental results from Oh et al [8]	Results from the present study using FEM
Location	Mean ± SD (μ €)	
Medial 1	1827 ± 601	1273.6
Medial 2	1419 ± 628	988.4
Lateral	1019 ± 404	623.2
Remarks	Results from strain gauge measurement	Predicted values from matching nodes

Table 3. Comparison of peak von Mises stress (PVMS) for the non-cemented (NC) and the cement-augmented (CA) cases

Model location	Non-cemented (NC)	Cement- augmented (CA)	relative %change in PVMS	Yield Strength (MPa)
	PVMS (MPa)	PVMS (MPa)		
Hip screw region	322.61	174.97	-45.76	860.00
Cement region		9.16		30.00
Cortical region	48.81	54.63	11.92	560.00
Cancellous region	27.74	4.88	-82.41	22.00

4 Discussion

This study investigated the biomechanical advantages of adding bone cement to reinforce the hip screw fixation during the surgical treatment of intertrochanteric fracture of the femur by using clinically relevant finite element (FE) models. Although the intertrochanteric fractures often occur in elderly patients with osteoporosis, the material properties of the bone in this study used those of the normal person in order to elucidate its feasibility in a simplified way. It would be very interesting to conduct similar studies for varying degrees of osteoporosis to assess its feasibility by introducing appropriate conversion values for the elastic stiffness in accordance with the BMD data of a given patient.

Our results clearly suggested the biomechanical advantages of bone cement augmentation. In particular, the most drastic reduction in stresses was seen at the cancellous bone. Almost 80% of reduction was noted. This makes the cut-out of the screw that has been advocated as one of the major complications of the hip screw systems far less likely to take place. In fact, the PVMS with non-cemented case was higher than its yield strength (27MPa vs. 22 MPa), a sign of impending loosening of the hip screw at the screw-bone interface that may progress to cut-out of the screw.

No difference in micromotion at the fracture plane was assessed between two cases (CA and NC). At the cement-bone interface, we assumed that interdigitization of cement took place with the irregularities of the bone (i.e., microinterlock). Therefore, we did not assign any friction coefficient. Rather we took it as continuity (i.e., the nodes are shared) and no micromotion was assessed here. It should be noted that at the interface between screw and cement, we adopted the friction coefficient value of 0.3 as suggested by Mann et al [13], which was lower than the that of screw-bone interface ($\mu = 0.5$, suggested by Shrazi-Adl et al [14]). The initial micromotion assessed immediately after surgery has been cited as one of the indicators whether successful bony healing can take place. Studies have shown that micromotion should be less than (0.15mm) for bony fusion at fracture sites [12]. Based on the magnitudes of the micromotion predicted from this study (0.008mm with CA), cement augmentation could reduce the risk of non-union or delayed union of the fracture that might have had progressed otherwise (0.275mm in NC). Our results also demonstrated that despite the lower friction coefficient assigned for the screw-cement interface (CA) its micromotion was substantially less than that of non-cemented (NC). This may due to the fact that the deformation is inversely proportional to the elastic

modulus. It appears higher modulus of bone cement ($E = 22000\text{MPa}$) as opposed to the cancellous bone ($E = 550\text{MPa}$) overcame the relative lack of friction and resulted in less micromotion.

The peak stress found in cement mantle was about less than $1/3$ of its fatigue strength. This suggests that the bone cement mantle may theoretically withstand the repeated load indefinitely under the loading condition provided in this study. Although much in-depth study should be done, this result may relieve clinicians from worries about the likelihood of debris-causing fatigue failure that may initiate the vicious cycles of osteolysis and implant loosening.

5 Conclusion

Our finite element analysis was able to show the following advantages of bone cement augmentation technique in the treatment of IFF. They can be summarized as:

- A significant reduction in stresses in the cancellous bone: This indicates that further fractures of the cancellous bone or the screw cut-outs are far less likely to take place.
- Cement mantle could withstand against the physiological load and further against the repeated fatigue load.
- Reduced micromotions at the interfacial regions suggest stronger chance of bony healing.

This work was supported by a grant from Korea Science & Engineering Foundation (R01-2001-0500).

References

1. Sledge, C.B.: *Master technique; in orthopaedic surgery, The Hip*, Lippincott-Raven, Philadelphia (1998) 106
2. Choueka, J., Koval, K.J., Kummer, F.J., Crawford, G.: Biomechanical comparison of the sliding hip screw and the dome plunger. *The Journal of Bone and Joint Surgery (Br)*, 77(2) (1995) 277-283
3. Choueka, J., Koval, K.J., Kummer, F.J., Zuckerman, J.D.: Cement augmentation of intertrochanteric fracture fixation. *Acta Orthop Scand* 67(2) (1996) 153-157
4. Bartucci, E.J., Gonzalez, M.H., Cooperman, D.R., Freedberg, H.I., Barmada, R., Laros, G.S.: The effect of adjunctive methylmethacrylate on failures of fixation and function in patients with intertrochanteric fractures and osteoporosis. *The Journal of Bone and Joint Surgery (Am)* 60(7) (1985) 1094-1107
5. Hartog, B.D., Bartal, E., Cooke, F., Kansas, W.: Treatment of the unstable intertrochanteric fracture. *The Journal of Bone and Joint Surgery (Am)*, 73-A(5) (1991) 726-733
6. Kyle, R.F., Wright, T.M., Burstein, A.H.: Biomedical analysis of the sliding characteristics of compression hip screws. *The Journal of Bone and Joint Surgery (Am)*, 62-A(8) (1980) 1308-1314

7. Cheng, C.L., Chow, S.P., Pun, W.K., Leong, J.C.: Long-term results and complications of cement augmentation in the treatment of unstable trochanteric fractures. *Injury* 20(3) (1989) 134-138
8. Baumgaertner, M.R., Curtin, S.L., Lindskog, D.M., Keggi, J.M.: The value of the tip-apex distance in predicting failure of fixation of peritrochanteric fractures of the hip. *The Journal of Bone and Joint Surgery (Am)*, 77-A(7) (1995) 1058-1064
9. Brown, T.D., Ferguson, J.A.: Mechanical property distributions in the cancellous bone of the human proximal femur. *Acta Orthop Scand* 13 (1980) 687-699
10. Oh, I.D., Harris, W.H.: Proximal strain distribution in the loaded femur. *The Journal of Bone and Joint Surgery (Am)* 60(1) (1978) 75-85
11. Kuhn, K.D.: *Bone cement*, Springer; Berlin Heidelberg New York (2000) 19-20; 142-147
12. Pilliar, R.M., Lee, J.M., Maniopoulos, C.: Observations on the effect of movement on bone ingrowth into porous-surfaced implants. *Clinical Orthopaedics and Related Research* 208 (1986) 108-113
13. Mann, K.A., Bartel, D.L., Wright, T.M., Burstein, A.H.: Coulomb frictional interfaces in modeling cemented total hip replacements: a more realistic model. *J. Biomechanics* 28(9) (1995) 1067-1078
14. Shirazi-Adl, A., Dammak, M., Paiement, G.: Experimental determination of friction characteristics at the trabecular bone/porous-coated metal interface in cementless implants. *J. Biomed. Mater. Res.* 27 (1993) 167-175
15. Natarajan, R.N., Chen, B.H., An, H.S., Andersson, B.J.: Anterior Cervical Fusion; A finite element model study on motion segment stability including the effect of osteoporosis *Spine* 25(8) (2000) 955-961