Optimal Acquisition Schemes in High Angular Resolution Diffusion Weighted Imaging

V. Prčkovska¹, A.F. Roebroek², W.L.P.M. Pullens^{2,3},
A. Vilanova¹, and B.M. ter Haar Romeny¹

¹ Dept. of Biomedical Engineering, Eindhoven Univ. of Technology, The Netherlands
{V.Prckovska, A.Vilanova, B.M.terHaarRomeny}@tue.nl

² Maastricht Brain Imaging Center, Dept. of Cognitive Neuroscience, Faculty of
Psychology, Maastricht University, The Netherlands
A.Roebroek@PSYCHOLOGY.unimaas.nl

³ Brain Innovation B.V., Maastricht, The Netherlands
pullens@brainvoyager.com

Abstract. The recent challenge in diffusion imaging is to find acquisition schemes and analysis approaches that can represent non-gaussian diffusion profiles in a clinically feasible measurement time. In this work we investigate the effect of b -value and the number of gradient vector directions on Q-ball imaging and the Diffusion Orientation Transform (DOT) in a structured way using computational simulations, hardware crossing-fiber diffusion phantoms, and in-vivo brain scans. We observe that DOT is more robust to noise and independent of the b -value and number of gradients, whereas Q-ball dramatically improves the results for higher b -values and number of gradients and at recovering larger angles of crossing. We also show that Laplace-Beltrami regularization has wide applicability and generally improves the properties of DOT. Knowledge of optimal acquisition schemes for HARDI can improve the utility of diffusion weighted MR imaging in the clinical setting for the diagnosis of white matter diseases and presurgical planning.

1 Introduction

Diffusion-weighted Magnetic Resonance Imaging (DW-MRI) is a clinical medical imaging technique that provides a unique view on the structure of brain white matter in-vivo. White matter fiber-bundles are probed indirectly by measuring the directional specificity (anisotropy) of local water diffusion. Post-processing of diffusion weighted images is fundamentally aimed at calculating the probability distribution function (pdf) for the displacement of water molecules in each imaging voxel. In general, the task is to find the transform Φ that takes the measurements $S(\mathbf{g}_i)$, for a finite set of 3D diffusion gradient vectors $\mathbf{g}_i = \{x_i, y_i, z_i\}$ $i \in \{1, \dots, N\}$, to the desired pdf $P(\mathbf{r}_i)$, as in $P(\mathbf{r}_i) = \Phi[S(\mathbf{g}_i)](\mathbf{r}_i)$. Here the pdf $P(\mathbf{r}_i)$ is a function of the 3D displacement vector $\mathbf{r}_i$, and we generally have measured signals $S(\mathbf{g}_i)$. In Diffusion Tensor Imaging (DTI) the pdf is assumed to be a 3D gaussian distribution represented as a rank-2 tensor and the effect of the

diffusion gradients on the MR signal is assumed to be of an exponentially decaying form. However, the diffusion tensor cannot resolve more than a single fiber population per voxel, because partial volume effects will cause a non-gaussian diffusion profile with multiple maxima [1].

It has been shown if we re-express our signal as $E(\mathbf{q}_i) = S(\mathbf{q}_i)/S_0$, that the sought relation is the Fourier Transform, where $\mathbf{q}_i$ is the diffusion wavevector, defined as $\mathbf{q}_i = (2\pi)^{-1}\gamma\delta\mathbf{g}_i$ (where γ is the gyromagnetic ratio and δ is the duration of the diffusion gradients), and S_0 is the unweighted or zero-weighted baseline signal obtained without any applied diffusion gradients. However, directly calculating the Fourier Transform, as in Diffusion Spectrum Imaging (DSI) [2], requires a very large number of gradients, thus ensuing long measurement times. In High Angular Resolution Diffusion Imaging (HARDI) a moderate amount (from about 60 to a few hundred) of diffusion gradients are scanned that together sample a sphere of given radius [3]. Among the analysis techniques that transform this data to certain probability function (Orientation Distribution Function, Fiber Orientation or Probability Function given a position, etc.) are Q-ball imaging [4], Spherical Deconvolution (SD) [5], Diffusion Orientation Transform (DOT) [6] and Persistent Angular Structure (PAS-MRI) [7].

In this work we use computational simulations, hardware crossing-fiber diffusion phantoms, and in-vivo brain scans to investigate the effect of acquisition parameters on the ability to reconstruct non-gaussian diffusion profiles in crossing fiber regions. We quantify the angular resolution of two selected reconstruction methods Q-ball imaging, and the Diffusion Orientation Transform under different acquisition schemes. We vary b -value (defined as $b = \gamma^2\delta^2(\Delta - \delta/3)$, where Δ is the time between the two complementary diffusion gradients) and number of gradients directions in a structured way to investigate their effects and interaction.

2 Methods

2.1 Ground Truth Synthetic Data

We generate synthetic data by simulating the diffusion-weighted MR signal attenuation from molecules, with free diffusion coefficient D_0 , restricted inside a cylinder of radius ρ and length L as in [8]. Two fiber crossings were simulated under $40^\circ, 45^\circ, 50^\circ, 55^\circ, 60^\circ$ and 90° , with the following set of parameters (see [6]): $L = 5\text{mm}$, $\rho = 5\text{ }\mu\text{m}$, $D_0 = 2.02 \cdot 10^{-3}\text{mm}^2/\text{s}$. All simulated imaging parameters were chosen to be the same as in our MRI acquisition protocol, described in section 2.3. Simulations were performed with and without noise. For the noise simulations, Gaussian noise was added to the real and complex part of the signal, with standard deviation according to the SNR calculated in our MRI acquisition protocol (Fig. 1) for the corresponding b -values. Mean and standard deviation were calculated over 100 noise realizations for each set of gradient directions, b -values and simulated angle. The angular error was taken as the mean of the individual absolute differences between the simulated and recovered angles.

2.2 Ground Truth Hardware Phantoms

Hardware phantoms were constructed using the method described earlier [9]. Two bundles, each composed of 25 sub-bundles of 400 fibers (KUAG Diolen, Germany; Each fiber consists of 22 filaments of $\approx 10\mu\text{m}$) were positioned interdigitated to create artificial crossings. Three phantoms were constructed, with crossings at 30° , 50° and 65° respectively. The phantoms were placed in a container, filled with a $0.03 \text{ g/l MnCl}_2 \cdot 4 \text{ H}_2\text{O}$ solution (Siemens, Erlangen, Germany) to obtain T_2 relaxation of approximately 90ms, corresponding with human white matter T_2 . 2.4 g/l NaCl was added for resistive coil loading.

2.3 MRI Data Acquisition

Human: DW-MRI acquisition was performed on subject VP (25 yrs, female) using a twice refocused spin-echo echo-planar imaging sequence on a Siemens Allegra 3T scanner (Siemens, Erlangen, Germany). Informed consent was obtained prior to the measurement. FOV $208 \times 208 \text{ mm}$, voxel size $2.0 \times 2.0 \times 2.0 \text{ mm}$. 10 horizontal slices were positioned through the body of the corpus callosum and centrum semiovale. Custom gradient direction schemes, created with the electrostatic repulsion algorithm [10] were used for DW-MRI. The diffusion-weighted volumes were interleaved with b_0 volumes every 12th scanned diffusion gradient directions. Data sets were acquired with #vols(#dirs): 132(120) 106(96), 80(72), 54(48) directions, each at b-values of 1000, 1500, 2000, 3000, 4000 s/mm^2 , using gradient timing δ : gradient pulse duration, and Δ : gradient spacing as given in Fig. 1. In the same session, two anatomical data sets (192 slices, voxel size $1 \times 1 \times 1 \text{ mm}$) were acquired using the ADNI protocol. Total scanning time was 75 minutes.

parameters	b (s/mm^2)				
	1000	1500	2000	3000	4000
Δ (ms)	32.44	35.94	38.94	43.44	47.44
δ (ms)	25.34	28.84	31.84	36.34	40.34
SNR	15.3	14.1	13.3	12.0	11.9

Fig. 1. Parameters from our acquisition protocol

Hardware phantom: The hardware phantom was scanned using exactly the same DW-MRI protocol as the human subject. The 54 direction scheme and ADNI were omitted from the protocol. Slices were positioned orthogonal to the legs of the phantom, through the crossing region.

2.4 Analysis

We implemented the analytical Q-ball imaging [11] and the parametric and non-parametric DOT [6]. Both of the implemented DOT techniques gave identical results. It is important to note that DOT has two extra tunable parameters: the effective diffusion time t , and the radius of a sphere R_0 that defines the

probability function. As the results of DOT depend on these values, we experimentally found the optimal R_0 and extracted the effective diffusion time from the imaging parameters as $t = \Delta - \delta/3$ (see Fig. 1). For finding the optimal R_0 , we varied the value of R_0 over a wide range of discrete values, and in our analysis we considered the ones that gave smallest angular error (e.g., in the simulation cases the "good candidates" were $R_0 = \{0.022, 0.024, 0.026, 0.028, 0.03\} \mu\text{m}$). We varied the order of the Spherical Harmonics ℓ , between 4 and 8. As a preprocessing step for the noisy data sets, we included Laplace-Beltrami smoothing on the signal, for both of the methods with $\lambda = 0.006$ as in [12]. We assume that the fiber directions are simply given by the local maxima of the normalized $[0,1]$ ODF/probability profile where the function surpasses a certain threshold (here, we use 0.5). To ensure that the minimal expected error related to the sphere tessellation is less than 7.2° [13], we use 4th order of tessellation of an icosahedron.

3 Results

3.1 Noiseless Synthetic Simulations

Results of the angular errors are shown on Fig. 2 for the different simulated direction tables and b -values. Both of the methods managed to recover most angles with identical angular error over a wide range of b and ℓ . Naturally as the angle of crossing increases the angular error decreases. Tables on Fig. 2b summarize the optimal b value and model order ℓ for Q-ball and DOT respectively, where as optimal, we consider the *lowest* combination of b and ℓ for which the angular error for each simulated angle is minimal. From the table we can conclude that DOT is able to recover the simulated angles with the same angular error as

Fig. 2. a) Detected angular difference for DOT and Q-ball in noiseless synthetic data, for all sets of gradient directions. b), c) Optimal set of b -value and order ℓ , for the minimal angular error (per corresponding simulated angle) w.r.t. number of gradient directions for DOT and Q-ball respectively.

Q-ball, but with lower combination of b and ℓ order. Furthermore, DOT reconstructs the correct profiles, for each simulated angle, generally at lower values of ℓ than Q-ball.

3.2 Noisy Synthetic Simulations and Hardware Phantom Data

In Fig. 3, the angular error of the minimal recovered angle for the different acquisition schemes is shown. Our criteria for minimal detected angle is *the smallest angle for which the angular error is no more than 20° , given that the standard deviation (σ) over the noise realizations is no more than 20°* . It should be noted that a liberal threshold on σ was needed as it was generally very high over 100 realizations at the simulated low but realistic SNRs. If multiple angles are found within this criteria, we choose the one whose probability is highest. The results from Fig. 3 coincide with the conclusion from the noiseless data sets. DOT is more robust to noise, thus it manages to recover, in most of the cases, angles of 45° , whereas Q-ball is quite dependent on the b -value and the number of gradient directions in case of small angles of crossings. Furthermore for the same set of parameters, the angular error of the same recovered angle is smaller in DOT (e.g., In Fig. 3, for 54 gradients and $b = 1000s/mm^2$ DOT found angle of 90° with smaller error than the reported one in Q-ball. Yet we report angle of 50° for DOT, since that is the minimal angle that was recovered

Fig. 3. The angular error is plotted for the smallest found angle within our criteria, for each combination of b -value and direction table (given on the x-axis), both for DOT (upper panel) and Q-ball (lower panel). On the x-axis are given (from top to bottom): the best detected angle, order ℓ , b -value and number of gradient directions. Different colors correspond to different b -values.

Fig. 4. The found angular error of the recovered angles in phantom data, by DOT and Q-ball respectively, for each combination of number of gradients, b-value and simulated angle. On the x-axis are given (from top to bottom): order ℓ , b-value, number of gradient directions and simulated angle. Different colors correspond to different b-values.

with this parameter combination and within our criteria. In this concrete case Q-ball failed to accurately recover angles smaller than 90°).

In Fig. 4 the analysis of the phantom data is shown. Using the same criteria as above on 60 voxels located in the crossings, we show that DOT is able to recover even an angle with 30° of crossing, at a cost of very high ℓ order and number of gradient directions. Again we observe that DOT is more robust to noise and independent of the b -value and number of gradients, whereas Q-ball improves the results for higher b -value and number of gradients and at recovering larger angles of crossing.

3.3 In-Vivo Human Brain Data

The centrum semiovale was used for the qualitative analysis of the Q-ball and DOT techniques. It is a challenging region for DW-MRI analysis techniques, since fibers of the corpus callosum, corona radiata, and superior longitudinal fasciculus form a three-fold crossing there. Region-of-interest (ROI) was defined on a coronal slice.

Fig. 5 shows the normalized Q-ball and DOT glyph reconstructions of the DW-MRI datasets with 4th order of Spherical Harmonics, for 54 and 132 gradient directions and b -values of 1000 and 4000 s/mm². All the images are from similar

Fig. 5. Q-ball and DOT representations of the DW-MRI data, with 54 and 132 gradient directions, and $b=1000$ and $b=4000$ s/mm^2 , of a human subject in a ROI defined on a coronal slice in the centrum semiovale. The glyphs are shown for 4th order of Spherical Harmonics.

region in the different datasets, but no registration was done, so the shown glyphs do not correspond exactly. In the background, the color coded FA map is visible, with the corpus callosum (CC) in red, going from lower right to upper left and corona radiata (CR) in blue, from lower left to upper right. The crossing region in the middle is shown in purple.

Both CC and CR structures can be clearly separated. Overall, the data shows an increase in quality when raising the number of gradient directions. When comparing the Q-ball images, the higher b -value (4000 s/mm^2) case shows more detail than the lower b -value (1000 s/mm^2) across all data sets. The DOT images do not show a decrease in quality when comparing both b -values and are rather invariant to the number of the sampled gradient directions.

It's important to note that tuning the optimal R_0 parameter for the DOT, is not trivial in *in-vivo* data sets, and by simplicity and speed of calculations of the reconstructed ODFs, Q-ball has significant advantage.

4 Discussion and Conclusions

In the comparison between the two tested methods there seems to be a tendency for DOT to be more robust to noise and relatively independent of the b -value and number of gradients, whereas Q-ball improves the results for higher b -value and number of gradients and at recovering larger angles of crossing. DOT is able to recover the same (in the noiseless simulations) or even smaller angles (in the noisy simulations) as the used Q-ball implementation, with smaller b -values, and generally at a smaller ℓ . In the hardware phantom data DOT seems to be able to recover even an angle with 30° of crossing, at a high ℓ order and number of gradient directions. This makes DOT comparable to today's state-of-the art Spherical Deconvolution, and future work is addressed in finding the minimum angle at which DOT can still distinguish between two crossings. A few cautions

are in place here. First, these results are specific to the implementations of the methods used here. Particularly, a spherical harmonic transform was used on the data as a first preprocessing step that served to smooth the data and regularize the fit of the models in a unified and standard way for both methods. Our results are thus dependent on choices for the SH order ℓ and Laplace-Beltrami regularization coefficient λ , which we fixed to 0.006, as detailed in [12]. It is very interesting to see that this regularization approach originally constructed for fast and robust Q-ball imaging seems to have a more general utility, since the DOT also benefits from its smoothing properties. Future work should investigate in details the optimal values for Laplace-Beltrami regularization coefficient λ for the DOT, or consider different regularization approaches that might improve DOT's properties. A second important aspect of implementation is the R_0 parameter for the DOT. The results of DOT are highly dependent on the choice of R_0 , which we tried to resolve by computing the DOT at several values of R_0 and then reporting the best result. Future work should be aimed at robustly finding the optimal R_0 under different acquisition protocols. Last aspect is on maxima detection. Current maxima detection approaches in HARDI are quite poor. Finding some more general and robust algorithms, can improve the accuracy of the HARDI methods, as well as help in developing better fibertracking techniques. This work shows that validation of HARDI methods in the ranges of noise present in actual clinical data sets is highly important. Here we compared only two of the many available (and clinically applicable) HARDI techniques. It would be very interesting if techniques as Spherical Deconvolution and PAS-MRI can be subject of similar comparison.

Acknowledgements

We greatly acknowledge Maxime Descoteaux for many fruitful discussions and sharing of code. Furthermore we are thankful to Evren Özarıslan, for all valuable explanations on DOT and synthetic data generation. Finally we thank Paulo Rodrigues for many useful advices on the implementation issues.

References

1. Alexander, A.L., Hasan, K.M., Lazar, M., Tsuruda, J.S., Parker, D.L.: Analysis of partial volume effects in diffusion-tensor MRI. *Magn. Reson. Med.* 45, 770 (2001)
2. Wedeen, V.J., Hagmann, P., Tseng, W.Y., Reese, T.G., Weisskoff, R.M.: Mapping complex tissue architecture with diffusion spectrum magnetic resonance imaging. *Magn. Reson. Med.* 54(6), 1377–1386 (2005)
3. Frank, L.R.: Anisotropy in high angular resolution diffusion-weighted MRI. *Magn. Reson. Med.* 45(6), 935–939 (2001)
4. Tuch, D.S.: Q-ball imaging. *Magn. Reson. Med.* 52, 1358–1372 (2004)
5. Tournier, J.D., Calamante, F., Gadian, D.G., Connelly, A.: Direct estimation of the fiber orientation density function from diffusion-weighted MRI data using spherical deconvolution. *Neuroimage* 23(3), 1176–1185 (2004)

6. Özarslan, E., Shepherd, T.M., Vemuri, B.C., Blackband, S.J., Mareci, T.H.: Resolution of complex tissue microarchitecture using the diffusion orientation transform (DOT). *NeuroImage* 36(3), 1086–1103 (2006)
7. Jansons, K.M., Alexander, D.: Persistent angular structure: new insights from diffusion magnetic resonance imaging data. *Inverse Problems* 19, 1031–1046 (2003)
8. Söderman, O., Jönsson, B.: Restricted diffusion in cylindrical geometry. *J. Magn. Reson. B* 117(1), 94–97 (1995)
9. Pullens, W., Roebroek, A., Goebel, R.: Kissing or crossing: validation of fiber tracking using ground truth hardware phantoms. In: *Proc. ISMRM*, p. 1479 (2007)
10. Jones, D., Horsfield, M., Simmons, A.: Optimal strategies for measuring diffusion in anisotropic systems by magnetic resonance imaging. *Magn. Reson. Med.* 42, 515–525 (1999)
11. Descoteaux, M., Angelino, E., Fitzgibbons, S., Deriche, R.: Regularized, fast and robust analytical q-ball imaging. *Magn. Reson. Med.* 58, 497–510 (2007)
12. Descoteaux, M., Angelino, E., Fitzgibbons, S., Deriche, R.: Apparent diffusion coefficients from high angular resolution diffusion imaging: Estimation and applications. *Magn. Reson. Med.* 56(2), 395–410 (2006)
13. Savadjiev, P., Campbell, J.S., Pike, G.B., Siddiqi, K.: 3D curve inference for diffusion MRI regularization