Abstract: Automatic Plane Adjustment in Surgical Cone Beam CT-volumes

Celia Martín Vicario¹, Florian Kordon^{1,2,3}, Felix Denzinger^{1,2},
Markus Weiten², Sarina Thomas⁵, Lisa Kausch⁵, Jochen Franke⁶, Holger Keil⁷,
Andreas Maier^{1,3,4}, Holger Kunze^{2,1}

¹Pattern Recognition Lab, Universität Erlangen-Nürnberg (FAU), Erlangen

²Siemens Healthcare GmbH, Forchheim

³Erlangen Graduate School in Advanced Optical Technologies (SAOT), Universität Erlangen-Nürnberg (FAU), Erlangen

⁴Machine Intelligence, Universität Erlangen-Nürnberg (FAU), Erlangen

⁵Division of Medical Image Computing, German Cancer Research Center, Heidelberg

⁶Department for Trauma and Orthopaedic Surgery, BG Trauma Center Ludwigshafen, Ludwigshafen

⁷Department of Trauma and Orthopedic Surgery, University Hospital Erlangen, Universität Erlangen-Nürnberg (FAU), Erlangen

Celia.Martin@fau.de

Cone beam computed tomography (CBCT) is used intra-operatively to assess the result of surgery. Due to limitations of patient positioning and the operating theater in general, the acquisition usually cannot be performed such that the axis-aligned multiplanar reconstructions (MPR) of the volume match the anatomically oriented MPRs. This needs to be corrected manually, which is a time-consuming and complex task and requires the surgeon to interact with non-sterile equipment. To this end, this study investigates a fully-automatic solution to directly regress the standard plane parameters from a CBCT volume of the calcaneus and ankle regions. A PoseNet convolutional neural network (CNN) is adapted and trained, comparing a 6D-, Euler angle- and quaternion-based approach to represent the plane rotation [1]. In addition, a cost function is optimized to incorporate orientation constraints. The best-performing CNN – which uses the 6D representation – estimates the plane normal with a median accuracy of 5°, the in-plane rotation with a median accuracy of 6°, and the position with a median accuracy of 6 mm. The inference time is less than 0.05 s.

References

1. Martin Vicario C, Kordon F, Denzinger F, et al. Automatic plane adjustment of orthopedic intraoperative flat panel detector CT-volumes. Proc MICCAI. 2020; p. 486–495.