$$C/B = 0.438$$

By calculation,

$$\begin{aligned}\frac{C}{B} &= 1 - \left\{ 1 - \frac{9.98 \cdot 2}{\pi 10(2 \cdot 5^2 + 5^2)} \right\}^{\frac{1}{2}} \\ &= 1 - (1 - \frac{9}{10})^{\frac{1}{2}} \\ &= 1 - 0.5624 = 0.4376\end{aligned}$$

With three linear measurements and a simple volume calculation, the chart should provide a fair approximation to the concentration-gradient curve required.

Eqn. 1, used above for cylinders, shows that C/B varies with the height of the instantaneous liquid level to the height of the initial level. Cylinders had been considered; but clearly this is applicable to right prisms, hence widening the combinations of vessels that could be utilised, e.g. cuboid to cuboid, triangular prism to rectangular tank etc. In this event a^2 would be the sectional area of the second vessel and b^2 the sectional area of the first vessel.

Acknowledgment—The chart was produced by the programming services on the ICL 1903A installation at the Polytechnic, Wolverhampton.

References

- ALM, R. S., TISELIUS, A. and WILLIAMS, R. J. P. (1952) Gradient elution analysis. *Acta Chem. Scand.* **6**, 826.

BAKER, R. G., ELLIS, D. A., HAWLEY, F. J., TAYLOR, S. N. and VOKES, B. J. (1969) A concentration gradient problem. *Math. Gaz.* **53** (386), 382.

BROWN, E. G. (1962) The acid soluble nucleotides of mature pea seeds. *Biochem. J.* **85**, 633.

HURLBERT, R. B., SCHMITZ, H., BRUMM, A. F. and POTTER, V. R. (1954) Nucleotide metabolism II chromatographic separation of acid-soluble nucleotides. *J. Biol. Chem.* **209**, 23.

WREN, J. J. (1963) A simple general method for gradient elution using eluents of unequal density. *J. Chromatography* **12**, 32.

R. G. BAKER and D. A. ELLIS

Midland Centre for Neurosurgery and Neurology
Smethwick, Staffs., England

S. N. TAYLOR

Department of Mathematics and Computing
The Polytechnic
Wolverhampton, Staffs., England

B. J. VOKES

Department of Mathematics and Computing
The Polytechnic
Leeds, Yorks., England

Erratum

TOLL, M. O. (1972) An isolated transient-free solid-state a.c. power switch. *Med. & Biol. Engng.* **10**, 532–535.

In the above paper, Fig. 2 should have appeared as shown here.

Fig. 2 Typical response of a.c. power switch as related to an input trigger pulse. Horizontal scale is 20 ms/division
(a) input-trigger pulse (10 V/division)
(b) triac voltage (200 V/division)
(c) load voltage (200 V/division)