A COMPARISON BETWEEN THE BAIN AND MAGILL ANAESTHETIC SYSTEMS DURING SPONTANEOUS BREATHING

M.J. UNGERER

A CO-AXIAL VERSION of the Mapleson D anaesthetic system was described by Bain and Spoerel in 1972.¹ The main feature of the Bain circuit was the incorporation of the fresh gas inflow tubing inside the exhalation limb. Their results suggested that a fresh gas inflow of 70 ml/kg would be adequate to maintain normocarbida in the majority of patients during controlled ventilation, provided that the minute volume is high enough.² These results have been confirmed by Henville and Adams.³

It has also been claimed that the same fresh gas flow which will maintain normocarbida in patients on controlled ventilation will also maintain a normal P_{aCO_2} in patients breathing spontaneously if respiratory depression is avoided.⁴ However, the Mapleson D system has been shown to be much less efficient than the Mapleson A (Magill) system during spontaneous breathing. A fresh gas flow exceeding twice the patient's minute volume has been suggested.⁵

This study was devised in an effort to further investigate flow requirements by comparing the efficiency of the Bain system with that of the Magill system in conscious subjects breathing spontaneously under identical conditions.

METHODS

The experiment was performed in ten healthy conscious volunteers. A conventional Magill attachment was compared with a Bain system as modified by Henville and Adams.⁶ This modification incorporates an expiratory valve and reservoir bag attached by an adaptor to the outer tube of the Bain system at the end distant from the patient.

Air was supplied from cylinders through pre-calibrated rotameters. The seated subject, with a nose-clip applied, breathed air through a rubber mouthpiece from one of the systems. Both systems were evaluated at the same session.

Expired minute volume was measured continuously with a pre-calibrated direct-reading

Wright electronic spirometer. Gas was taken intermittently from a point close to the patient's mouth and analyzed for carbon dioxide concentration on an infra-red analyzer (Godart Capnograph) and the signals were recorded. To minimize the effects of the withdrawn gas on the behaviour of the systems, samples were not taken continuously, but only for a short time during each observation period. Rate of withdrawal of gas was 200 ml/minute.

Initially, the flow rate was set at 150 ml/kg for several minutes to establish a baseline. Thereafter, the flow rate was reduced stepwise by 10 ml/kg at four-minute intervals until the fresh gas flow was reduced to 50 ml/kg. Observations were taken at the end of each four-minute period.

All subjects acted as their own controls and the results were analyzed statistically using a paired Student's t-test.

RESULTS

The appearance of carbon dioxide in the inspired gas was taken as evidence of rebreathing. No rebreathing was evident with the Magill system until the flow was reduced below approximately 80 ml/kg. However, rebreathing could be demonstrated with the Bain system at all the fresh gas flow rates used in this study. Recordings of the carbon dioxide concentration in one subject is shown in Figure 1.

FIGURE 1 Recordings of carbon dioxide concentration in one subject:

FGF = fresh gas flow rate in ml/kg

V_E = expired minute volume in litres/minute.

TABLE I
MEAN EXPIRED MINUTE VOLUMES (LITRES/MINUTE) IN TEN SUBJECTS BREATHING AIR THROUGH A BAIN OR A MAGILL SYSTEM. P VALUES (PAIRED T-TEST) REFER TO MEAN VALUES

Fresh gas flow ml/kg	Bain		Magill		P
	Mean	SD	Mean	SD	
150	9.44	2.82	8.98	2.94	n.s.
130	11.42	2.75	8.60	2.41	<0.025
110	11.94	2.44	8.78	2.27	<0.01
100	13.61	2.32	8.89	2.43	<0.005
90	14.8	2.96	9.10	2.40	<0.0005
80	16.55	3.99	9.50	2.60	<0.0005
70	18.50	3.69	10.45	2.74	<0.0005
60	21.40	4.40	12.80	3.43	<0.0005
50	24.33	4.18	15.88	6.11	<0.01

TABLE II
MEAN END-TIDAL CARBON DIOXIDE CONCENTRATIONS IN TEN SUBJECTS BREATHING AIR THROUGH A BAIN OR A MAGILL SYSTEM. P VALUES (PAIRED T-TEST) REFER TO MEAN VALUES

Fresh gas flow ml/kg	Bain		Magill		P
	Mean	SD	Mean	SD	
150	6.48	0.55	5.92	0.67	<0.025
130	6.61	0.54	5.99	0.68	<0.01
110	6.80	0.61	5.93	0.78	<0.0025
100	6.82	0.53	6.03	0.88	<0.0025
90	6.81	0.58	6.09	0.76	<0.0025
80	6.92	0.62	6.09	0.69	<0.0005
90	7.21	0.68	6.31	0.63	<0.0005
60	7.45	0.72	6.42	0.59	<0.0005
50	7.72	0.88	6.67	0.66	<0.025

The mean expired minute volumes of all subjects are shown in Table I. The larger minute volumes when breathing through a Bain system as compared with the Magill circuit were highly significant at all flow rates except at 150 ml/kg. Mean end-tidal carbon dioxide concentrations of the subjects are shown in Table II. These values were significantly higher when breathing through a Bain circuit at all gas flow rates.

The expired minute volumes when no rebreathing took place were higher than the accepted normal values for subjects at sea level. However, this study was conducted at an altitude of 1,422 metres and the values are within normal limits for subjects at high altitudes.⁷

DISCUSSION

The appearance of carbon dioxide in the inspired air, a rise in end-tidal carbon dioxide concentration and an increase in expired minute volume are indices that rebreathing is occurring.

This investigation supports the view that the Magill system is highly efficient during spontaneous breathing. With this system, rebreathing became apparent only at a fresh gas flow rate of approximately 80 ml/kg. These results confirm the findings of other reports.^{8,9}

In comparison, the Bain system is less efficient with regard to fresh gas economics during spontaneous breathing. Rebreathing was evident at all flow rates used in this study (150 ml/kg and below). Mean expired minute volume was 4.9 per cent higher than with the Magill system at a flow rate of 150 ml/kg and increased progressively to 40 per cent higher at a flow rate of 60 ml/kg. Mean end-tidal carbon dioxide concentration increased from 6.48 per cent to 7.45 per cent with the Bain circuit as compared to an increase from 5.92 per cent to 6.42 per cent with the Magill circuit during the same period.

Since the carbon dioxide production is lower in the anaesthetized patient it may be assumed that lower fresh gas flow would be required to prevent

rebreathing during anaesthesia. However, it was shown in a recent case report that considerable rebreathing took place with the Bain circuit in an anaesthetized patient breathing spontaneously. A much higher minute volume than was the case with the Magill circuit was necessary to maintain an acceptable Pa_{CO_2} level.¹⁰

Nevertheless, the Bain circuit has been safely used in many patients breathing spontaneously.¹⁻⁴ This may be explained by the fact that the response to carbon dioxide remains adequate during anaesthesia and that carbon dioxide may be kept within acceptable limits through active hyperventilation. The question whether the additional effort and work of breathing demanded from the patient is deleterious remains to be answered. The conventional view has always been that minimal rebreathing should be permitted in spontaneously breathing patients. However, it has recently been suggested that the administration of a low concentration of carbon dioxide may be innocuous during spontaneous breathing under anaesthesia and may even be useful to support ventilation and circulation and may reduce cerebrovascular vasoconstriction in hypertensive patients.¹¹

Other studies^{2,3} have demonstrated that the Bain circuit may be used safely and economically during controlled ventilation. Although clinical experience has shown it to be safe in many patients breathing spontaneously, in view of the results of this study its universal use in spontaneously breathing patients, especially with low fresh gas flow rates, cannot be commended. In certain situations such as increased intracranial pressure and chronic lung disease, its use may be regarded as hazardous during spontaneous breathing.

SUMMARY

The efficiency of the Bain system has been compared with that of the Magill system in ten conscious subjects breathing spontaneously. Air was supplied at fresh flow rates of 150 ml/kg and decreased stepwise at four-minute intervals until a flow of 50 ml/kg was attained. Expired minute volume and end-tidal carbon dioxide concentrations were measured.

No rebreathing could be demonstrated with the Magill system at flow rates above approximately 70 ml/kg. In contrast, rebreathing was evident at all flow rates with the Bain system.

It is concluded that acceptable carbon dioxide levels during spontaneous breathing with the Bain circuit can only be maintained by considerable active hyperventilation when using flow rates of 150 ml/kg and less.

RÉSUMÉ

Les systèmes de Bain et de Magill ont été comparés chez dix sujets conscients en respiration spontanée. Les débits de gaz frais de 150 ml/kg/minutes étaient diminués en palier aux quatre minutes, jusqu'à un niveau de 50 ml/kg. On mesurait le volume expiré-minute ainsi que la CO_2 de fin d'expiration.

On n'a pu mettre en évidence de rebreathing avec le système de Magill lorsque les débits de gaz frais étaient au-dessus de 70 ml/kg/minute. Par contre, un rebreathing était évident à tous les débits administrés avec le système de Bain.

Nous concluons qu'avec des débits de gaz frais inférieurs à 150 ml/kg avec un circuit de Bain, des taux acceptables de PCO_2 ne peuvent être obtenus qu'au coût d'une hyperventilation active et considérable.

REFERENCES

1. BAIN, J.A. & SPOEREL, W.E. A streamlined anaesthetic system. *Can. Anaesth. Soc. J.* 19: 426 (1972).
2. BAIN, J.A. and SPOEREL, W.E. Flow requirements for a modified Mapleson D system during controlled ventilation. *Can. Anaesth. Soc. J.* 20: 629 (1973).
3. HENVILLE, J.D. & ADAMS, A.P. The Bain anaesthetic system. *Anaesthesia* 31: 247 (1976).
4. SPOEREL, W.E. & BAIN, J.A. Correspondence. *Can. Anaesth. Soc. J.* 22: 626 (1975).
5. SYKES, M.K. Rebreathing circuits. *Brit. J. Anaesth.* 40: 666 (1968).
6. HENVILLE, J.D. & ADAMS, A.P. A co-axial breathing circuit and scavenging value. *Anaesthesia* 31: 257 (1976).
7. GOLDMAN, H.I., ZWI, S., LEVIN, N.W., RUBENSTEIN, A.H., STABLES, D.P., VAN AS, A.W.W., & ELLIOTT, G.A. Normal values for arterial blood pH, carbon dioxide tension and content, at an altitude of 5,750 feet. *South Afr. Med. J.* 4: 97 (1967).
8. KAIN, M.L. & NUNN, J.F. Fresh gas economics of the Magill circuit. *Anesthesiology* 29: 964 (1968).
9. NORMAN, J., ADAMS, A.P., & SYKES, M.K. Rebreathing with the Magill attachment. *Anaesthesia* 23: 75 (1968).
10. MANSWELL, W.H. Spontaneous breathing with the Bain circuit at low flow rates: a case report. *Can. Anaesth. Soc. J.* 23: 432 (1976).
11. DOBKIN, A.B., BYLES, P.H., AFRICA, B.F., & LEVY, A.A. A comparison of eleven general anaesthetics administered with 7.5 per cent carbon dioxide during spontaneous breathing. *Can. Anaesth. Soc. J.* 23: 408 (1976).