LETTER TO THE EDITOR

DEAR SIR:

A hazard has recently been experienced with the Patient Air Intake Valve of the Cape Ventilator. The expiratory hose was disconnected from its port in the ventilator cabinet, in order to drain condensed water from the hose. It was then reconnected to the ventilator, and the patient was noted to be agitated and the pressure gauge indicated a sustained positive inspiratory pressure of more than 60 centimeters of water. The patient was disconnected from the ventilator and the machine checked. It was discovered that the expiratory hose had been connected to the air inlet of the Patient Air Intake Valve, and consequently while inspiration continued, expiration could not occur. The Patient Air Intake Valve has now been modified by welding a bar across the air inlet of the valve. It is now impossible to insert the expiratory hose into this port.

This hazard would appear to merit dissemination among Canadian users of the ventilator. The Company is intending to insert a letter in the British literature.

F.R.H. WRIGLEY, M.B., B.S.

ERRATA

MALIGNANT HYPERTHERMIA: AN INVESTIGATION OF FIVE PATIENTS

BEVERLEY A. BRITT, M.D., F.R.C.P. (C), WERNER KALOW, M.D., ALAN GORDON, M.D.,
JOHN G. HUMPHREY, M.D., F.R.C.P. (C) AND N. BARRY REWCASTLE, M.D., F.R.C.P. (C)

Canadian Anaesthetists' Society Journal 20: 431, July 1973.

TABLE XII – The Title should read

“Values Represent Calcium Uptake Into the Sarcoplasmic Reticulum Expressed as $\mu\text{M}\text{Ca}^{45}/\text{mg Protein}/5 \text{ Minutes}$.”

FIGURE 10 – The Legend should read

“Abscissa indicates concentration of halothane expressed as volume percent added to a homogenate of sarcoplasmic reticulum. Ordinate indicates uptake into sarcoplasmic reticulum of Ca^{45}Cl in $\mu\text{M}/\text{mg protein}$ during five minutes.”

PHANTOM LIMBS AND THE BODY SCHEMA

PHILIP R. BROMAGE, F.F.A.R.C.S. AND RONALD MELZACK, PH.D.

Canadian Anaesthetists' Society Journal 21: 267–272.

Page 268, Line 6 – Delete “t” from inducted to read “induced”

Page 272, Line 10 – Insert “u” to read “nouminal”