significant advantages over the top-up technique to mother, foetus, nurse, obstetrician and anaesthetist. The most important of these is the safety of the technique.

Hilary J.C. Taylor, M.D., FRCP (C) FFARCS
Scarborough Centenary Hospital,
Scarborough, Ontario.

REFERENCES

1. DAVIES, A.O. & FETTES, I.W. A simple safe method for continuous infusion epidural analgesia in obstetrics. *Can. Anaesth. Soc. J.* 28: 484-487 (1981).
2. RALSTON, D.H. & SHNIDER, S.M. The fetal and neonatal effects of regional anesthesia in obstetrics. *Anesthesiology* 48: 34-64 (1978).
3. CRAWFORD, J.S. Obstetrics, analgesia and anesthesia. *Br. J. Anesth.* 49: 19-23 (1977).
4. COUSINS, M.D. & BRIDENBOUGH, P.O. *Neural blockade 1st ed.* Philadelphia J.B. Lippincott Co. p. 467 (1980).
5. BROMAGE, P.R. *Epidural analgesia. 1st ed.* Philadelphia: WB Saunders Co. pp. 235-236 (1978).
6. SMITH, R.N. Invasive pressure monitoring. *Amer. J. of Nursing* 78: 1514-1521 (1978).

MERA F AND BAIN CIRCUITS

SIR,

Re: *Resistance to Breathing*

While appreciating the description of carbon dioxide responses and elimination during anaesthesia using the Bain and Mera F circuits (Byrick, R.J., Janssen, E. and Yamashita, M., *Can. Anaesth. Soc. J.*, 28: 321, 1981) we believe that insufficient attention has been given to the measurable resistance. The Mera F is a valveless system, but the Bain has an exhaust valve on the expiratory limb which adds considerably to resistance in clinical use. This should, therefore, be included when comparing one system with another. We presume that the figure, intended, of 0.004 kPa (0.04 cm H₂O) refers to the Bain outer pathway alone, as our measurements of the whole expiratory pathway in two systems with valve open, at 20 litres per minute, were 0.196 kPa (2.0 cm H₂O) and 0.181 kPa (1.85 cm H₂O). When circle valves and an absorber are attached to the Mera F as recommended for spontaneous ventilation, their resistance must be taken into account and will raise the figure above that of 0.025 kPa (0.25 cm H₂O) given for this system. Moreover, the effect of the fresh gas inflow is to add further

resistance to breathing, as shown for Bain systems.¹

Resistance of a single Heidbrink valve is an extremely unreliable guide against which to make comparisons. In a study of these "standard" valves,² the mean resistance of 70 valves in use was 0.267 kPa (2.72 cm H₂O) at 20 litres air per minute, with one standard deviation of 0.180 kPa (1.83 cm H₂O) and at 30 litres per minute, as published, was 0.318 ± 0.175 kPa (3.24 ± 1.78 cm H₂O). Thirty new valves at 30 litres per minute measured 0.397 ± 0.100 kPa (4.05 ± 1.02 cm H₂O). A total of 54 out of the 100 valves were found to have a resistance exceeding the limit suggested by Nunn.³

Inspiratory resistance was not mentioned. With a Bain system, the inner tube for fresh gas supply has a very high resistance which patients are not expected to overcome. In the Mera F when connected to a circle attachment "the patient can then inspire gas only from the inner tube" and the dimensions of this tube are thus important. Both internal diameter and length affect the resistance to breathing and may be the reason why the length as supplied is inconveniently short.

M.R. Nott, Consultant,
St. Richard's Hospital,
Chichester, Sussex,
J. Norman, PhD, Professor,
Department of Anaesthetics,
University of Southampton, England

REFERENCES

1. ADAMS, A.P. & SALT, R.H. The resistance of the Bain and Penlon co-axial circuits. *Anaesthesia* 31: 562 (1976).
2. NOTT, M.R. & NORMAN J. Resistance of Heidbrink-type expiratory valves. *British Journal of Anaesthesia* 50: 477 (1978).
3. NUNN, J.F. *Applied Respiratory Physiology*, second edition, London, Butterworths, 103 (1977).

TO THE EDITOR,

We appreciate the remarks of Drs. Nott and Norman concerning our comparison of the re-breathing characteristics of the Bain and Mera F co-axial circuits.¹ They have very appropriately indicated that any discussion of the increased resistance to breathing for anaesthetized patients caused by different circuits must include the entire system (valves, absorber and tubing). However, the main purpose of our paper was to