We conclude that the Laryngeal Mask Airway[®], with modification, would be as effective as the Williams Airway Intubator[®] in fiberoptic laryngoscopy for the purpose of tracheal intubation.

The modifications which we would recommend are enlargement of the diameter of the Laryngeal Mask Airway[®] so that it could accommodate a larger than 6 mm endotracheal tube and its shortening to facilitate removal of the Laryngeal Mask Airway[®] over the endotracheal tube once the trachea has been intubated.² We have used such a device with very good results and would encourage the manufacturer of the Laryngeal Mask Airway[®] to look at modifying this very useful anaesthetic device³ for this purpose.

Allison Crichlow MD
Rock Locken MD FRCPC
Joanne Todesco MD FRCPC
Department of Anaesthesia
Foothills Hospital
University of Calgary

REFERENCES

- 1 Loken RG, Moir C. The Laryngeal Mask Airway as an aid to blind orotracheal intubation: use and removal (in press).
- 2 Brian AIF. Three cases of difficult intubation overcome by the Laryngeal Mask Airway. *Anaesthesia* 1985; 40: 353–5.
- 3 Maltby JR, Loken RG, Watson CN. Clinical appraisal of the Laryngeal Mask Airway. *Can J Anaesth* 1990; 37: S108.

Regurgitation and the laryngeal mask

To the Editor:

Your editorial and the report by Nanji and Maltby^{1,2} on the laryngeal mask (LM) raise important issues.

The patient described was unsuitable for the LM. Gastric emptying is delayed by opiates and trauma, vomiting raises the suspicion of ileus and Alzheimer patients are unreliable historians.

The lack of published mortality with the LM after more than four years in the UK suggests it is safe to use in fasting patients and its role in the difficult airway is firmly established.³

However, we do need to reduce the chances of regurgitation, recognise its occurrence and treat it appropriately. Instructions for reducing the risks in fasting patients are given in the Instruction Manual p 35, 39.⁴

H2 blockers and metoclopramide are wise in the obese. Gross obesity is normally a contra-indication to using the LM. *Induction*: Avoid etomidate and any agent known to be associated with vomiting. A small dose of droperidol is useful with short-acting opiates. *Maintenance*: Avoid inadequate anaesthesia. *Recovery*: Leaving the LM inflated and in place, maintain adequate anaesthesia until the patient is breathing spontaneously and positioned correctly for recovery.

Recognising regurgitation

Fluid may suddenly appear inside the LM tube. This should be watched for. The hypopharynx is filled by the inflated LM, so that even 7–10 ml is immediately apparent, giving early warning of regurgitation.

Straining during anaesthesia may produce high intra-abdominal pressures, overcoming the limited resistance offered by the distal LM cuff which lies against the upper oesophageal sphincter. Gastric fluid may thus enter the mask interior where it can be aspirated during inspiration or blown up the LM tube during expiration. The glottis may close limiting the amount of aspiration but causing stridor.

Treatment

- Head down tilt
- Disconnect circuit to allow fluid to drain
- Reconnect circuit to give 100% oxygen
- Avoid forceful ventilation attempts
- Deliver *small* tidal volumes
- Deepen anaesthesia – *iv* propofol works well
- Perform suction through the LM (fibrescope is best), avoid high vacuum/prolonged suction.

Further treatment will depend on the extent of aspiration but chest x-ray is mandatory. Avoid removing or deflating the LM if possible – it is likely to be holding back more fluid.⁵ Laryngeal mask removal followed by paralysis and intubation create new hazards including loss of airway potency, obstruction by biting, further aspiration, loss of cough reflex and possible bronchospasm. Of four cases receiving the above treatment without intubation, none required admission to an intensive care unit.⁶

A.I.J. Brain
Northwick Park Hospital
London, England

REFERENCES

- 1 Fisher JA, Ananthanarayan C, Edelist G. Role of the laryngeal mask in airway management. *Can J Anaesth* 1992; 39: 1–3.
- 2 Nanji GM, Maltby JR. Vomiting and aspiration

pneumonitis with the laryngeal mask airway. *Can J Anaesth* 1992; 39: 69-70.

- 3 Editorial: Laryngeal mask airway. *The Lancet* 1991; 338: 1046-7.
- 4 The Intavent Laryngeal Mask Instruction Manual, 2nd Ed. Available from: Vitaid Ltd, PO Box 123, Stn St, Toronto, Ontario M5M 4L6.
- 5 Griffin RM, Hatcher IS. Aspiration pneumonia and the laryngeal mask airway. *Anaesthesia* 1990; 45, 12: 1039-40.
- 6 Brain AIJ. The laryngeal mask and the oesophagus. *Anaesthesia* 1991; 46: 701-2.

Early detection of airway obstruction with a capnographic probe attached to an oxygen mask

To the Editor:

Capnometry and pulse oximetry are effective in preventing anaesthetic mishaps^{1,2} and ensuring adequate ventilation³ in sedated patients. We report a case in which airway obstruction was detected by capnography, but not by pulse oximetry.

A 71-yr-old man, scheduled to undergo inguinal hernia repair, was continuously monitored with ECG, pulse oximeter (SpO₂), and capnography. The capnographic probe (Capnomac, Datex) was inserted into and attached to an oxygen face mask near the nostrils (Figure 1). After establishment of epidural anaesthesia, midazolam 5 mg and butorphanol 1 mg were given *iv*. One minute later, capnographic waves were abruptly distorted and flattened (Figure 2). The patient was apnoeic for 25 sec.

FIGURE 1 Capnographic probe attached inside the oxygen face mask near the nostrils of the patient.

FIGURE 2 Capnographic tracings. After sedatives were given, CO₂ value decreased markedly and reached baseline level. The patient was apnoeic for 25 sec.

During this period, SpO₂ (99%) and haemodynamic variables remained unchanged. Immediately after moving the lower jaw to open the airway, the capnographic waves reappeared, indicating a patent upper airway of the patient.

In our monitoring system no additional devices are required. Although a leak around the face mask may cause dilution of the expiratory CO₂, we believe that our technique can monitor trends of CO₂ output, and thus aid in the early detection of airway obstruction.

Shinichi Inomata MD

Toshiaki Nishikawa MD

Department of Anaesthesiology

Institute of Clinical Medicine

University of Tsukuba

Tsukuba City, Ibaraki 305

Japan

REFERENCES

- 1 Tinker JH, Dull DL, Caplan RA, Ward RJ, Cheney FW. Role of monitoring devices in prevention of anesthetic mishaps: a closed claims analysis. *Anesthesiology* 1989; 71: 541-6.
- 2 Bailey PL, Pace NL, Ashburn MA, Moll JWB, East KA, Stanley TH. Frequent hypoxemia and apnea after sedation with midazolam and fentanyl. *Anesthesiology* 1990; 73: 826-30.
- 3 Goldman JM. A simple, easy, and inexpensive method for monitoring ETCO₂ through nasal cannulae. *Anesthesiology* 1987; 67: 606.

Technology transfer and monitoring practices

To the Editor:

The penultimate stage in the life of a successful monitoring device is its routine use by clinicians. This is