Some related expenses which do not feature in this Table include: intravenous drugs;¹² choice of breathing circuit;¹³ replenishing soda lime absorbers;¹¹ and operating room staff salaries.¹⁴ Analysis of total operating room costs in our hospital for conducting a common surgical procedure (laparoscopic cholecystectomy) shows that surgical consumables contribute 38%, while staff wages costs 53%. Anaesthetic costs (drugs and depreciation of monitoring equipment) constitute the remaining 9%. In particular, isoflurane delivery makes up less than 1% of the total cost!

Using our calculations, accurate estimates of inhalational agent and carrier gas costs have allowed us to keep the cost of delivering anaesthesia in better perspective in today's cost-conscious environment. Our formula can also be used when newer, and inevitably more expensive, agents become available, to allow rational discussion of their economic impact.

Julian Loke MB BS (Hons) B Med Sc (Hons)
William A.J. Shearer MB BS FANZCA
Department of Anaesthesia
Dandenong Hospital
Victoria, Australia

REFERENCES

- 1 Dion P. The cost of anaesthetic vapours. *Can J Anaesth* 1992; 39: 633.
- 2 Tarpey J, Lawler PG. Volatile agent use: perception and practice – a survey of agent use over a 3-year period. *Anaesthesia* 1989; 44: 596–9.
- 3 Rigby R. Closed system use of isoflurane and enflurane. *Anaesthesia* 1990; 45: 500.
- 4 Jones RM. Desflurane and sevoflurane: inhalational anaesthetics for this decade? *Br J Anaesth* 1990; 65: 527–36.
- 5 Wood M. Inhalational anesthetic agents. In: Wood M, Wood AJJ (Eds.). *Drugs and Anesthesia: Pharmacology for Anesthesiologists*, 2nd ed., Baltimore: Williams & Wilkins, 1990; 255–70.
- 6 Rampil IJ, Lockhard SH, Zweiss MS, et al. Clinical characteristics of desflurane in surgical patients: minimum alveolar concentrations. *Anesthesiology* 1991; 74: 429–33.
- 7 Katoh T, Ikeda K. The minimum alveolar concentration (MAC) of sevoflurane in humans. *Anesthesiology* 1987; 66: 301–3.
- 8 Scheller MS, Partridge BL. MAC of sevoflurane in humans and the New Zealand white rabbit. *Can J Anaesth* 1988; 35: 153–6.
- 9 Matjasko J. Economic impact of low-flow anesthesia. *Anesthesiology* 1987; 67: 863–4.
- 10 Virtue RW. Comparison of cost of high and low flows of anaesthetic agents. *Can Anaesth Soc J* 1981; 28: 182–4.
- 11 Christensen KN, Thomsen A, Jørgensen S, Fabricius J. Analysis of costs of anaesthetic breathing systems. *Br J Anaesth* 1987; 59: 388–91.
- 12 Bengtson JP, Sonander H, Stenqvist O. Comparison of costs of different anaesthetic techniques. *Acta Anaesthesiol Scand* 1988; 32: 33–5.
- 13 Humphrey D, Downing JW, Brock-Utne JG. Cost of anaesthesia. *BMJ* 1983; 286: 800.
- 14 Nelson AH, Hanson RL. OR cost survey – results of a nationwide study. *Journal of the Association of Operating Room Nurses* 1985; 41: 569–77.

Erratum

Eisen LB, Fisher JA. Cost of soda lime (Correspondence). *Can J Anaesth* 1993; 40: 81–3.

Please note that the following material was incorrectly displayed, the amendments are underlined:

The cost of soda lime as a function of flow is therefore;

$$\begin{aligned} \text{CS (cents)} &= 1.965 \dot{V}\text{CO}_2 T \left(1 - \frac{F}{\dot{V}} \right) \\ &= -1.965 \left(\frac{\dot{V}\text{CO}_2}{\dot{V}} \right) \text{TF} + 1.965 \dot{V}\text{CO}_2 T \end{aligned}$$

Note that for $F \geq \dot{V}$, no soda lime is needed, so CS = 0.

Converting Dr. Dion's equation for halothane cost from dollars to cents, we obtain the cost of halothane (CH) as a function of fresh gas flow, where P is the vaporizer concentration of halothane in percent:

Also, in the captions for Figures 1 and 2 the symbols for "minute ventilation" should appear as ($\dot{V}$).