market and was not detected by our normal QC, production and QA inspection methods. We have production quality controls in place, as product is inspected during manufacture and at the end of production by packagers. Following our 100% in-line quality inspection, our Quality Assurance department inspects each batch using the G2 statistical sample plan as per Mil Standard 105.

Since this occurrence, we have reviewed our methods of manufacturing and assembly and consider that an error was made in inspection which was not detected. We have implemented a second 100% inspection station in production prior to packaging to ensure all future products meet specification.

K.E. Little
President
SIMS Canada Ltd.

Erratum

J.G. Ramsay. Continuous cardiac output: myth or reality? / La mesure continue du débit cardiaque: mythe ou réalité? Can J Anaesth 1993; 40: 98–102.

Please note that the second paragraph on page 98 should read as follows:

Recently several investigators have demonstrated that elevated cardiac output and oxygen delivery in critically ill surgical patients,¹ patients with adult respiratory distress syndrome² and patients with septic shock,³ is associated with improved survival. In two of these studies,^{1,3} cardiac output measurements were used to guide a therapy (fluids and inotropes) which then appeared to improve outcome. It seems likely that the demand for reliable, frequent or continuous cardiac output measurement will continue to increase.