over to fiberoptic laryngoscope blades to facilitate reliability when subjected to high level sterilization.

James Fletcher MB BS MRCP FRCA
Department of Anesthesia, SUNY at Buffalo
Children's Hospital of Buffalo
219 Bryant Street
Buffalo
NY 14222

REFERENCES

- 1 *Tousignant G, Tessler MJ.* Light intensity and area of illumination provided by various laryngoscope blades. *Can J Anaesth* 1994; 41: 865-9.
- 2 *Fletcher JE, Summerfield RJ, Bowen DJ.* Audit of Laryngoscope Light Intensity. Presented at Wessex Regional Medical Audit Conference, England. April, 1992.

Erratum

Bernards CM. Human dura mater permeability (Letter). *Canadian Journal of Anaesthesia* 1994; 41: 1125.

Please note that in paragraph numbered 1, the phrase "highly permeable" should have read "highly impermeable." We apologize for the error.