Mapleson D circuit mechanics confirm the need to differentiate between the patient and the breathing system when considering CO₂ elimination.

Alveolar ventilation remains the only way to remove CO₂ from the patient; whether or not CO₂ is removed from the Bain or any other version of the Mapleson D system depends upon the relationship between fresh gas flow and the duration of the expiratory phase. Whether the ventilation is spontaneous or controlled is irrelevant.

If this flow/time relationship is appropriate, V_F may approach V_E but, if either expiratory time or fresh gas flow is too low, rebreathing will occur and *must* hinder CO₂ washout. Spoerel may regard this as a passive phenomenon if he wishes, it certainly is inevitable. It can equally certainly be regarded as a CO₂ load.

Spoerel's own data confirm this view as Figures 4 and 6 (p. 152-153) demonstrate:

- A Pre-naloxone $f = c 10 \cdot \text{min}^{-1}$, $\dot{V}_E = 70 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$, $F_{ET}\text{CO}_2 = 7.5$, $F_{i}\text{CO}_2 = 0.6$, $\dot{V}_F = 100 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$, $t_{\text{Exp}} c 3.5 \text{ sec}$.
- B Post-naloxone $f = 26 \cdot \text{min}^{-1}$, $\dot{V}_E = 133 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$, $F_{ET}\text{CO}_2 = 6.75$, $F_{i}\text{CO}_2 = 3.5$, $\dot{V}_F = 100 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$, $t_{\text{Exp}} < 1 \text{ sec}$.
- A Alveolar hypoventilation - $\dot{V}_A c 50 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$, $F_{ET}\text{CO}_2 7.5$, but at $\dot{V}_F 100 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$ and $t_{\text{Exp}} 3.5 \text{ sec}$, adequate CO₂ washout of the circuit $F_{i}\text{CO}_2 0.6$.
- B Alveolar ventilation - apparently adequate $\dot{V}_A 91 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$ (approximately $2 \times A$) but at $\dot{V}_F 100 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$ and $t_{\text{Exp}} < 1 \text{ sec}$, inadequate CO₂ washout. $F_{ET}\text{CO}_2 6.75$ (despite 80 per cent increase in $\dot{V}_A$).

Is it really possible to say that a V_F of $100 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$ is adequate for CO₂ elimination here or that the $F_{i}\text{CO}_2$ of 3.5 per cent does not contribute to CO₂ retention ($F_{ET}\text{CO}_2 6.75$) and hence is not a CO₂ load?

The Mapleson D system as originally designed by Philip Ayre⁴ was simple in construction and, being valveless, had minimal expiratory resistance, both desirable features for the paediatric and neurosurgical applications he intended. It was never expected to be either economical or to be used with "fixed" fresh gas flow rates.

Mapleson recognized almost 20 years ago that non-rebreathing in this system required correlation with the patient's minute ventilation ($>2 \times$). Since his own data, as well as that of many others,

confirms this, perhaps Dr. Spoerel will recognize it too!

David D. Imrie MB BS
Dept. of Anaesthesia,
Victoria General Hospital,
Halifax, Nova Scotia

REFERENCES

- 1 Mapleson WW. Br J Anaesth 1954; 26: 323.
- 2 Byrick RJ. Can Anaesth Soc J 1980; 27: 96.
- 3 Humphrey D. The Lack, Magill and Bain anaesthetic breathing system: a direct comparison in spontaneously breathing anesthetized adults. Journal of the Royal Society of Medicine 1982; 75: 513-24.
- 4 Ayre P. Lancet 1937; 1: 561.

REPLY

Thank you for the opportunity to reply to Dr. Imrie's letter. I am distressed by the thought that I may have repeatedly contributed to the misunderstandings of the Mapleson D breathing system. It was all done with the good intention to make observations and to report data which should eventually lead to an end of the confusion about which Dr. Imrie complains in his letter. Why is this unjustifiable?

I am not aware that I ever disagreed with Mapleson that a V_E:V_F ratio of about 0.5 was required to avoid rebreathing, as indicated by the presence of CO₂ during inspiration (F_iCO₂). But a ratio has two sides; I showed in my paper that the same ratio (i.e., fresh gas flow [V_F] equals two times minute ventilation [V_E]) can also be achieved and rebreathing be avoided by a suitable depression of V_E with a narcotic.

Furthermore, since the tidal volume varies from breath to breath in spontaneous respiration, the V_E:V_F ratio and also F_iCO₂ change with each breath. This can be readily demonstrated in the capnogram of a conscious person breathing through a Bain system with a V_F of $100 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$ and voluntarily alternating slow normal breathing with short periods of marked hyperventilation (see Figure); the F_iCO₂ changes with the breathing pattern but does not alter the end tidal CO₂.

The association of a raised F_iCO₂ with a near normal PaCO₂ in spontaneous breathing under anaesthesia which I reported, was also seen by Byrick¹ who found no correlation between PaCO₂ and a calculated inspiratory CO₂ load. Meakin and Coates² evaluated spontaneous breathing with the Bain system and remarked: "paradoxically, those patients in whom the degree of rebreathing was greatest had the lowest end-tidal CO₂." If rebreathing in the Bain system would actively contribute to hypercarbia, as assumed by Dr. Imrie, it has yet to be documented; the presently available evidence supports the "dangerous" concept that the F_iCO₂ does not affect CO₂ homeostasis.

The assumed danger of rebreathing led to the demand

FIGURE Capnogram of a conscious volunteer breathing through a Bain circuit with a fresh gas flow of $100 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$. 1. normal quiet breathing; 2. breath holding in inspiration; 3. rapid deep breathing.

that the $\dot{V}_F$ be increased to a level where functional non-rebreathing was achieved. Baracca³ in 1968 was the first to demonstrate that, with controlled ventilation, partial rebreathing was safe and that the $\dot{V}_F$ could be reduced to a level near the patient's volume of alveolar ventilation. The use of a $\dot{V}_F$ well below the resting minute ventilation (i.e., $70 \text{ ml} \cdot \text{kg}^{-1} \cdot \text{min}^{-1}$) has now been widely accepted. I agree with Dr. Imrie that we should discard the idea that the function of the Bain system during spontaneous breathing is fundamentally different from controlled ventilation; the distribution of CO_2 within the Bain system is in principle not different with the two types of ventilation.⁴

Certainly the alveolar ventilation ($\dot{V}_A$) is the only way to remove CO_2 from the patient. This implies that a volume of CO_2 free gas equal to the required $\dot{V}_A$ must reach the alveoli. The only CO_2 free gas is supplied from the anaesthetic machine and, to accommodate the dead space, $\dot{V}_F$ must always be greater than $\dot{V}_A$. Since a volume of gas equal to $\dot{V}_F$ is vented from the Bain system, the volume of gas which is washing CO_2 out of the breathing system must always be greater than the volume of gas which is removing CO_2 from the patient.

There is no evidence that the degree of mixing of fresh gas and expired gas would influence the volume of CO_2 removed from the breathing system. Also the duration of expiration, at a constant alveolar ventilation, does not affect CO_2 removal; reversal of the I:E ratio,⁴ although it did raise the FiCO_2 , did not alter the end tidal CO_2 nor the volume of CO_2 vented from the system.

I believe the difficulties in understanding the Mapleson D system which Dr. Imrie expressed in his letter, originate in the assumption that rebreathing in this system is functionally the same as rebreathing induced by an increased equipment deadspace. While both mechanisms of rebreathing are associated with a raised FiCO_2 , their effect on CO_2 elimination and their potential hazard are different.

W.E. Spoerel MD FRCP(C)
Department of Anaesthesia,
University of Western Ontario,
London, Ontario

REFERENCES

- 1 Byrick RJ, Janssen EG. Respiratory wave form and rebreathing in T-piece circuits: A comparison of enflurane and halothane waveforms. *Anesthesiology* 1980; 53: 371.
- 2 Meakin G, Coates A.L. An evaluation of rebreathing with the Bain system during anaesthesia with spontaneous ventilation. *Br J Anaesth* 1983; 55: 487.
- 3 Baraka A. PaCO_2 control by fresh gas flow during controlled ventilation with a semi-open circuit. *Br J Anaesth* 1967; 41: 527.
- 4 Spoerel W.E. Rebreathing and carbon dioxide elimination with the Bain circuit. *Can Anaesth Soc J* 1980; 27: 357.

Hazard of interchangeable oxygen connectors

To the Editor:

Patient safety is of utmost importance whenever using medical equipment. This is especially important when administering medical gases. The Diameter Index Safety System (DISS) was developed to prevent incorrect attachment of gas-administering equipment to the wrong gas.¹⁻⁴

We would like to give a brief account of an error involving the administration of oxygen while using a system which conformed to DISS.

A middle-aged woman had a tracheostomy because of long-term respiratory failure secondary to a neurological disease. She was able to maintain her own ventilatory function for only short periods during the day and at night was maintained on a Bird respirator. During one episode of daytime spontaneous ventilation she developed acute respiratory difficulty, secondary to a mucus plug.

Her ventilation was immediately assisted by a Hope 1 resuscitation bag which was then connected to wall oxygen because of the patient's increasing cyanosis. At that time she was noted to be pulseless and cardiopulmonary resuscitation was quickly initiated. It was noticed that the Hope 1 bag was difficult to deflate and appeared to be overdistended. Large amounts of gas were heard hissing around the connections and through the tracheostomy site. EKG was showing sinus tachycardia but there was no palpable pulse.