473872 - A NOVEL CLOSED-LOOP SYSTEM TO ADMINISTER PROPOFOL-COMPARISON WITH MANUAL CONTROL

Thomas Hemmerling, MD¹, Samer Charabati, BEng², Benjamin Dubois, cand med³, David Bracco, MD¹, Pierre Mathieu, PhD²

1. Anesthesiology, McGill, Montreal, QC, Canada

2. Institut de Genie Biomedical, Universite de Montreal, Montreal, QC, Canada

3. Medecine, Universite de Montreal, Montreal, QC, Canada

Introduction: The purpose of this project is to assess the performance of a novel, expert-based adaptive closed-loop system for propofol administration using Bispectral Index (BIS).

Methods: After IRB approval and written consent, 50 patients undergoing orthopedic or general surgery are randomly assigned to 2 groups of equal size. After manual propofol induction (1.5 mg/kg), closed loop control was used to maintain anesthesia at a target BIS of 45 (Closed-loop group); in the Control group, propofol was administered manually using a syringe pump by an experienced TIVA-anaesthesiologist in order to maintain a target of 45 as closely as possible. The performance of the system was defined as excellent, good, poor or inadequate, when the BIS was within 10%, between 10 and 20%, between 20 and 30% or above 30% of the target BIS, respectively.

Results: In the Closed-loop group, 11 patients (2 f, 9 m; age 50 ± 16 y; weight 80 ± 9 kg) underwent anesthesia of 178 ± 68 min and received propofol of 120 ± 35 μ g/kg/min.

During 48 %, 30% and 10% of the total anesthesia time, the system showed excellent, good and poor control, respectively. The time from the end of infusion to extubation was 9 ± 2 min. In the Control group, 11 patients (4 f, 7 m; age 65 ± 15 y; weight 70 ± 15 kg) underwent anesthesia of 115 ± 68 min and mean propofol of 105 ± 23 μ g/kg/min.

Excellent, good and poor control were noted during 26%, 32% and 15% of the time. In the Closed-loop group, excellent control of anesthesia occurred significantly more often ($P=0.001$) and inadequate control less often than in the Control group ($P=0.01$).

Discussion: In comparison to manual control, our system seems to be able to maintain anesthesia closer to a given target.

References: NO REFERENCES

Closed-loop system performance

Excellent Control

Poor Control

Good Control

Inadequate Control