An unusual case of laryngeal mask airway obstruction

To the Editor:

A healthy 46-yr-old woman was booked for bunion surgery under general anesthesia. After the *iv* induction of anesthesia, a #4 laryngeal mask airway (LMA) was inserted uneventfully but ventilation was impossible. The LMA was removed and the patient was ventilated easily by facemask. The LMA was noted to have a foreign object in the lumen. Another LMA was inserted and the patient was ventilated without difficulty. No desaturation occurred throughout the incident. The patient made an uneventful recovery.

The LMA was totally obstructed by a cylindrical plastic plug, which we were able to push out the mask end with a stylet (Figure).

A visit to our sterilization department revealed the cause. The object was a plastic cap on the stainless steel posts of the "Utility Rack Insert" (Getinge, Castle corporation). The sterilizing process had recently been changed to mount the LMA vertically; connector end first, onto the plastic capped posts before the wash cycle. When finished the mask was removed – apparently taking the cap with it. It was then packaged in a sterile bag with a transparent window for steam autoclaving. After the incident, the caps were easily removed from the posts and discarded. Sterilizing techniques were changed to prevent a recurrence.

A closer inspection of the LMA would have detected the plug but, if situated near either end, this would be less obvious. It could have been disastrous for the

patient if the object had been forced out through the LMA bars into the patient's airway.

There are other reports of foreign objects in LMA's after sterilization^{1–3} including one where an entire steel cleaning post was found in the LMA.⁴

This report highlights the danger of inventing new processes or using a tool for a purpose it was not intended and the importance of a "pre-flight" inspection of all airway equipment.

Douglas Hardy MD FRCPC
Ottawa, Ontario

References

- 1 Conacher ID. Foreign body in a laryngeal mask airway (Letter). *Anaesthesia* 1991; 46: 164.
- 2 Cameron D, Onslow J. Obstruction of airway equipment (Letter). *Anaesthesia* 2002; 57: 188–9.
- 3 Riley RH, Browning FS. Another foreign body in a laryngeal mask airway (Letter). *Anaesthesia* 1996; 5: 286–7.
- 4 Srikanth K. Yet another foreign body in a laryngeal mask airway (Letter). *Anaesthesia* 2002; 57: 189–90.

Successes and failures with the laryngeal mask airway (LMA) in patients with Treacher Collins syndrome – a case series

To the Editor:

Treacher Collins syndrome (TCS) is a disease known to be associated with difficult airway.^{1,2} Some authors reported that the laryngeal mask airway (LMA) enabled ventilation in TCS patients.^{3,4} However, it remains unclear whether or not the LMA always provides a patent airway in TCS patients. In the present study, clinical and fiberoptic assessment of the positioning of the LMA was performed in TCS patients. After obtaining institutional approval and informed consent, we studied seven consecutive TCS patients scheduled for elective surgery. Anesthesia was induced with propofol, sevoflurane, and/or fentanyl. No muscle relaxants were used. The LMA was inserted in a standard fashion. In all patients, the LMA was inserted into the hypopharynx on the first attempt. When manual ventilation via the LMA produced sufficient tidal volume ($> 8 \text{ mL}\cdot\text{kg}^{-1}$), it was considered an acceptable patent airway. The pressure at which an audible leak occurred was documented. The position of the LMA was ascertained with a fibroscope, and the view was graded as follows: grade 1, only larynx seen; grade 2, epiglottis and larynx

FIGURE Laryngeal mask airway (LMA) mounted on the stainless steel post of the "Utility Rack Insert" prior to sterilization.

TABLE Clinical and fiberoptic assessment of LMA positioning

Case	Sex	Age (yr)	Weight (kg)	Size of LMA	Ventilation via LMA	Leak pressure cm H ₂ O	Fiberoptic findings
1	F	32	40	3	acceptable	18	grade 1
2	M	12	33	3	acceptable	16	grade 3
3	F	12	23	2.5	unacceptable	-	grade 4
4	M	19	43	3	acceptable	15	grade 2
5	F	12	36	3	acceptable	16	grade 1
6	F	9	22	2.5	unacceptable	-	grade 4
7	F	6	20	2	acceptable	15	grade 2

LMA = laryngeal mask airway.

seen; grade 3, epiglottis down-folded or impinging on grille, larynx seen by manipulation of the fibroscope; grade 4, epiglottis down-folded, larynx not seen. The results are shown in the Table. In five of the seven patients, the LMA provided an acceptable patent airway. In the remaining two patients (cases three and six), however, down-folding of the epiglottis caused severe airway obstruction. In case three, establishment of a patent airway via the LMA was unsuccessful despite three attempts at insertion. Subsequently, tracheal intubation succeeded by direct laryngoscopy. In case six, the down-folding of the epiglottis was partially relieved by moving the LMA in an oral direction. However, this correction of LMA placement did not provide an acceptable patent airway. In this case, SpO₂ decreased to 80% after insertion of the LMA; it immediately increased (> 97%) after recovery of spontaneous breathing. At that point, fibroscope-aided tracheal intubation through the LMA was performed.

Despite the small sample size in this study, failure to establish a patent airway with the LMA in two patients shows that the LMA cannot always provide a patent airway in TCS patients. Accordingly, careful consideration is necessary in the elective use of the LMA in TCS patients.

Koichi Takita MD
Shigeaki Kobayashi MD
Masahito Kozu MD
Yuji Morimoto MD PhD
Osamu Kemmotsu MD PhD
Sapporo, Japan

References

- 1 Stone DJ, Gal TJ. Airway management. In: Miller RD (Ed.). *Anesthesia*, 5th ed. Philadelphia: Churchill-Livingstone Inc.; 2000: 1414–51.
- 2 Stoeling RK, Dierdorf SF. *Anesthesia and Co-Existing Disease*, 4th ed. Philadelphia: Churchill-Livingstone Inc.; 2000.
- 3 Bahk JH, Han SM, Kim SD. Management of difficult airways with a laryngeal mask airway under propofol anaesthesia. *Paediatr Anaesth* 1999; 9: 163–6.
- 4 Ebata T, Nishiki S, Masuda A, Amaha K. Anaesthesia for Treacher Collins syndrome using a laryngeal mask airway. *Can J Anaesth* 1991; 38: 1043–5.

Severe bradycardia in a patient undergoing endovascular stent-graft repair for abdominal aortic aneurysm with vena cava balloon occlusion

To the Editor:

Endovascular stent-graft deployment (SGD) for repair of an aortic aneurysm has recently been preferred to aneurysmectomy in certain cases.¹ We encountered severe dysrhythmias in a patient undergoing SGD.

A 73-yr-old male with infrarenal abdominal aortic aneurysm underwent endovascular SGD due to reduced pulmonary function arising from a previous right thoracoplasty. The preoperative electrocardiogram revealed complete right bundle-branch block. Anesthesia was induced with propofol, fentanyl and vecuronium and maintained with sevoflurane and supplemental fentanyl.

A double-balloon catheter was inserted through the femoral vein and its position confirmed under fluoroscopic guidance. Subsequently, a sheath charged with the self-expandable stent was introduced via the left femoral artery.

Hypotension was induced by inflation of the occlusion balloon in the inferior vena cava (IVC) followed by gradual inflation of the balloon in the superior vena cava (SVC). Supraventricular arrhythmia ensued, followed by severe bradycardia (Figure). The mean arterial blood pressure decreased to 30 mmHg. Cardiac massage was performed accompanied by the *iv* administration of atropine (0.5 mg) and methoxamine (3