Ensure that the intensity and pitch of auditory signals allow them to be heard easily by device users.

Consider the effects of ambient noise.

Ensure that the brightness of visual signals is sufficient to be perceived by users working under various conditions of ambient illumination. Also, brightness contrast and colour contrast can help to optimize legibility.

Make labels and displays so that they can be easily read from typical viewing angles and distances.

Symbol size, contrast, color, and display depth are important considerations.

Ensure that the abbreviations, symbols, text, and acronyms placed on, or displayed by, the device are also used consistently in the instructional manual.

They also should correspond to standard nomenclature, if possible.

Design control knobs and switches so that they correspond to the conventions of the user population (as determined by user studies and existing medical device standards).

Arrange and design knobs, switches, and keys in a way that reduces the likelihood of inadvertent activation.

Use colour and shape coding, where appropriate, to facilitate the rapid identification of controls and displays. Colours and codes should not conflict with universal industry conventions.

Space keys, switches, and control knobs sufficiently apart for easy manipulation. This will also reduce the likelihood of inadvertent activation.

Make sure that controls provide tactile feedback.

This site is important to anesthesiologists because it clearly shows how "the application of user interface design principles and participation of healthcare practitioners in design analyses and tests are very important."

The authors also point out that "in addition to increased safety, an added benefit of such practices is the likelihood that good user interface design will reduce training costs to healthcare facilities."

BRIEF OUTLINE

HUMAN FACTORS: AN HISTORICAL PERSPECTIVE

WHY HUMAN FACTORS ENGINEERING IS IMPORTANT

THE USER INTERFACE

Control/Display Arrangement and Design

Device Logic and Microprocessing

Component Installation

Alarms

Other Important Issues

HUMAN FACTORS ENGINEERING

Document Reviews

Exploratory Studies

Analysis

Usability Testing

SPECIAL ISSUES

Initiating a Company Human Factors Program

Advice to Healthcare Facilities

GLOSSARY

FURTHER READING

In conclusion, this document offers much-needed guidance aimed at increasing the understanding of human factors issues in medical device design. The message applies most strongly to technology intended for use in surgical patients and patients in Intensive Care Units. I highly recommend this site as a starting point for all individuals interested in human factors issues in anesthesia technology and its impact on patient safety.

D. John Doyle MD PHD FRCP
Toronto, Ontario

Clinical Video Review

Intubating the Newborn.

(c) 1996 Judith A. Littleford md frcpc

32 min. VHS in NTSC, PAL or SECAM formats.

\$99 USD from American Academy of Pediatrics

(<http://www.aap.org>)

According to the author, this video "is designed for practitioners who assist with, teach, and perform tracheal intubation." In the presentation, relevant anatomy and "the procedural sequence" for intubation are covered in detail using a combination of illustrations, computer animations, video clips and demonstrations using a manikin. Also nicely covered are practical issues like equipment selection and setup, and how best to secure the tube once placed.

The video is narrated by the author in a soft but well-paced and well-articulated manner that shows every sign of a professional level production effort.

The following educational objectives are provided for those individuals who watch the video with a view to having a serious learning experience:

1. Know the anatomical structures of the infant airway,
2. Compare infant and adult anatomy,
3. Select, and conduct a pre-use check of infant intubation equipment,
4. Understand the procedural sequence involved in tracheal intubation, in confirmation of tube placement, and in securing an endotracheal tube,
5. Identify measures that can be employed to improve visualization of the infant glottis,
6. List the indications for intubation of the new born.

After some introductory remarks, the difference between the airways in neonates and adults is covered in detail, with professional quality illustrations and computer-animation sequences. The required equipment for intubation is reviewed. Real views at laryngoscopy are also offered, complete with graphic annotations where appropriate. Indications for intubation are mentioned, including intubation for the purpose of tracheal suction and the management of the neonate with suspected meconium aspiration. A detailed discussion of securing the endotracheal tube after intubation is also presented, and no doubt is based on practical concerns about the hazards of inadvertent extubation.

This is a high-quality video that does a superb job of distilling the essentials of neonatal airway management into a mere 32 min. In particular, the video does a great job at helping individuals meet the stated educational objectives. Every resident in training (and many others) would benefit from this video. Highly recommended.

D. John Doyle MD PHD FRCPC
Toronto, Ontario