Erratum to: Rapid and simple method for the quantification of flavan-3-ols in wine

Maria A. Silva · Isabelle Ky · Michael Jourdes ·
Pierre-Louis Teissedre

Published online: 28 December 2011
© Springer-Verlag 2011

Erratum to: Eur Food Res Technol DOI 10.1007/s00217-011-1628-0

In the original publication, Table 2 has not been published completely as the last line in the first column was found to be omitted. The complete version of Table 2 is given below.

Table 2 Validation parameters determined for the proanthocyanidins liquid chromatography coupled with fluorescence detection method

Compound	Retention time (min)	R^2	Linearity range (mg/L)	Repeatability ($n = 6$), CV%		Intermediate precision ($n = 6$), (%)		Accuracy ^a ($n = 3$), (%)			LOD (mg/L)	LOQ (mg/L)
				0.10 (mg/L)	20 (mg/L)	0.10 (mg/L)	20 (mg/L)	0.1 (mg/L)	5 (mg/L)	20 (mg/L)		
B1	6.4	0.9999	0.004–48.13	0.25	0.13	3.41	3.41	90.2	90.2	90.2	0.002	0.008
B3	6.8	0.9999	0.005–52.12	0.46	0.23	4.13	4.13	89.1	89.1	89.1	0.003	0.009
(+)-Catechin	8.5	0.9999	0.001–40.02	0.21	0.12	3.96	3.96	95.4	95.4	95.4	0.001	0.015
B4	10.4	0.9993	0.004–63.32	1.25	0.99	2.21	2.21	92.3	92.3	92.3	0.002	0.036
B2	12.6	0.9999	0.001–51.04	0.33	0.31	1.12	1.12	91.0	91	91.0	0.001	0.014
(–)-Epicatechin	13.8	0.9999	0.001–40.87	0.56	0.51	3.45	3.45	94.6	94.6	94.6	0.001	0.025
C1	16.7	0.9999	0.052–48.88	0.49	0.37	4.85	4.85	91.1	91.1	91.1	0.045	0.036
B2-gallate	17.1	0.9993	0.102–45.65	1.23	1.01	5.60	5.60	93.1	93.1	93.1	0.113	0.189
ECG	22.6	0.9991	0.087–57.24	3.39	1.03	7.51	7.51	87.3	87.4	87.3	0.236	0.277

^a SD on measurement is less than to 0.1%

The online version of the original article can be found under doi:[10.1007/s00217-011-1628-0](https://doi.org/10.1007/s00217-011-1628-0).

M. A. Silva · I. Ky · M. Jourdes · P.-L. Teissedre (✉)
Unité de Recherche Oenologie, EA 4577, USC 1219 INRA,
Faculté d'Oenologie, Institut des Sciences de la Vigne et du Vin,
Université Bordeaux Segalen, 210 chemin de Leysotte,
CS 50008, 33882 Villenave d'Ornon Cedex, France
e-mail: p.teissedre@u-bordeaux2.fr