Gated SPET quantification of small hearts: mathematical simulation and clinical application

Kenichi Nakajima¹, Junichi Taki¹, Takahiro Higuchi¹, Masaya Kawano¹, Masashi Taniguchi², Keiko Maruhashi², Shinobu Sakazume², Norihisa Tonami¹

¹ Department of Nuclear Medicine, Kanazawa University Hospital, Kanazawa, Japan

² Department of Pediatrics, Kanazawa University Hospital, Kanazawa, Japan

Published online: 27 October 2000

© Springer-Verlag 2000

Eur J Nucl Med (2000) 27:1869

DOI 10.1007/s002590000431

Eur J Nucl Med (2000) 27:1372–1379

In Table 2, two side headings were inadvertently omitted from the table. The whole table is reproduced here.

Table 2. Results of gated SPET and echocardiography

	Group 1	Group 2	Group 3	Statistics ($P<0.05$)
Echocardiography				
End-diastolic dimension (mm)	30±7	40±3	47±6	1 vs 2, 2 vs 3, 1 vs 3
Fractional shortening (%)	37±6	33±5	32±7	NS
End-diastolic volume (ml)	30±19	72±25	108±43	1 vs 3 2 vs 3
Ejection fraction (%)	75±8	69±7	68±10	NS
Gated SPET				
End-diastolic volume (ml)	23±15	63±25	108±29	1 vs 2, 2 vs 3, 1 vs 3
Ejection fraction (%)	67±11	63±14	60±12	NS
Maximum count in the anterior projection				
Count/pixel	31±12 ^a	41±7	35±7	NS
Count/pixel/min × no. of frames	302±166 ^a	465±218	511±218	NS
Image quality				
Good	2 (28%)	11 (100%)	8 (89%)	$P=0.008$
Fair	4 (57%)	0 (0%)	1 (1%)	
Poor	1 (14%)	0 (0%)	0 (0%)	

The online version of the original article can be found at
<http://doi.org/10.1007/s002590000299>