Erratum to: Pressure cycling technology (PCT) reduces effects of inhibitors of the PCR

Pamela L. Marshall · Jonathan L. King ·
Nathan P. Lawrence · Alexander Lazarev ·
Vera S. Gross · Bruce Budowle

Published online: 1 August 2014
© Springer-Verlag Berlin Heidelberg 2014

Erratum to: International Journal of Legal Medicine
DOI 10.1007/s00414-012-0770-y

Table 1 of the original version of this article, unfortunately, contained errors.

The correct table is presented here.

The online version of the original article can be found at <http://dx.doi.org/10.1007/s00414-012-0770-y>.

P. L. Marshall (✉) · J. L. King · B. Budowle
Institute of Applied Genetics,
Department of Forensic and Investigative Genetics,
University of North Texas Health Science Center,
3500 Camp Bowie Blvd.,
Fort Worth, TX 76107, USA
e-mail: pamela.marshall@unthsc.edu

N. P. Lawrence · A. Lazarev · V. S. Gross
Pressure BioSciences Incorporated,
14 Norfolk Avenue,
South Easton, MA 02375, USA

Table 1 Effect of Pressure Treated Inhibitors on Amplification of DNA

Sample	Average Quant (ng/μl)	StDev Quant	Average IPC	StDev IPC
No Hem - PCT	0.96	0.12	27.50	0.08
No Hem - NPC	0.71	0.11	27.84	0.12
2.5 μM Hem - PCT	0.76	0.18	27.90	0.09
2.5 μM Hem - NPC	0.56	0.03	27.63	0.26
5 μM Hem - PCT	0.23	0.13	34.33	1.42
5 μM Hem - NPC	0.14	0.03	Undetermined	0
7 μM Hem - PCT	Undetermined	0	Undetermined	0
7 μM Hem - NPC	Undetermined	0	Undetermined	0
No HA - PCT	0.99	0.09	27.43	0.13
No HA - NPC	0.90	0.16	27.41	0.05
2.5 ng HA - PCT	0.75	0.03	33.21	0.12
2.5 ng HA - NPC	Undetermined	0	Undetermined	0
5 ng HA - PCT	Undetermined	0	Undetermined	0
5 ng HA - NPC	Undetermined	0	Undetermined	0

Quant Quantifiler Results, *StDev* Standard Deviation, *IPC* Internal PCR Control, *Hem* Hematin, *HA* Humic Acid, *PCT* Pressure Cycling Technology, *NPC* No Pressure Cycling