Erratum to: A disposable biosensor for the determination of alpha-amylase in human saliva

Mika Mahosenaho · Felice Caprio · Laura Micheli ·
Adama M. Sesay · Giuseppe Palleschi · Vesa Virtanen

Published online: 19 February 2013
© Springer-Verlag Wien 2013

Erratum to: Microchim Acta (2010) 170:243–249
DOI 10.1007/s00604-010-0360-y

The original version of this paper, unfortunately, contained an error in Table 1. Most of the values given in the table are incorrect.

The correct Table 1 is presented below.

Table 1 Linear measurement ranges, detection limits and sensitivity for glucose, maltose and maltotriose activity in phosphate buffer using the same electrode ($n=6$)

	Glucose	Maltose	Maltotriose
Range	5–200 μM	1–100 mM	10–100 mM
LOD	5 μM	1 mM	20 mM
Sensitivity	30 A/Mcm ²	240 mA/Mcm ²	4 mA/Mcm ²

The online version of the original article can be found at <http://dx.doi.org/10.1007/s00604-010-0360-y>.

M. Mahosenaho · A. M. Sesay · V. Virtanen
Laboratory of Biotechnology, Kajaani University consortium,
Oulu University, Salmelantie 43,
Sotkamo 88600, Finland

F. Caprio · L. Micheli (✉) · G. Palleschi
Dipartimento di Scienze e Tecnologie Chimiche,
University of Rome “Tor Vergata”, Via della Ricerca Scientifica,
00133 Rome, Italy
e-mail: micheli@uniroma2.it