Håkan Källgren
Margreet Lauwaars
Bertil Magnusson
Leslie Pendrill
Phillip Taylor

Role of measurement uncertainty in conformity assessment in legal metrology and trade

Published online: 28 February 2004
© Springer-Verlag 2004

The online version of the original article can be found at <http://dx.doi.org/10.1007/s00769-003-0707-8>

H. Källgren · B. Magnusson · L. Pendrill (✉)
National Metrology Institute,
SP Swedish National Testing & Research Institute,
Box 857, SE-50115 Borås, Sweden
e-mail: leslie.pendrill@sp.se
Tel.: +46-33-165444
Fax: +46-33-165620

M. Lauwaars · P. Taylor
IRMM European Commission Joint Research Centre,
Institute for Reference Materials and Measurements,
Retieseweg, B-2440 Geel, Belgium

Accred Qual Assur (2003) 8:541–547

Unfortunately the equation on the bottom of the left column on page 545 was incomplete. The correct equation is given below:

$$m_o - (MPE - U) < m_c < m_o + (MPE - U)$$