Erratum to: Decontamination of dental implant surfaces by means of photodynamic therapy

**Juliana Marotti • Pedro Tortamano • Silvana Cai •
Martha Simões Ribeiro •
João Eduardo Miranda Franco •
Tomie Toyota de Campos**

Published online: 24 January 2013
© Springer-Verlag London 2013

**Erratum to: Lasers Med Sci
DOI 10.1007/s10103-012-1148-6**

The original version of this article, unfortunately, contains an error.

The energy was erroneously calculated based on an irradiation of 40mW (input power), not with **30mW** as it was originally presented (output power). For this reason, the energy should be **5.4 J** (for 3 min of irradiation) and **9 J** (for 5 min) and not 7.2 and 12J, respectively.

The online version of the original article can be found at <http://dx.doi.org/10.1007/s10103-012-1148-6>.

J. Marotti
Department of Prosthodontics and Dental Materials, Medical
Faculty, University Hospital RWTH Aachen, Pauwelsstrasse 30,
52074 Aachen, Germany

P. Tortamano • J. E. M. Franco • T. T. de Campos
Department of Prosthodontics, School of Dentistry,
University of São Paulo, Av. Prof. Lineu Prestes, 2227,
05508-000 São Paulo, SP, Brazil

P. Tortamano
e-mail: tortamano@usp.br

J. E. M. Franco
e-mail: joaofranco@usp.br

T. T. de Campos
e-mail: tncampos@usp.br

S. Cai
Department of Microbiology, Biomedical Institute,
University of São Paulo, Av. Prof. Lineu Prestes, 2415,
05508-000 São Paulo, SP, Brazil
e-mail: silvanac@icb.usp.br

M. S. Ribeiro
Center for Lasers and Applications, IPEN-CNEN/SP,
University of São Paulo, Av. Prof. Lineu Prestes, 2242,
05508-000 São Paulo, SP, Brazil
e-mail: marthasr@usp.br

J. Marotti (✉)
Departamento de Prótese, Faculdade de Odontologia,
Universidade de São Paulo, Av. Prof. Lineu Prestes, 2227,
05508-000 São Paulo, SP, Brazil
e-mail: marotti@usp.br