Erratum to: Universal hydrophilic coating of thermoplastic polymers currently used in microfluidics

Caterina Zilio · Laura Sola · Francesco Damin ·
Lucia Faggioni · Marcella Chiari

Published online: 15 February 2014
© Springer Science+Business Media New York 2014

Erratum to: Biomed Microdevices (2014) 16:107–114
DOI 10.1007/s10544-013-9810-8

The original version of this article unfortunately contained a mistake. Tables 1 and 2 are missing in the article and the tables are shown below:

Table 1 Thickness, mass and density of polymer coatings

Surface Coating	Th (nm)	Mass (ng/mm ²)	Density (g/cm ³)
Poly(DMA-GMA-MAPS)	9.2 +/- 0.9	1.6 +/- 0.02	0.179 +/-0.015
Poly(DMA-NAS-MAPS)	13.6+/-3.8	1.4 +/-0.01	0.105 +/-0.03

Table 2 Thickness, mass and density of the protein layer adsorbed to both coated and uncoated surfaces

Surface Coating	Th (nm)	Mass (ng/mm ²)	Density (g/cm ³)
Poly(DMA-NAS-MAPS)	0.4 +/-0.02	0.2+/- 0.002	0.486 +/-0.008
Uncoated	5.7 +/- 0.2	2.2 +/-0.03	0.382 +/-0.007

The online version of the original article can be found at <http://dx.doi.org/10.1007/s10544-013-9810-8>.

C. Zilio · L. Sola · F. Damin · L. Faggioni · M. Chiari (✉)
Istituto di Chimica del Riconoscimento Molecolare, Consiglio
Nazionale delle Ricerche, Via Mario Bianco 9, 20131 Milan, Italy
e-mail: marcella.chiari@icrm.cnr.it