Signal-dependent wavelets for electromyogram classification

A. Maitrot · M.-F. Lucas · C. Doncarli · D. Farina

Published online: 10 July 2007
© International Federation for Medical and Biological Engineering 2007

**Erratum to: Med Bio Eng Comput (2005) 43(4):487–492
DOI 10.1007/BF02344730**

All results reported in the paper are valid since they have been obtained with the correct expression. The mistake in the reported expression was only due to a typing error.

The following formula:

$$\begin{aligned}i = 0, 1 : h[i] &= 4\sqrt{2}[(1 + (-1)^i \cos \alpha + \sin \alpha) \\&\quad \times (1 - (-1)^i \cos \beta - \sin \beta) \\&\quad + (-1)^i 2 \sin \beta \cos \alpha] \\i = 2, 3 : h[i] &= 2\sqrt{2}[1 + \cos(\alpha - \beta) + (-1)^i \sin(\alpha - \beta)] \\i = 4, 5 : h[i] &= 1/\sqrt{2} - h(i - 4) - h(i - 2)\end{aligned}$$

$$\begin{aligned}i = 0, 1 : h[i] &= [(1 + (-1)^i \cos \alpha + \sin \alpha) \\&\quad \times (1 - (-1)^i \cos \beta - \sin \beta) \\&\quad + (-1)^i 2 \sin \beta \cos \alpha] / (4\sqrt{2}) \\i = 2, 3 : h[i] &= [1 + \cos(\alpha - \beta) + (-1)^i \sin(\alpha - \beta)] / (2\sqrt{2}) \\i = 4, 5 : h[i] &= 1/\sqrt{2} - h(i - 4) - h(i - 2)\end{aligned}$$

reported at page 488, should read:

The online version of the original article can be found under
doi:[10.1007/BF02344730](https://doi.org/10.1007/BF02344730).

A. Maitrot · M.-F. Lucas · C. Doncarli
Institut de Recherche en Communication et Cybernétique
de Nantes (IRCCyN), Nantes, France

D. Farina (✉)
Center for Sensory-Motor Interaction (SMI),
Department of Health Science & Technology,
Aalborg University, Aalborg, Denmark
e-mail: df@hst.aau.dk