Pressure Pulse Contour-derived Stroke Volume and Cardiac Output in the Morbidly Obese Patient

Donald P. Bernstein

Published online: 4 June 2008
© Springer Science + Business Media, LLC 2008

Erratum to: OBES SURG
DOI 10.1007/s11695-007-9378-6

Subsequent to e-publication of the above Review Article (which is the preceding article), the author noted that he had made an error in the description of Eq. 8. The sentence above Eq. 8 currently reads, K is the ratio of uncalibrated pressure pulse contour-derived SV (SV_{UNCAL}) and transpulmonary TDCO-derived SV (i.e. $K=SV_{\text{UNCAL}}/SV_{\text{TD}}$). Thus,

$$SV_{\text{CAL}} = SV_{\text{UNCAL}} \times \left(\frac{SV_{\text{UNCAL}}}{SV_{\text{TD}}} \right) \quad (8)$$

The sentence and corresponding equation should read, K is the ratio of transpulmonary TDCO-derived SV and uncalibrated pressure pulse-derived SV (i.e. $K=SV_{\text{TD}}/SV_{\text{UNCAL}}$). Thus,

$$SV_{\text{CAL}} = SV_{\text{UNCAL}} \times \left(\frac{SV_{\text{TD}}}{SV_{\text{UNCAL}}} \right). \quad (8)$$

The online version of the original article can be found at
<http://dx.doi.org/10.1007/s11695-007-9378-6>.

D. P. Bernstein (✉)
Department of Anesthesia, Palomar Medical Center
and Palomar-Pomerado Research Institute,
Escondido, CA 92025, USA
e-mail: strokevolumedon@aol.com