Correction to: Metrology for stable isotope reference materials: $^{13}\text{C}/^{12}\text{C}$ and $^{18}\text{O}/^{16}\text{O}$ isotope ratio value assignment of pure carbon dioxide gas samples on the Vienna PeeDee Belmnite- CO_2 scale using dual-inlet mass spectrometry

Abneesh Srivastava¹ · R. Michael Verkouteren²

Published online: 14 March 2019

© Springer-Verlag GmbH Germany, part of Springer Nature 2019

Correction to: Anal Bioanal Chem

<https://doi.org/10.1007/s00216-018-1064-0>

The authors would like to call the reader's attention to the fact that unfortunately the formula for the O17 correction parameter "a" is reported incorrectly as

$$\frac{R17(sam)}{R18(sam)} = \left(\frac{R17(RM-O)}{R18(RM-O)} \right)^a$$

The formula can be found in the section titled "Data normalization: Realization of VPDB scale for traceability" and should read:

$$\frac{R17(sam)}{R17(RM-O)} = \left(\frac{R18(sam)}{R18(RM-O)} \right)^a$$

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

The online version of the original article can be found at <https://doi.org/10.1007/s00216-018-1064-0>

✉ Abneesh Srivastava
abneesh.srivastava@nist.gov

¹ Gas Sensing Metrology Group, Chemical Sciences Division, National Institute of Standards and Technology, 100 Bureau Drive, Gaithersburg, MD 20899, USA

² Surface and Trace Chemical Analysis Group, Materials Measurement Science Division, Materials Measurement Laboratory, National Institute of Standards and Technology, 100 Bureau Drive, Gaithersburg, MD 20899, USA