

Erratum to: Validation of the *custo screen pediatric* blood pressure monitor according to the European Society of Hypertension International Protocol revision 2010

Beate Beime¹ · Cornelia Deutsch¹ · Ralf Krüger¹ · Andreas Wolf² · Peter Müller² · Gertrud Hammel³ · Peter Bramlage¹

Published online: 4 May 2017
© Springer-Verlag Berlin Heidelberg 2017

Erratum to: Eur J Pediatr
DOI 10.1007/s00431-017-2874-3

The original version of this article, unfortunately, contained errors.

In Table 5, entries “Passed” and “Passed and Result” were not properly aligned under Grade 2 and Grade 3. The “Passed” entries in column 4 should be moved under Grade 2 and “Passed and Result” entries under Grade 3.
The correct Table is as follows:

The online version of the original article can be found at <http://dx.doi.org/10.1007/s00431-017-2874-3>

✉ Beate Beime
beate.beime@ippmed.de

Cornelia Deutsch
cornelia.deutsch@ippmed.de

Ralf Krüger
ralf.krueger@ippmed.de

Andreas Wolf
wolf@ms-gmbh.de

Peter Müller
mueller@ms-gmbh.de

Gertrud Hammel
gertrud.hammel@estimate.de

Peter Bramlage
peter.bramlage@ippmed.de

¹ Institute for Pharmacology and Preventive Medicine, Bahnhofstrasse 20, 49661 Cloppenburg, Germany

² Müller & Sebastiani GmbH, Ottobrunn, Germany

³ Estimate GmbH, Augsburg, Germany

Table 5 Validation Results

Part 1	≤ 5 mmHg	≤ 10 mmHg	≤ 15 mmHg	Grade 1	Mean \pm SD (mmHg)
Pass requirements					
Two of	73	87	96		
All of	65	81	93		
Achieved					
SBF	95	99	99	Passed	-1.4 ± 3.0
DBF	91	99	99	Passed	-0.7 ± 3.2
Part 2	$2/3 \leq 5$ mmHg	$0/3 \leq 5$ mmHg		Grade 2	Grade 3
Pass requirements	≥ 24	≤ 3			
Achieved					
SBF	32	0		Passed	
DBF	33	0		Passed	
Part 3					Result Passed

SBP systolic blood pressure, *DBP* diastolic blood pressure, *SD* standard deviation