

## Correction to: Uncertainty estimation of the determination of chemical sum parameters in water

Tony Venelinov<sup>1</sup>

Published online: 10 October 2017  
© Springer-Verlag GmbH Germany 2017

### Correction to: Accred Qual Assur DOI 10.1007/s00769-017-1276-6

Unfortunately, Eq. 3 was incorrectly presented in the original publication. The correct version of Eq. 3 is

$$u_{ip} = \frac{s_d}{\sqrt{d}} = \sqrt{\frac{MS_{\text{betweengroup}} - MS_{\text{withingroup}}}{n_{\text{pergroup}}}},$$

where  $u_{ip}$  is the uncertainty due to intermediate precision,  $s_d$  is the day-to-day variation,  $d$  is the number of measurement days,  $MS_{\text{betweengroup}}$  is mean of squares between the groups,  $MS_{\text{withingroup}}$  is mean of squares within the groups, and  $n_{\text{pergroup}}$  is the number of replicates.

---

The online version of the original article can be found under doi:[10.1007/s00769-017-1276-6](https://doi.org/10.1007/s00769-017-1276-6).

---

✉ Tony Venelinov  
tvenelinov\_fhe@uacg.bg

<sup>1</sup> Faculty of Hydraulic Engineering, Chair of Water, Sewerage, Water and Wastewater Treatment, University of Architecture, Civil Engineering and Geodesy, 1 Hr. Smirnenski blvd., 1164 Sofia, Bulgaria