



Correction to: Intervendor variability of carotid intima–media thickness measurement: validation study using newly developed ultrasound phantom

Tomoko Ishizu¹ · Hirotochi Hamaguchi² · Naotaka Nitta³ · Yoshihiro Seo⁴ · Hiroshi Matsuo⁵ · Tsuyoshi Shiina⁶

Published online: 1 July 2020

© The Japan Society of Ultrasonics in Medicine 2020

Correction to:

Journal of Medical Ultrasonics (2020) 47:155–165

<https://doi.org/10.1007/s10396-019-00995-7>

In the Original publication of the article, Table 3 was published incorrectly. The correct Table 3 is given in this Correction.

Publisher's Note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

The original article can be found online at <https://doi.org/10.1007/s10396-019-00995-7>.

✉ Tsuyoshi Shiina
shiina.tsuyoshi.6w@kyoto-u.ac.jp

¹ Department of Clinical Laboratory Medicine, Faculty of Medicine, University of Tsukuba, 1-1-1 Tennodai, Tsukuba, Ibaraki 305-0817, Japan

² Department of Neurology, Kita-Harima Medical Center, 926-253 Ichiba-cho, Ono, Hyogo 675-1392, Japan

³ Health Research Institute, National Institute of Advanced Industrial Science and Technology (AIST), 1-2-1 Namiki, Tsukuba, Ibaraki 305-8564, Japan

⁴ Department of Cardiology, Nagoya City University, 1 Kawasumi, Mizuho-cho, Mizuho-ku, Nagoya 467-8601, Japan

⁵ Matsuo Medical Clinic, 2-15-25 Kita-Honmachi, Yao City, Osaka 581-0802, Japan

⁶ Department of Human Health Sciences, Graduate School of Medicine, Kyoto University, 53 Kawahara-cho, Shogoin, Sakyo-ku, Kyoto 606-8507, Japan

Table 3 Vendor participating in the study with the version of equipment provided and exported 2D image information

Vendor	Ultrasound machine	Probe	Frequency bands	Exported 2D image resolution ^a , mm/pixel	Number of the pixel in 10-mm length of IMT
Fuji-Film	FC1	HFL38	13–6 MHz	0.063	159
Fukuda Denshi	UF-760AG+	FUT-LA385-12P	12–5 MHz	0.056	179
GE	LOGIQ S8	ML-15-D	12 MHz	0.038	263
Hitachi	ARIETTA 70	L441	11–2 MHz	0.052	192
Konica Minolta	SONIMAGE HS1	L18-4	18–4 MHz	0.057	175
Philips	EPIC7	L12-3	12–3 MHz	0.046	217
Siemens	ACUSON S2000	14L5	11 MHz	0.044	227
Canon	Aplio i900	11L4 PLU-704BT	11–4.8 MHz	0.060	167

^aExported 2D image resolution was presented as the tissue length/pixels of the picture from the depth set to 3.0 cm in each ultrasound machine preset for the carotid ultrasonography