



# Correction to: Radiation exposure for intraoperative 3D scans in a hybrid operating room: how to reduce radiation exposure for the surgical team

K. Schuetze<sup>1</sup> · M. Kraus<sup>2</sup> · A. Eickhoff<sup>1</sup> · F. Gebhard<sup>1</sup> · P. H. Richter<sup>1</sup>

Published online: 15 June 2018  
© CARS 2018

## Correction to:

**International Journal of Computer Assisted Radiology  
and Surgery**

<https://doi.org/10.1007/s11548-018-1747-1>

The original version of this article unfortunately contained a mistake. In abstract, results section should read:

**Results** The highest scatter radiation was measured during 3D scans of the pelvis on the side of the surgical assistant (39.5  $\mu$ Sv in height of 1 m) compared to the side of the main surgeon (23  $\mu$ Sv in height of 1 m). Scatter radiation was less on the position of the scrub nurse (6.8  $\mu$ Sv in height of 1 m) and during 3D scans of the other anatomical regions. The radiation dosage was about 66% less with the DRP. Low values of scatter radiation were measured behind a radiation protection wall and with open doors in the adjacent rooms.

---

The original article can be found online at <https://doi.org/10.1007/s11548-018-1747-1>.

---

✉ K. Schuetze  
konrad.schuetze@uniklinik-ulm.de

<sup>1</sup> Department of Trauma-, Hand- and Reconstructive Surgery, Ulm University, Albert-Einstein-Allee 23, 89081 Ulm, Germany

<sup>2</sup> Department of Orthopaedic Surgery, Ulm University, Oberer Eselsberg 45, 89081 Ulm, Germany