The fight against COVID-19: disinfection protocol and turning over of CleanSpace® HALO™ in a Singapore Hospital

Frederick H. Koh¹ · Moarie G. Tan² · Min-Hoe Chew¹

Received: 8 April 2020 / Accepted: 18 May 2020 / Published online: 27 May 2020
© Italian Society of Surgery (SIC) 2020

To the Editor-in-Chief,

The COVID-19 pandemic has infected more than 3 million patients, caused > 230,000 deaths, and have greatly impacted the lives of millions [1]. Singapore was one of the first countries outside China to be affected by the virus. Aggressive contact tracing and isolation with national legislations have helped limit the spread of the virus to just over 1400 people over the last 10 weeks of the pandemic. Nonetheless, there has been increasing community spread with a high number of unlinked cases detected. This has necessitated a partial lockdown to enforce distancing measures. Healthcare workers (HCW) meanwhile are rapidly ramping up capabilities and preparing for a worsening of the crisis.

COVID-19 has been well established to spread through droplets and potentially, aerosol-generating procedures (AGP) [2]. The protection of the health of HCW is of paramount importance. One of the few protective measures is the strict compliance to personal protective equipment (PPE). On top of the utility of surgical face masks and N95 respirators, powered air-purifying respirators (PAPR) is another tool in our armamentarium. Of significant importance is the worldwide shortage of disposable N95 respirators due to production and supply chain cuts. Protectionist policies and stockpiling have also greatly reduced availability [3].

Sengkang General Hospital was a newly opened 1400 bed hospital in August 2018. To date, we have managed 748 COVID-19 positive cases in a “mixed COVID” hospital. Currently, 0% of the hospital’s HCW have been infected by

COVID-19. As routine stocks for disaster and emergency preparedness, we had purchased 686 CleanSpace® HALO™ (CleanSpace Technology Pte Ltd, Artarmon, NSW, Australia) PAPR. In the pandemic outbreak, these PAPRs were provisioned to clinical areas with COVID patient contact or had a risk of exposure to AGP—Operating Theatre (OT), Endoscopy Centre, Emergency Department (ED), Intensive Care Unit (ICU) and isolation wards.

Each CleanSpace® HALO™ consists of a detachable reusable mask (half/full), HEPA filter, the power system, reusable harness and neck support. When compared to the HAZMAT suits and 3 M™ Jupiter™ PAPR Helmet, which were used commonly during the SARS-CoV-2003 outbreak, the CleanSpace® HALO™ is marketed as the lightest PAPR, potentially offering doctors better manoeuvrability and comfort. This PAPR with face-pieces also has a higher Assigned Protection Factor (workplace level respiratory protection) of 50 compared to hooded PAPR with loose-fitted PAPR which has an APR of 25 [4].

The utility, disinfection and turnover of these PAPR are critical in ensuring the availability of the device during high risks procedures. To date, there has been limited knowledge of the processes and recommendations for such PAPR. This is especially pertinent as COVID-19 pandemic is likely to be protracted and traditional N95 respirators will continue to be restricted supply.

Our hospital’s Central Sterile Supply Unit (CSSU) has advised for the surface disinfection of individual parts of the PAPR upon its removal, using Isopropyl Alcohol 70% wipes in keeping with UK guidelines.[5] While bleach may reduce contaminants more, handling of bleach in the clinical area is not as common as compared to the alcohol wipes [6]. The masks and harnesses of the CleanSpace® HALO™ are then collected in biohazard-waste-bags-lined containers before being dispatched to the CSSU 3 times a day (8 a.m., 2 p.m., 8 p.m.), to ensure a full-load disinfection is performed each time. When masks and harnesses are dropped off with a dispatch form (Fig. 1), they would be exchanged with the same

Electronic supplementary material The online version of this article (<https://doi.org/10.1007/s13304-020-00809-3>) contains supplementary material, which is available to authorized users.

✉ Min-Hoe Chew
chew.min.hoe@singhealth.com.sg

¹ Department of General Surgery, Sengkang General Hospital, 110 Sengkang East Way, Singapore 544886, Singapore

² Operating Theatre Services, Division of Nursing, Sengkang General Hospital, Singapore, Singapore

Fig. 1 PAPR dispatch form with pictorial instructions on surface disinfection and packaging before thermal disinfection by CSSU

CleanSpace HALO Mask & Harness Dispatch Form

Unit : _____
 Contact Number : _____
 Contact Person : _____

Dispatch Timings:
 8am, 2pm, 8pm

Put this form into a clean plastic bag, and staple above the cable tie on the biohazard bag.

1 Wipe PAPR (Mask/Harness/Power Blower) after use.

Mask & Harness

Power Blower

Use the above disinfectant wipes available in unit.

2 Bag the Mask & Harness

Bag the Mask as shown using the following items:
 x2 Biohazard Bag
 x1 Cable Tie
 Place x1 Dispatch Form in a clean plastic bag, and staple above the cable tie on the biohazard bag.

3 Store at individual unit

DO NOT SEND TO CSSU

4 Indicate Quantity DISPATCHED to CSSU

	Size S	Size M	Size L
Half Mask, CleanSpace®			
Harness			
CSSU Staff Received : _____			
Date & Time : _____			

5 Indicate Quantity RETURNED to Unit

	Size S	Size M	Size L
Half Mask, CleanSpace®			
Harness			
CSSU Staff Returned : _____			
Date & Time : _____			
Porter Name : _____			

**Dispatch form is kept in respective ward*

number of cleaned masks and harnesses and sent back to facilitate quick turnover and sustained availability of PAPR in clinical areas.

In the CSSU, further thermal disinfection, which was found to be superior to surface alcohol disinfection, would be performed for the masks and harnesses. This includes a 25 min wash cycle, followed by a 1-min thermal disinfection at 90 °C. The masks have been validated to tolerate up to 30 such cycles up to a maximum of 90 °C. To safeguard against silicone degradation, the drying cycle of the disinfection, which is at a temperature > 90 °C, has been replaced by blow drying by pressurized air-guns. A total of 45 min is required for this entire process.

Random adenosine triphosphate (ATP) test swabs, especially around the exhalation valves where most contamination are expected, are conducted to ensure quality control of the cleaning efficacy before the parts are repackaged. The mean Relative Light Unit (RLU) before and after the disinfection process is 430.9 and 23.6. (Figure S1) Masks and harnesses are tagged digitally, and upon completion of 30 cleaning cycles, would be identified and discarded.

The fight against COVID-19 will be protracted [7]. The protection of HCW and thereby the preservation of the workforce cannot be further emphasized. The proper utility and cleansing of PAPR are but just 1 crucial element in the cog. We hope that by sharing the rationale and method of reprocessing of CleanSpace® HALO™, hospitals who are making use of this PAPR would be able to use this as a springboard to facilitate readily available PAPR to the HCW participating in high-risk procedure during the pandemic.

Funding Nil.

Compliance with ethical standards

Conflict of interest The author declares that they have no conflict of interest.

Research involving human participants and/or animals The study does not involve any human/animal subjects.

Informed consent Informed consent is not necessary for this type of study.

Code availability NA.

Reference

1. Channelnewsasia (2020) Novel coronavirus cases—live updates. <https://infographics.channelnewsasia.com/covid-19/map.html>. Accessed 8 Apr 2020
2. Lai CC, Shih TP, Ko WC et al (2020) Severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) and coronavirus disease-2019 (COVID-19): the epidemic and the challenges. *Int J Antimicrob Agents* 55(3):105924
3. World Health Organization (2020) Media Statement: the role and need of masks during COVID-19 outbreak. <https://www.who.int/indonesia/news/detail/06-03-2020-media-statement-the-role-and-need-of-masks-during-covid-19-outbreak>. Accessed 8 Apr 2020
4. Occupational Safety and Health Administration; US Department of Labor (2009) Assigned protection factors for the revised respiratory protection standard—2009. <https://www.osha.gov/Publications/3352-APF-respirators.pdf>. Accessed 8 Apr 2020
5. Department of Health (2016) Health Technical Memorandum 01–01: Management and Decontamination of Surgical Instruments (Medical Devices) Used in Acute Care. https://assets.publishing.service.gov.uk/government/uploads/system/uploads/attachment_data/file/545864/HTM0101PartD.pdf. Accessed 2 May 2020
6. Verbeek JH, Rajamaki B, Ijaz S et al (2020) Personal protective equipment for preventing highly infectious diseases due to exposure to contaminated body fluids in healthcare staff. *Cochrane Database Syst Rev* 4:CD011621
7. Tan A (2020) Covid-19 likely to last till end-2020 at least: experts. <https://www.straitstimes.com/singapore/health/coronavirus-covid-19-likely-to-last-till-end-2020-at-least-experts>. Accessed 22 Mar 2020

Publisher's Note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.