

Erratum to: Evaluation of a new device for measurement of hemoglobin A1c for Japanese subjects

Seiya Shimoda · Takako Maeda · Noboru Furukawa ·
Kenshi Ichinose · Kayo Taketa · Motoyuki Igata ·
Takafumi Senokuchi · Takeshi Matsumura · Eiichi Araki

Published online: 5 November 2013
© The Japan Diabetes Society 2013

Erratum to: Diabetol Int (2013) 4:112–116
DOI 10.1007/s13340-012-0102-y

In the original publication, there are errors in the ‘Results’ section of the ‘Abstract’. The correct sentence should appear as given below:

There was a strong positive correlation between HbA1c-HPLC (Y) and HbA1c-new device (X) ($Y = 0.952X + 0.147$, $r = 0.989$; $p < 0.001$; 95 % confidence interval of the slope 0.929–0.976; 95 % confidence interval of the intercept -0.036 to 0.329).

The online version of the original article can be found under doi:[10.1007/s13340-012-0102-y](https://doi.org/10.1007/s13340-012-0102-y).

S. Shimoda · T. Maeda · N. Furukawa · K. Ichinose · K. Taketa ·
M. Igata · T. Senokuchi · T. Matsumura ·
E. Araki (✉)
Department of Metabolic Medicine, Faculty of Life Sciences,
Kumamoto University, 1-1-1 Honjo, Chuo-ku,
Kumamoto 860-8556, Japan
e-mail: earaki@gpo.kumamoto-u.ac.jp