

## Erratum to: Contribution of two molecular assays as compared to selective culture for MRSA screening in a low MRSA prevalence population

E. Nulens · P. Descheemaeker · R. H. Deurenberg ·  
E. E. Stobberingh · B. Gordts

Published online: 3 July 2010  
© Urban & Vogel 2010

**Erratum to: Infection (2010) 38:98–101**  
**DOI 10.1007/s15010-009-9117-0**

There was an error in the “Results” section of the abstract.

The corrected paragraph reads as follows:

**Results:** Eight MRSA carriers were detected by selective culture (1.6% prevalence). The sensitivity, specificity, **positive predictive value**, and **negative predictive value** were 62.5, 99.0, 50.0, and 99.4% for BD GeneOhm™ StaphSR and 62.5, 97.7, 31.3, and 99.4% for Cepheid Xpert™ MRSA, respectively.

---

The online version of the original article can be found under  
doi:[10.1007/s15010-009-9117-0](https://doi.org/10.1007/s15010-009-9117-0).

---

E. Nulens (✉) · P. Descheemaeker · B. Gordts  
Medical Microbiology, General Hospital Sint-Jan AV  
Brugge-Oostende, Ruddershove 10, 8000 Brugge, Belgium  
e-mail: [eric.nulens@azbrugge.be](mailto:eric.nulens@azbrugge.be)

E. Nulens · R. H. Deurenberg · E. E. Stobberingh  
Medical Microbiology, Maastricht University Medical Center,  
P.O. Box 5800, 6202 AZ Maastricht, The Netherlands