To the Editor

We have noticed a common shortcoming in computer programs used to calculate weights in Daltons of ions having elemental compositions entered as input. A similar problem arises when one tries to determine the composition of an ion given the accurate mass measured by the mass spectrometer. Both programs typically neglect the mass of the electron lost in the formation of a radical cation, the presence of a positive charge in an even electron ion, the electron gained in a radical anion formed by electron capture, and the negative charge present in ion adducts and in $[M - H^+]$ ions. In the case of large molecules such as peptides and proteins, this is of little consequence, but in the case of small molecules this can appear as an error of a few parts-per-million (ppm), even when the measurement of the ion mass is exactly correct. Some examples shown in Table 1 from the simple to the more complex will illustrate the nature of the problem.

Similar errors are also noted in many cases when one wishes to accurately calibrate the mass scale with a compound that produces a series of cluster ions such as glycerol. The accurate mass calculated for the $C_3H_8O_3 \cdot H^+$ ion which is one of the calibrating ions commonly used in FAB (as a simple example) is not the value in many lists of masses of calibrants, but rather that of the neutral $C_3H_8O_3$ which is too large by 0.00055 Da, the electronic mass (Table 1 above).

Related programs that determine elemental formulae from measured masses report the masses of the neutral for each composition. It is clear then that even if the mass measurement in the mass spectrometer was perfect (in the case of glycerol, suppose 93.05462 Da was measured for $C_3H_8O_3 \cdot H^+$), the closest match would be for neutral $C_3H_9O_3$ having a calculated mass of 93.05517, and an error in measurement of 5.9 ppm would be reported.

In manual peak matching, the ratios of known and unknown ion masses are measured in the MS. The software provides the masses of the corresponding neutrals. In the case of positive ions, one typically measures the ratio of the masses of ions X^+ and Y^+ , i.e., the ratio M_X^+ / M_Y^+ where M_X is greater than M_Y . Software calculated masses of X and Y yield the ratio $[M_X^+ + e] / [M_Y^+ + e]$ where e is the mass of the electron. Clearly the ratios are not equal. For example, if one were peak matching the molecular cations of benzene (the reference) against perdeuterobenzene (the unknown) in electron impact, we would measure the ion ratios as:

Table 1. Ion Mass Calculation (Th)

Compound	Composition	Calculated ^a	Correct	Error (ppm)
Methane	$CH_4^{+ \cdot}$	16.03130	16.03075	34.3
Methyl ⁺	CH_3^+	15.02348	15.02293	36.3
Methyl ⁻	CH_3^-	15.02348	15.02403	36.6
Glycerol	$C_3H_8O_3 \cdot H^+$	93.05517	93.05462	5.9
	$C_3H_7O_3^-$	91.03952	91.04007	6.0
PEO (10-mer)	$C_{20}H_{42}O_{11} \cdot NA^+$	481.26248	481.26193	1.1
GluFib	$C_{66}H_{95}N_{19}O_{26} \cdot H^+$	1570.67739	1570.67684	0.3
	$C_{66}H_{95}N_{19}O_{26} \cdot H_3^{+3}$	524.23101	524.23047	1.0

PEO: Polyethylene oxide. GluFib: [Glu¹]-Fibrinopeptide B (CAS 103213-49-6)

^aCalculated by an assortment of software programs supplied with commercial MS instruments using the ¹²C scale.

$$C_6D_6^{+ \cdot} / C_6H_6^{+ \cdot} = 84.08406 / 78.04640 = 1.077360$$

and the mass of the perdeuterobenzene positive ion would be obtained as

$$1.077360 \times 78.04695 = 84.08466$$

$$\text{Where neutral } C_6D_6 = 78.04695$$

$$\text{and } C_6D_6 = 84.08461$$

for an apparent error of 0.5 ppm, even though the matching was perfect.

In instances involving electrosprayed relatively small multiply protonated ions, the errors may be misleading. Triply-protonated GluFib for example (Table 1) presents a one ppm difference between the correct mass-to-charge ratio and the value calculated by ignoring the mass of three electrons. Authors reporting accurate ion mass data should ensure that the reported and calculated masses are correctly presented.

Some commercial software packages allow the calculation of the masses of positive and negative ions as well as neutrals (example: ChemDraw Pro, version 8, CambridgeSoft Corp, Cambridge MA, USA). Software purchased as a package from mass spectrometer manufacturers should be consistent and have the same capability, along with the ability to easily select for cations formed by adduction to neutrals by ammonium, sodium and potassium, as well as anion adducts with halides and pseudohalides (e.g., CN^- , OH^-).

O.A. Mamer

A. Lesimple

The Mass Spectrometry Unit

McGill University

Montréal, Quebec City, Canada H3A 1A3

Published online February 6, 2004