Peptide Fragmentation by Corona Discharge Induced Electrochemical Ionization

John R. Lloyd^a and Sonja Hess^b

^a Proteomics and Mass Spectrometry Facility, National Institute of Diabetes and Digestive and Kidney Diseases, National Institutes of Health, Bethesda, Maryland, USA

^b Proteome Exploration Laboratory, California Institute of Technology, Pasadena, California, USA

Fundamental studies have greatly improved our understanding of electrospray, including the underlying electrochemical reactions. Generally regarded as disadvantageous, we have recently shown that corona discharge (CD) can be used as an effective method to create a radical cation species $[M]^+$, thus optimizing the electrochemical reactions that occur on the surface of the stainless steel (SS) electrospray capillary tip. This technique is known as CD initiated electrochemical ionization (CD-ECI). Here, we report on the fundamental studies using CD-ECI to induce analytically useful in-source fragmentation of a range of molecules that complex transition metals. Compounds that have been selectively fragmented using CD-ECI include enolate forming phenylglycine containing peptides, glycopeptides, nucleosides, and phosphopeptides. Collision induced dissociation (CID) or other activation techniques were not necessary for CD-ECI fragmentation. A four step mechanism was proposed: (1) complexation using either Fe in the SS capillary tip material or Cu(II) as an offline complexation reagent; (2) electrochemical oxidation of the complexed metal and thus formation of a radical cation (e.g.; $Fe - e^- \rightarrow Fe^+$); (3) radical fragmentation of the complexed compound; (4) electrospray ionization of the fragmented neutrals. Fragmentation patterns resembling b- and y-type ions were observed and allowed the localization of the phosphorylation sites. (J Am Soc Mass Spectrom 2010, 21, 2051–2061) © 2010 American Society for Mass Spectrometry

Since its inception, fundamental studies have greatly improved our understanding of electrospray [1]. It is, for instance, now no longer questioned that the electrospray source also functions as an electrolytic cell, where electrochemical oxidations can happen as originally proven by Blades et al. [2] and extensively studied by van Berkel and Kertesz [3]. The focus of these studies was the reactions in the gas phase, and not those at the surface of the electrodes. It is thus not surprising that until recently, the connection between the electrolytic nature of electrospray and the onset of commonly observed discharges (i.e., the release of electrons) was not well understood. For instance, in a recent review, Kobarle et al. have associated the onset of these discharges with cosmic rays or background radiation. What we have recently shown is that the onset of the discharges is associated with the electrochemistry that Kobarle initially described [1, 2]. CD is initiated at the sharp edge of the capillary due to its high potential energy and follows the normal voltage-current characteristics as established by Penning, where dependent on the current, discharges are classified as dark discharges (10^{-10} A– 10^{-5} A), glow discharges (10^{-5} A–ca 1 A) and arc discharges (ca 1 A– 10^4 A) [4–6]. At dark discharges, very few electrons (10^{-10} A– 10^{-5} A) are released, light

emission is thus negligible. In contrast, corona discharge (CD) is characterized by a stable emission of blue or red light, depending on the electron densities of the plasma. CD is a special case in that it oscillates at steady-state between the Townsend region of the dark discharge and the glow discharge region, roughly between 10^{-7} A– 10^{-4} A. These oscillations are very well studied, experimentally by Trichel, hence the name Trichel pulses, and theoretically by Morrow [7, 8]. In addition, the flow of electrons out of the electrospray tip measured in ESI and CD-ESI techniques has been extensively studied and documented [6, 9]. We have recently shown that corona discharge (CD) can be used as an effective method to create a radical cation species $[M]^+$ using a regular ESI source and appropriate MS settings [6]. To achieve a stable CD, the exit tip of a stainless steel electrospray capillary was extended 3 mm beyond the desolvation gas tube, the nitrogen gas temperature was raised to at least 500 °C, and the ES high voltage to at least 5 KV. As shown in Figure 1, the ion plasma created was diverted away from the ESI spray path by a ground electrode placed behind the ESI capillary exit. CD at the capillary tip oxidizes the Fe in the stainless steel, which subsequently oxidizes redox active analytes when the ionization potential of the analyte is equal to or lower than that of Fe. Detection of N-ferrocenyl iodoacetamide (FcIAA) at the low zeptomolar level was demonstrated. The excellent sensitivity

Address reprint requests to Dr. S. Hess, California Institute of Technology, BI 211, MC139-74, Pasadena, CA 91125, USA. E-mail: shess@caltech.edu

Figure 1. Experimental setup shown in (a) side view: in CD-ECI, the plasma ion path was directed away from the analyte ion path to the ground plate behind the orifice. CD removed an electron from Fe at the sharp edge of the SS capillary resulting in radical Fe^{•+}. Analytes (M) with ionization potential lower than Fe/Fe^{•+} were ionized by electrochemical oxidation and fragmented. During this process, Fe⁰ was regenerated. The surface interaction is critical for the formation of the radical ions and for their subsequent fragmentation; (b) top view: right after the generation of the analyte and plasma, both paths are separated.

could be attributed to effective electrochemical ionization and the ability of the enolate form of FcIAA to bind and accumulate to the Fe in SS capillary [6]. We termed this process CD initiated electrochemical ionization (CD-ECI) [6]. The insight that electrochemical reactions at both electrodes are essential for the formation of CD was already recognized by Perrot in 1861 [10] and was summarized in the benchmarking book of Thomson, who also pointed out the similarities between current conducting gases and liquid electrolytes [11]. Surprisingly, the underlying surface chemistry was not intensively studied in either the engineering field or mass spectrometry until recently [6, 12].

Inspired by recent studies that used electrochemistry to generate radical ions to fragment peptides, we tested whether CD-ECI could be used in this regard [13–15].

Permentier and Bruins used an electrochemical flow cell coupled to a mass spectrometer to generate radicals that resulted in the specific cleavage of proteins at tyrosine and tryptophans [13]. Prudent and Girault used Cu(II) either in solution or a sacrificial Cu (0) anode to generate Cu-peptide complexes that would then undergo an electrochemical redox reactions. Depending on the formed complex, some fragmented without collision induced dissociation (CID) while others required CID [14]. CID was also required for generating and fragmenting the peptide radical cations from Cu(auxiliary ligand)-peptide complexes investigated by other groups [16–20]. Using a different concept, namely a CD needle placed in a flow reactor between the capillary and the MS ion entrance, Vilkov et al. recently introduced a novel dissociation technique. Electrochemically generated OH radicals from the CD needle fragmented peptides through ion/molecule reactions. Here, the peptides were sprayed directly at the CD needle (conceptually similar to an APCI setup) resulting in a mixture of diagnostically useful c-, b-, and y-ions [15]. Other groups relied on generation of radical ions through homolytic cleavage to effectively fragment peptides by either free radical initiated peptides sequencing (FRIPS) using an N-terminal Vazo 68 conjugate as pioneered by Beauchamp et al. [21], or side-chain lysine peroxycarbamates as developed by Porter et al. [22]. In both cases, CID was necessary to generate and fragment the peptide radical ion species.

Here, we report on the fundamental studies using CD-ECI to induce analytically useful in-source fragmentation of a range of molecules including glycopeptides, enolate forming phenylglycine containing peptides, nucleosides, and phosphopeptides, and study the involved mechanism. Unlike other in-source fragmentation techniques that are generally described to be non-selective [15, 23], CD-ECI showed selectivity for enolate-forming, metal complexing compounds. Experiments were performed using CD-ECI techniques on an LCT time-of-flight (TOF) instrument. Comparisons were done with ESI techniques on an LCT or, in select cases, on a hybrid QTOF II instrument.

Experimental

Materials

All reagents were of the highest purity available. BOC-L- α -phenyl glycine 99%, DMSO $\geq$ 99.9%, PPPPPPPPP-PPPPPP (P14R), calcineurin substrate phosphopeptide DLDVPIPGRFDRRVpSVAAE 97%, pyridine $\geq$ 99%, ristomycin A, HPLC grade acetonitrile, and formic acid were all obtained from Sigma-Aldrich, St. Louis, MO, USA. ⁴FUP14R ($\geq$ 95%), where ⁴FU was a 4-fluorophenyl glycine, was from Anaspec, Freemont, CA, USA. The HPLC water was obtained from Riedel-deHaen (Sigma-Aldrich). The MassPREP Enolase Digest with Phosphopeptide Mix was purchased from Waters, Milford, MA.

LC Instrumentation

A 1525 μ binary HPLC (Waters) was used. Samples (2 μ L) were injected using a 2777 Sample Manager (Waters). A Zorbax-SB C3 2.1 $\times$ 150 mm, 5 μ m particle size was used at a flow rate of 200 μ L/min. Solvent A was 0.2% formic acid, 2% acetonitrile in water, and Solvent B was 0.2% formic acid, 2% water in acetonitrile.

MS Instrumentation and ESI and CD-ECI Techniques

Experiments were done with a Waters LCT Premier High Resolution TOF mass spectrometer equipped with the API Z-spray ESI ionization source or where specifically noted with a Waters QTOF II mass spectrometer (Waters). The LCT was operated in W-mode with a nominal resolution of 10 K. The ionization mode was positive.

For ESI experiments the capillary voltage was 2 kV, the cone voltage was 30 V, and the desolvation gas was nitrogen at a flow rate of 300 L/h at 300 °C. The angle of the stainless steel capillary to the ion entrance of the MS was 80°. The tip of the capillary was extended 0.5 mm from the desolvation tube. The source temperature was 125 °C.

Instrument parameters for CD-ECI mass spectrometry and liquid chromatography were used as previously described [6]: briefly, the capillary voltage was 5 kV, the cone voltage was 30 V, and the desolvation gas was nitrogen at a flow rate of 300 L/h at 500 °C. The angle of the stainless steel capillary to the ion entrance of the MS was 90°. The tip of the capillary was extended 3 mm from the desolvation tube. The source temperature was 125 °C.

For the QTOF experiment using UP14R, a Waters QTOF II was operated in positive ion mode with a nominal resolution of 5 K. The ESI capillary voltage was 3.5 kV and the cone voltage was 20 V. The desolvation gas was nitrogen at a flow of 300 L/h and a temperature of 250 °C. The collision gas was nitrogen and the collision energy was 30 V.

Additional information for the individual MS experiments is available in the Supplementary data, which can be found in the electronic version of this article.

Results and Discussion

For CD-ECI to occur, we have previously determined that the interaction of the analytes with the Fe in the SS electrode is essential. Inhibition of this interaction led to a dramatic loss in sensitivity and replacement of the SS capillary with a Ni capillary led to a total loss in signal [6]. At the same time, a thorough understanding of these interactions can also be used to analytical advantage. For instance, we have recently shown that the sulfated mutremdamide A can be analyzed by CD-ECI, but binds so tightly to the SS capillary that ESI only detected the hydrolyzed compound [24].

Depending on the nature of the surface interaction between the analyte and the SS, the electrochemically

generated radicals can selectively induce fragmentation in a range of molecules, including glycopeptides, enolate forming phenylglycine containing peptides, phosphopeptides, and nucleosides that have been investigated in this study.

A most illustrative example of electrochemically-induced fragmentation using CD-ECI technique is the study of ristomycin A **1**. Figure 2 compares the chro-

Figure 2. Chromatogram and spectra of ristomycin A **1** using ESI and CD-ECI technique. (a) ESI chromatogram of **1**, and (b) ESI spectrum of **1** showing the expected doubly charged ions at $m/z = 1034.39$. (c) CD-ECI chromatogram of **1** showing two coeluting peaks; (d) spectrum of the first peak of **1** in CD-ECI, and (e) spectrum of the second peak of **1** in CD-ECI (all in TOF mode).

matograms and spectra of the glycopeptide **1** under ESI (Figure 2 a, b) and CD-ECI (Figure 2 c–e) technique. As expected, glycopeptide **1** eluted as one peak using ESI (Figure 2a). The corresponding ESI spectrum showed a predominantly doubly charged ion at $m/z = 1034.39$ (Figure 2b). In contrast, when investigated by CD-ECI, two LC peaks appeared in the chromatogram (Figure 2c) with differing mass spectra (Figure 2d, e). While the spectrum corresponding to the first peak (Figure 2d) showed similarities to the ESI spectrum, the spectrum corresponding to the second peak (Figure 2e) showed intense fragmentation representing the subsequent loss of the sugar moieties in **1** (Scheme 1). We propose that this differing behavior is intrinsically related to the involved mechanisms in ESI and CD-ECI. We have previously shown the importance of the surface chemistry of the SS capillary in CD-ECI [6, 24]. Glycopeptide **1** exists in equilibrium of keto- and enol-tautomers (Figure 2c inset). While the keto tautomer cannot be complexed by the Fe in the SS capillary, the enol tautomer can be complexed and, due to this online complexation, can be retained slightly longer on the SS tip than the keto form, leading to two co-eluting peaks in CD-ECI. Using ESI this phenomenon will not be observed, since it has previously been shown that compounds that are capable to chelate Fe (as is the case for the enolate, but not the keto form of ristomycin) are not detectable due to complexation when a SS capillary was used [25–27]. Therefore, the detection of the Fe-enol complex of **1** would not be expected in ESI, and indeed, only the earlier eluting keto form of **1** was detected, but

Scheme 1. Ristomycin A **1** and observed fragment assignments.

CD-ECI was able to detect both, the unfragmented keto (Figure 2d) and the fragmented enol (Figure 2e) tautomer of ristomycin. We propose this was enabled through the intrinsic CD-ECI mechanism: after complexation of the enol form with the Fe in the stainless steel capillary, the Fe can undergo oxidation via the CD-ECI process to form a radical cation organometallic moiety. This complexation with SS capillary is not unique and has previously been reported for phosphates [25–28]. In fact, it is the main reason why many proteomics researchers avoid SS capillary materials. After formation by CD-ECI, the radical cation can undergo free radical type fragmentation as evidenced by the type of fragmentation shown in Figure 2e. These radical fragments were no longer complexed by Fe and are thus released and detected by mass spectrometry. As can be seen in Figure 2e and Scheme 1, the observed product ions correspond to radically generated fragments of **1**. In addition, the normal ESI process ionized neutral fragments, so fragmentation and ionization are two distinct processes.

When we complexed the enol form of **1** with Na^+ instead of the Fe of the SS by adding NaI to the solution before the CD-ECI experiment, the described fragmentation did not occur, indicating that the complexation of Fe with the enol of **1** is necessary for both, radical ionization and fragmentation. In these experiments, the keto-enol equilibrium was shifted from the keto to the enol tautomer and then, while complexed with Na^+ , only one LC peak was observed corresponding to $[\text{M} + \text{Na}]^+$ (data not shown). In contrast to transition metals like Fe in SS, Na^+ would not be able to induce radical ionization and fragmentation since Na^+ has a 2p [6] outer electron shell is stable and highly unlikely to have another electron removed. This is also reflected in the high IP for the second removal of an electron from $\text{Na}^+/\text{Na}^{2+}$ (47.2 eV).

To further investigate the fragmentation mechanism, a number of experiments were designed using phosphates, nucleotides, and phenylglycine containing peptides that can either complex transition metals and fragment or, alternatively, cannot complex transition metals and thus cannot fragment using CD-ECI technique.

Phosphate containing compounds have previously been described to form complexes with Fe [25–28] and thus, Fe-complexed phosphopeptides should fragment using CD-ECI technique, but not using ESI technique. A standard phosphopeptide DLDVIPGRFDRRvPSVAAE **2** was used and investigated by infusion ESI and CD-ECI. As expected, Figure 3a shows the doubly, triply, and quadruply charged ions using ESI technique at m/z 1097.02, 731.34, and 549.01, respectively. In contrast, Figure 3b and c show additional fragmentation in CD-ECI technique not observed in ESI technique. This includes a series of b- and y-like ions listed in Table S1 and $[\text{M} + \text{H}]^+$, $[\text{M} + \text{H} - \text{H}_2\text{O}]^+$, $[\text{M} + \text{H} - \text{HPO}_3]^+$, $[\text{M} + \text{H} - \text{H}_3\text{PO}_4]^+$ (Figure 3b inset). Selected fragments in the m/z range from 540 to 700 are shown in

Figure 3. (a) ESI versus (b) CD-ECI spectrum of the phosphopeptide DLDVPIPIGRFDRRVpSVAAE 2 in TOF MS mode. The inset shows the expanded mass range $m/z = 2080$ – 2200 with the loss of H_2O , HPO_3 , and H_3PO_4 ; (c) expanded mass range 540 – 700 of CD-ECI spectrum of 2 indicating the phosphorylation site at b_{15}^{3+} (inverted filled triangle).

Figure 3c. The observed fragmentation ($b_{15}^{2+/3+/4+}$ and y_5^{1+} ; Table S1) allowed the unequivocal determination of the position of the phosphorylation site of 2. This

experiment showed that a direct infusion of phosphopeptides under CD-ECI could indeed fragment the peptide and, due to sufficient preservation of labile

phosphate groups, identify the phosphorylation site in a single TOF instrument.

The effect of exposure time to the SS capillary was investigated by lowering it to the timescale typical of an LC-MS experiment: A tryptically digested enolase spiked with phosphopeptides was separated on a C18 column and analyzed by CD-ECI. Figure 4 shows the chromatogram and the spectrum, respectively, of spiked phosphopeptide HLADLpSK 3 eluting at 7.48 min. The spectrum is characterized by backbone fragments of b_2 , b_3 , b_4 , b_5 , y_2 (indicating the site of phosphorylation), y_5 , y_6 , a_2 , and z_7 ions. In addition, loss of HPO_3 (-80) was observed for y_7 , y_5 , y_4 , y_3 , and y_2 . This experiment showed that indeed the complexation of phosphopeptides with Fe during LC is fast enough to allow specific fragmentation of phosphopeptides using CD-ECI technique. Somewhat similar to the MS^E approach [29, 30], where MS data are simultaneously collected using high and low collision energies, a co-eluting peak corresponding to IATAIEK did not interfere with the phosphopeptide assignment. In addition, there are already advanced software tools freely available that have been developed to evaluate concurrently

fragmented peptides, allowing the analysis of more complicated mixtures [31].

While the example phosphopeptides shown here are fragmented, there is a practical limitation to CD-ECI in that it requires the compound of interest to first bind to the SS capillary. This clearly excludes many compounds unless specific derivatizing reagents capable of causing diagnostic fragmentation of the desired compound are used that could circumvent the requirement for binding to the SS as a first step.

Towards this goal, we tested whether the addition of Cu(II) could function as such a derivatizing agent. Redox reactions of Cu complexes have previously been extensively studied in the gas phase [14, 16–20, 32–38]. It has previously been described that Cu(II) can be used to create molecular radical cations when electrosprayed as Cu(II)-amine complex. These cations were described to be different from those generated by ECD and metal-bearing peptide ternary complexes [18]. It was also shown that Cu can induce fragmentation [14]. Cu(II) was therefore added to adenosine congener 4 and analyzed by ESI and CD-ECI. When investigated by ESI, the nucleoside adenosine congener 4 ($C_{28}H_{32}N_8O_6$;

Figure 4. (a) LC-chromatogram of enolase spiked with phosphopeptide HLADLpSK 3. (b) CD-ECI spectrum of the LC-separated 3 in TOF MS mode. * Denotes a co-eluting peak corresponding to IATAIEK peptide; (inverted filled triangle) denotes the phosphorylated fragment ion at y_2 .

$M_R = 576.6$ Da) gave the spectrum shown in Figure 5a, b (without Cu(II) and with Cu(II)). In ESI, essentially no fragment ions were observed, and **4** gave $[M + H]^+$, $[M + 2H]^{2+}$, $[M + K]^+$, $[M + Na]^+$ ions.

In contrast, when investigated by CD-ECI, Fe or Cu(II)-chelated adenosine congener **4** showed diagnostically useful fragmentation with Cu(II) treatment leading to both, more intense total ion chromatograms (TICs) and fragmentation (Figure 5c, d and Supplementary Figure S1). Next to intact **4** represented by $[M + H]^+$, $[M + Na]^+$, $[M + K]^+$, $[M + 2H]^{2+}$, fragmentation was observed with the main fragment $[M + 2H-Rib]^+$ at $m/z = 445.18$ Da. A $\Delta 28$ impurity that was not identified in ESI was identified at $m/z = 473.20$ Da representing $[M + \Delta 28 + 2H-Rib]^+$ and at $m/z = 627.24$ Da representing $[M + \Delta 28 + Na]^+$. In addition, fragments were observed for $[M + 2H - Rib]^{2+}$, $[M + 3H - Rib - C_2H_4NH_2]^{2+}$, $[M + H - NH_2]^{2+}$, $[M + 2H - Rib - NH_2]^{2+}$ (Figure 5c, d).

What these experiments demonstrate is that chelation is necessary for fragmentation and that the fragmentation is the same regardless of whether **4** was chelated to the Fe in the SS (Figure 5c) or to Cu(II) before MS (Figure 5d). The increased TIC and fragmentation also indicated that Cu(II) addition can be used to induce fragmentation in compounds capable of chelating Cu. The obvious advantage of Cu(II) addition is that

one can increase incubation times before CD-ECI analysis and thus analyze samples that would otherwise not be amenable to CD-ECI.

To test this further, we synthesized UP14R **5** and 4F UP14R **6**, where U is phenylglycine and 4F U is 4-fluorophenylglycine and compared it with the commonly used standard P14R **7**, with and without prior Cu(II) treatment, both using ESI and CD-ECI. Phenylglycine and 4-fluorophenylglycine were specifically chosen because of their well-documented differences in exhibiting keto-enol tautomerism. Due to the electronegative fluorine in the para position, the α -hydrogen in 4F U is easily abstracted and the created enol structure is resonance-stabilized and energetically favored (as shown in Figure S2) [39]. In contrast, phenylglycine preferentially exists in the keto form, but can be trapped in the enol form under suitable chemical conditions, such as formation of a copper complex. While oligoproline **7** is prone to dissociation in CID experiments [40], out of these 12 experiments, only those three experiments that favor enolate complexation showed fragmentation in CD-ECI: first, CD-ECI after Cu(II) treatment of **5**, but not without prior Cu(II) treatment; second, CD-ECI of **6** with Cu(II) treatment; and third, CD-ECI of **6** without Cu(II) treatment. As expected from the mechanism that is dependent on the surface interaction of the enolate

Figure 5. Adenosine congener **4** investigated by (a) ESI without Cu(II), (b) ESI with Cu(II), (c) CD-ECI without Cu(II), and (d) CD-ECI with Cu(II). Fragmentation was only observed in CD-ECI technique regardless whether Cu(II) was added or not, indicating that Cu(II) can be used as a reagent to replace Fe.

In this study, we mainly observed b- and y-type ions as they would normally be expected during collision-induced dissociation according to the mobile proton theory [42, 43] and not necessarily in odd-electron driven pathways, although depending on the charge carrier and peptides structure, b-ion formation has also been reported for electron capture dissociation (ECD) [36, 44–48]. The previously described electron driven pathways [e.g., for electron capture dissociation (ECD) or electron-transfer dissociation (ETD)] were generated by the addition of an electron (reduction) as described in eq 1.

This is fundamentally different to CD-ECI, where ions were generated by the removal of an electron (oxidation) as shown in eqs 2 and 3. While the analyte is oxidized to $[M]^{+\cdot}$, $\text{Fe}^{+\cdot}$ is reduced to Fe, so that a new oxidation can be initiated.

There is precedence for radical oxidation reactions using Cu-complexes, where unusual radical-driven dissociation pathways including b- and y-ion formation, have previously been described after CID [20, 34, 35], when bi- or tridentate ligands (L) such as 2,2'-bipyridine and 2,2':6', 2''-terpyridine were part of the complex (eqs 4 and 5) [20, 34, 35].

While the bi- and tridentate ligands used in these studies form very stable complexes that can be detected by MS, we obtained weaker complexes and, thus, Cu(II) dissociated and only b- and y-ions were observed. This observation is in agreement with quantum chemical calculations for hypervalent nitrogen-protonated peptide radicals [49]. We thus propose that a Cu-complex was formed by reacting with the enolate of the peptide. The radical $\text{Cu}(\text{II})^{+\cdot}$ removes an electron from the peptide and thus initiates fragmentation into b- and y-type ions as indicated in eqs 6 and 7.

As has been observed for photodissociation, the fragmentation of the electron-deficient species is more sim-

ilar to CID even though no activation techniques were applied [50, 51]. With a few exceptions, such as z_1 and z_2 in **5**, fragment ions typical for ECD and ETD (c- and z-ions) were generally not observed. While we have initially considered ion-molecule reactions or electron impact ionization mechanism, the selectivity of the observed fragmentation exclude these mechanisms.

Conclusions

This study demonstrates that CD-ECI can be effectively used to fragment nucleosides, glycopeptides, and phosphopeptides into diagnostically useful ions in a single TOF instrument (without activation techniques). A four step mechanism was proposed: (1) complexation using either Fe in the SS capillary tip material or Cu(II) as an offline complexation reagent; (2) electrochemical oxidation of the complexed metal and thus formation of a radical cation (e.g., $\text{Fe} - e^- \rightarrow \text{Fe}^{+\cdot}$); (3) radical induced fragmentation of the complexed compound; (4) electrospray ionization of the fragmented neutrals. Fragmentation patterns resembling b- and y-type ions were observed. Like ECD or ETD techniques [52–56], the radical-driven fragmentation is softer than CID fragmentation. Therefore, phosphate groups were retained in fragment ions allowing the localization of the phosphorylation site. Unlike ECD or ETD, CD-ECI created an electron deficient species. The observed fragmentation is thus unique. Future investigation may be directed towards studies expanding the applications of the CD-ECI technique. The CD-ECI fragmentation platform is mechanically very simple and thus lends itself for miniaturization that can further increase achievable sensitivities [57]. In addition, it is our goal to determine the boundaries of the CD-ECI technique by exploring additional compound classes amenable to CD-ECI. Our current investigations focus on using CD-ECI to fragment proteins in a top-down like fashion.

Acknowledgments

This article is dedicated to Dr. Henry M. Fales, National Heart, Lung, and Blood Institute, National Institutes of Health. The authors acknowledge research support from the National Institute of Diabetes and Digestive and Kidney Diseases (Z01 DK070004-04), the Beckman Institute, and the Gordon and Betty Moore Foundation is gratefully acknowledged. The authors thank Wai-Ming Yau for synthesizing peptide UP14R. They thank Anastasia Kalli and Robert Graham for critical review of the manuscript.

Appendix A Supplementary Material

Supplementary material associated with this article may be found in the online version at [doi:10.1016/j.jasms.2010.08.018](https://doi.org/10.1016/j.jasms.2010.08.018).

References

1. Kebarle, P.; Verkerk, U. H. Electrospray: From Ions in Solution to Ions in the Gas Phase, What We Know Now. *Mass Spectrom. Rev.* **2009**, *28*, 898–917.
2. Blades, A. T.; Ikononou, M. G.; Kebarle, P. Mechanism of Electrospray Mass-Spectrometry—Electrospray as an Electrolysis Cell. *Anal. Chem.* **1991**, *63*, 2109–2114.
3. Van Berkel, G. J.; Kertesz, V. Using the Electrochemistry of the Electrospray Ion Source. *Anal. Chem.* **2007**, *79*, 5510–5520.
4. Druyvesteyn, M. J.; Penning, F. M. The Mechanism of Electrical Discharges in Gases of Low Pressure. *Rev. Mod. Phys.* **1940**, *12*, 0087–0174.
5. Baars-Hibbe, L. Characterization and Applications of High Frequency Discharges in the near-Atmospheric Pressure Range Using Micro-Structured Electrode Arrays; Cuvillier Verlag: Goettingen, 2005; p. 163.
6. Lloyd, J. R.; Hess, S. A Corona Discharge Initiated Electrochemical Electrospray Ionization Technique. *J. Am. Soc. Mass Spectrom.* **2009**, *20*, 1988–1996.
7. Trichel, G. W. The Mechanism of the Positive Point-to-Plane Corona in Air at Atmospheric Pressure. *Phys. Rev.* **1939**, *55*, 0382–0390.
8. Morrow, R. The Theory of Positive Glow Corona. *J. Phys. D Appl. Phys.* **1997**, *30*, 3099–3114.
9. Marginean, I.; Kelly, R. T.; Prior, D. C.; LaMarche, B. L.; Tang, K.; Smith, R. D. Analytical Characterization of the Electrospray Ion Source in the Nanoflow Regime. *Anal. Chem.* **2008**, *80*, 6573–6579.
10. Perrot, A. Recherches Sur L'action Chimique De L'ectincelle D'induction De L'appareil Ruhmkorff. *Annales de Chimie et de Physique.* **1861**, *61*, 161–199.
11. Thomson, J. J. *Notes on Recent Researches in Electricity and Magnetism, Intended as a Sequel to Professor Clerk-Maxwell's Treatise on Electricity and Magnetism.* The Clarendon Press: Oxford, 1893; pp 177–207.
12. Goldman, M.; Goldman, A.; Sigmond, R. S. The Corona Discharge, Its Properties and Specific Uses. *Pure Appl. Chem.* **1985**, *57*, 1353–1362.
13. Permentier, H. P.; Bruins, A. P. Electrochemical Oxidation and Cleavage of Proteins with On-Line Mass Spectrometric Detection: Development of an Instrumental Alternative to Enzymatic Protein Digestion. *J. Am. Soc. Mass Spectrom.* **2004**, *15*, 1707–1716.
14. Prudent, M.; Girault, H. H. On-Line Electrogeneration of Copper-Peptide Complexes in Microspray Mass Spectrometry. *J. Am. Soc. Mass Spectrom.* **2008**, *19*, 560–568.
15. Vilkov, A. N.; Laiko, V. V.; Doroshenko, V. M. Peptide Fragmentation Induced by Radicals at Atmospheric Pressure. *J. Mass Spectrom.* **2009**, *44*, 477–484.
16. Barlow, C. K.; Wee, S.; McFadyen, W. D.; O'Hair, R. A. Designing Copper(II) Ternary Complexes to Generate Radical Cations of Peptides in the Gas Phase: Role of the Auxiliary Ligand. *Dalton Trans.* **2004**, *20*, 3199–3204.
17. Chu, I. K.; Lam, C. N. W.; Siu, S. O. Facile Generation of Tripeptide Radical Cations In Vacuo Via Intramolecular Electron Transfer in Cu-II Tripeptide Complexes Containing Sterically Encumbered Terpyridine Ligands. *J. Am. Soc. Mass Spectrom.* **2005**, *16*, 763–771.
18. Chu, I. K.; Rodriguez, C. F.; Lau, T. C.; Hopkinson, A. C.; Siu, K. W. M. Molecular Radical Cations of Oligopeptides. *J. Phys. Chem. B* **2000**, *104*, 3393–3397.
19. Chu, I. K.; Siu, S. O.; Lam, C. N. W.; Chan, J. C. Y.; Rodriguez, C. F. Formation of Molecular Radical Cations of Aliphatic Tripeptides from Their Complexes with Cu-II(12-Crown-4). *Rapid Commun. Mass Spectrom.* **2004**, *18*, 1798–1802.
20. Lam, C. N.; Ruan, E. D.; Ma, C. Y.; Chu, I. K. Non-Zwitterionic Structures of Aliphatic-Only Peptides Mediated the Formation and Dissociation of Gas Phase Radical Cations. *J. Mass Spectrom.* **2006**, *41*, 931–938.
21. Hodyss, R.; Cox, H. A.; Beauchamp, J. L. Bioconjugates for Tunable Peptide Fragmentation: Free Radical Initiated Peptide Sequencing (FRIPS). *J. Am. Chem. Soc.* **2005**, *127*, 12436–12437.
22. Masterson, D. S.; Yin, H. Y.; Chacon, A.; Hachey, D. L.; Norris, J. L.; Porter, N. A. Lysine Peroxycarbamates: Free Radical-Promoted Peptide Cleavage. *J. Am. Chem. Soc.* **2004**, *126*, 720–721.
23. Chapman, J. R. Protein and Peptide Analysis by Mass Spectrometry; In *Methods in Molecular Biology*; Chapman, J.R.; Ed. Humana Press: Totowa, NJ, 1996; pp 9–28.
24. Plaza, A.; Bifulco, G.; Masullo, M.; Lloyd, J. R.; Keffer, J. L.; Colin, P. L.; Hooper, J. N.; Bell, L. J.; Bewley, C. A. Mutremdamide A and Koshikamides C-H, Peptide Inhibitors of HIV-1 Entry from Different Theonella Species. *J. Org. Chem.* **2010**, *132*, 9069–9077.
25. Asakawa, Y.; Tokida, N.; Ozawa, C.; Ishiba, M.; Tagaya, O.; Asakawa, N. Suppression Effects of Carbonate on the Interaction between Stainless Steel and Phosphate Groups of Phosphate Compounds in High-Performance Liquid Chromatography and Electrospray Ionization Mass Spectrometry. *J. Chromatogr. A* **2008**, *1198/1199*, 80–86.
26. Kim, J.; Camp, D. G. II; Smith, R. D. Improved Detection of Multi-Phosphorylated Peptides in the Presence of Phosphoric Acid in Liquid Chromatography/Mass Spectrometry. *J. Mass Spectrom.* **2004**, *39*, 208–215.
27. Tuijten, R.; Lemiere, F.; Witters, E.; Van Dongen, W.; Slegers, H.; Newton, R. P.; Van Onckelen, H.; Esmans, E. L. Stainless Steel Electrospray Probe: A Dead End for Phosphorylated Organic Compounds? *J. Chromatogr. A* **2006**, *1104*, 209–221.
28. Liu, S.; Zhang, C.; Campbell, J. L.; Zhang, H.; Yeung, K. K.; Han, V. K.; Lajoie, G. A. Formation of Phosphopeptide-Metal Ion Complexes in Liquid Chromatography/Electrospray Mass Spectrometry and Their Influence on Phosphopeptide Detection. *Rapid Commun. Mass Spectrom.* **2005**, *19*, 2747–2756.
29. Geiger, T.; Cox, J.; Mann, M. Proteomics on an Orbitrap Benchtop Mass Spectrometer Using All Ion Fragmentation. *Mol. Cell. Proteom.* **2010**, in press.
30. Plumb, R. S.; Johnson, K. A.; Rainville, P.; Smith, B. W.; Wilson, I. D.; Castro-Perez, J. M.; Nicholson, J. K. Uplc/Ms(E): A New Approach for Generating Molecular Fragment Information for Biomarker Structure Elucidation. *Rapid Commun. Mass Spectrom.* **2006**, *20*, 1989–1994.
31. Wong, J. W.; Schwahn, A. B.; Downard, K. M. Etiseq—an Algorithm for Automated Elution Time Ion Sequencing of Concurrently Fragmented Peptides for Mass Spectrometry-Based Proteomics. *BMC Bioinformatics* **2009**, *10*, 244.
32. Bagheri-Majidi, E.; Ke, Y. Y.; Orlova, G.; Chu, I. K.; Hopkinson, A. C.; Siu, K. W. M. Copper-Mediated Peptide Radical Ions in the Gas Phase. *J. Phys. Chem. B* **2004**, *108*, 11170–11181.
33. Chu, I. K.; Rodriguez, C. F.; Rodriguez, F.; Hopkinson, A. C.; Siu, K. W. M. Formation of Molecular Radical Cations of Enkephalin Derivatives Via Collision-Induced Dissociation of Electrospray-Generated Copper (II) Complex Ions of Amines and Peptides. *J. Am. Soc. Mass Spectrom.* **2001**, *12*, 1114–1119.
34. Karnezis, A.; Barlow, C. K.; O'Hair, R. A.; McFadyen, W. D. Peptide Derivatization as a Strategy to Form Fixed-Charge Peptide Radicals. *Rapid Commun. Mass Spectrom.* **2006**, *20*, 2865–2870.
35. Lam, C. N. W.; Siu, S. O.; Orlova, G.; Chu, I. K. Macrocyclic Effect of Auxiliary Ligand on the Gas-Phase Dissociation of Ternary Copper(II)-Gx Complexes. *Rapid Commun. Mass Spectrom.* **2006**, *20*, 790–796.
36. Laskin, J.; Yang, Z. B.; Lam, C.; Chu, I. K. Charge-Remote Fragmentation of Odd-Electron Peptide Ions. *Anal. Chem.* **2007**, *79*, 6607–6614.
37. Vaisar, T.; Gatlin, C. L.; Turecek, F. Oxidation of Peptide-Copper Complexes by Alkali Metal Cations in the Gas Phase. *J. Am. Chem. Soc.* **1996**, *118*, 5314–5315.
38. Vaisar, T.; Gatlin, C. L.; Turecek, F. Metal-Ligand Redox Reactions in Gas-Phase Quaternary Peptide-Metal Complexes by Electrospray Ionization Mass Spectrometry. *Int. J. Mass Spectrom. Ion Processes* **1997**, *162*, 77–87.
39. Ramamoorthy, A. NMR Spectroscopy of Biological Solids; CRC/Taylor and Francis Group: Boca Raton, 2006; p. 353.
40. Paizs, B.; Suhai, S. Fragmentation Pathways of Protonated Peptides. *Mass Spectrom. Rev.* **2005**, *24*, 508–548.
41. Dodds, E. D.; Hagerman, P. J.; Lebrilla, C. B. Fragmentation of Singly Protonated Peptides Via a Combination of Infrared and Collisional Activation. *Anal. Chem.* **2006**, *78*, 8506–8511.
42. Burlet, O.; Orkiszewski, R. S.; Ballard, K. D.; Gaskell, S. J. Charge Promotion of Low-Energy Fragmentations of Peptide Ions. *Rapid Commun. Mass Spectrom.* **1992**, *6*, 658–662.
43. Dongre, A. R.; Jones, J. L.; Somogyi, A.; Wysocki, V. H. Influence of Peptide Composition, Gas-Phase Basicity, and Chemical Modification on Fragmentation Efficiency: Evidence for the Mobile Proton Model. *J. Am. Chem. Soc.* **1996**, *118*, 8365–8374.
44. Cooper, H. J. Investigation of the Presence of B Ions in Electron Capture Dissociation Mass Spectra. *J. Am. Soc. Mass Spectrom.* **2005**, *16*, 1932–1940.
45. Haselmann, K. F.; Schmidt, M. Do B-Ions Occur from Vibrational Excitation Upon H-Desorption in Electron Capture Dissociation? *Rapid Commun. Mass Spectrom.* **2007**, *21*, 1003–1008.
46. Lee, S.; Chung, G.; Kim, J.; Oh, H. B. Electron Capture Dissociation Mass Spectrometry of Peptide Cations Containing a Lysine Homologue: A Mobile Proton Model for Explaining the Observation of B-Type Product Ions. *Rapid Commun. Mass Spectrom.* **2006**, *20*, 3167–3175.
47. Lee, S.; Han, S. Y.; Lee, T. G.; Chung, G.; Lee, D.; Oh, H. B. Observation of Pronounced B* Y Cleavages in the Electron Capture Dissociation Mass Spectrometry of Polyamidoamine (Pamam) Dendrimer Ions with Amide Functionalities. *J. Am. Soc. Mass Spectrom.* **2006**, *17*, 536–543.
48. Liu, H.; Hakansson, K. Abundant B-Type Ions Produced in Electron Capture Dissociation of Peptides without Basic Amino Acid Residues. *J. Am. Soc. Mass Spectrom.* **2007**, *18*, 2007–2013.
49. Bakken, V.; Helgaker, T.; Uggerud, E. Models of Fragmentations Induced by Electron Attachment to Protonated Peptides. *Eur. J. Mass Spectrom.* **2004**, *10*, 625–638.
50. Debois, D.; Giuliani, A.; Laprevote, O. Fragmentation Induced in Atmospheric Pressure Photoionization of Peptides. *J. Mass Spectrom.* **2006**, *41*, 1554–1560.
51. Delobel, A.; Halgand, F.; Laffranchise-Gosse, B.; Snijders, H.; Laprevote, O. Characterization of Hydrophobic Peptides by Atmospheric Pressure Photoionization-Mass Spectrometry and Tandem Mass Spectrometry. *Anal. Chem.* **2003**, *75*, 5961–5968.
52. Hakansson, K.; Cooper, H. J.; Emmett, M. R.; Costello, C. E.; Marshall, A. G.; Nilsson, C. L. Electron Capture Dissociation and Infrared Multiphoton Dissociation Ms/Ms of an N-Glycosylated Tryptic Peptide to Yield Complementary Sequence Information. *Anal. Chem.* **2001**, *73*, 4530–4536.
53. Kelleher, R. L.; Zubarev, R. A.; Bush, K.; Furie, B.; Furie, B. C.; McLafferty, F. W.; Walsh, C. T. Localization of Labile Post-Translational Modifications by Electron Capture Dissociation: The Case of Gamma-Carboxyglutamic Acid. *Anal. Chem.* **1999**, *71*, 4250–4253.
54. Mirgorodskaya, E.; Roepstorff, P.; Zubarev, R. A. Localization of O-Glycosylation Sites in Peptides by Electron Capture Dissociation in a Fourier Transform Mass Spectrometer. *Anal. Chem.* **1999**, *71*, 4431–4436.

55. Shi, S. D. H.; Hemling, M. E.; Carr, S. A.; Horn, D. M.; Lindh, I.; McLafferty, F. W. Phosphopeptide/Phosphoprotein Mapping by Electron Capture Dissociation Mass Spectrometry. *Anal. Chem.* **2001**, *73*, 19–22.
56. Syka, J. E. P.; Coon, J. J.; Schroeder, M. J.; Shabanowitz, J.; Hunt, D. F. Peptide and Protein Sequence Analysis by Electron Transfer Dissociation Mass Spectrometry. *Proc. Natl. Acad. Sci. U.S.A.* **2004**, *101*, 9528–9533.
57. Sikanen, T.; Franssila, S.; Kauppila, T. J.; Kostainen, R.; Kotiaho, T.; Ketola, R. A. Microchip Technology in Mass Spectrometry. *Mass Spectrom. Rev.* **2010**, *29*, 351–391.