



OPEN

Ultrasensitive tapered optical fiber refractive index glucose sensor

Erem Ujah^{1,2,3}, Meimei Lai^{1,3} & Gymama Slaughter^{1,2}✉

Refractive index (RI) sensors are of great interest for label-free optical biosensing. A tapered optical fiber (TOF) RI sensor with micron-sized waist diameters can dramatically enhance sensor sensitivity by reducing the mode volume over a long distance. Here, a simple and fast method is used to fabricate highly sensitive refractive index sensors based on localized surface plasmon resonance (LSPR). Two TOFs ($l = 5$ mm) with waist diameters of $5\ \mu\text{m}$ and $12\ \mu\text{m}$ demonstrated sensitivity enhancement at $\lambda = 1559$ nm for glucose sensing (5–45 wt%) at room temperature. The optical power transmission decreased with increasing glucose concentration due to the interaction of the propagating light in the evanescent field with glucose molecules. The coating of the TOF with gold nanoparticles (AuNPs) as an active layer for glucose sensing generated LSPR through the interaction of the evanescent wave with AuNPs deposited at the tapered waist. The results indicated that the TOF ($\varnothing = 5\ \mu\text{m}$) exhibited improved sensing performance with a sensitivity of 1265%/RIU compared to the TOF ($\varnothing = 12\ \mu\text{m}$) at 560%/RIU towards glucose. The AuNPs were characterized using scanning electron microscopy and ultraviolet-visible spectroscopy. The AuNPs-decorated TOF ($\varnothing = 12\ \mu\text{m}$) demonstrated a high sensitivity of 2032%/RIU toward glucose. The AuNPs-decorated TOF sensor showed a sensitivity enhancement of nearly 4 times over TOF ($\varnothing = 12\ \mu\text{m}$) with RI ranging from 1.328 to 1.393. The fabricated TOF enabled ultrasensitive glucose detection with good stability and fast response that may lead to next-generation ultrasensitive biosensors for real-world applications, such as disease diagnosis.

One of modern medicine's biggest challenges is developing cost-effective technologies that can diagnose a disease in a timely and accurate manner and unaffected by electromagnetic interference (EMI). The label-free optical sensor offers a promising approach to biochemical sensing in nearly any environment, including those with EMI^{1–3}. Most binding events, such as DNA hybridization, antibody-antigen recognition, chemical reactions, and changes in concentration, usually lead to changes in the optical sensor's surrounding environment and are known to change the sensing environment refractive index (RI). Moreover, this change in RI can quantitatively reflect the biosensors' detection ability. Therefore, measuring the small changes in RI that may result from a biochemical process is critical for biomarker detection^{4,5}. Fiber optic-based RI sensors are characterized as label-free, with a variety of configurations. The most common fiber-based RI sensor are the Bragg grating (FBG) structures⁶, long-period gratings (LPG) forming a Mach–Zehnder interferometer⁷, micro-interferometers based on chemical etching⁸, and microstructured fibers⁹, and taper optical fiber¹⁰.

The material of the optical fiber is usually silica, which is non-toxic, eco-friendly, and much more resistant to corrosion than most materials, thus being a good candidate for sensing in harsh environments. Optical fibers provide excellent sensing versatility where they can be decorated with various materials (polymers, nanomaterials, etc.) or simply by changing the fiber configuration (interferometer, optrode, whispering gallery mode, etc.). Recently, tapered optical fibers (TOF) have attracted significant attention due to their ease of fabrication and enhanced optical properties^{11–13}. TOF produces remarkably high optical intensity over a long distance, from several millimeters to a few centimeters. The long interaction length and high intensity of the TOF can enhance the light-matter interaction, thereby increasing the sensor's sensitivity. Although there are various approaches to fabricating TOFs, as illustrated in Table 1, the flame brushing method is the most used, wherein a single mode or multimode glass optical fiber with the cladding removed is heated at the center of the fiber while simultaneously stretching the fiber at both ends to produce a symmetrical “waist” in the fiber¹⁰. This is the most straightforward and low-cost approach to fabricating a TOF.

Furthermore, TOFs have been demonstrated to contain less than 100 photons at a time in the interaction (tapered) region using ultralow power level nonlinear spectroscopy¹⁴. This low dose of radiation also can reduce

¹Center for Bioelectronics, Old Dominion University, Norfolk, VA 23508, USA. ²Department of Electrical and Computer Engineering, Old Dominion University, Norfolk, VA 23508, USA. ³These authors contributed equally: Erem Ujah and Meimei Lai. ✉email: gslaught@odu.edu

Fabrication method	Advantages	Disadvantages
Flame brushing ¹⁰	Simple way and low-cost Varying lengths and diameters	Lack of uniformity of applied heat Rough surfaces/insertion loss
Laser processing ⁸	Sub-micron dimensions Smooth surfaces	Requires specialized equipment Insertion loss
Electrical heating ⁹	No turbulence in comparison to flame brushing	Non-even heating distribution
Chemical etching ¹⁰	High-quality Highly reproducible	Precision in hydrofluoric acid: time ratio Limited control over the shape

Table 1. Tapered optical fiber fabrication methods.

the risk of sample damage in the sensing environment. Another key advantage to using TOF is that they are very steady over time and can self-clean via fluid motion¹⁵. TOF systems have been reported for biochemical sensing^{16–23}. In addition, the functionalization of nanoparticles on TOF has been reported to enhance fiber optical-based RI sensor's property by creating localized surface plasmonic resonance (LSPR) that engages the electrons in the nanoparticles to oscillate coupling to the evanescent wave resulting in a significant change in refraction index²⁴. Different types of nanoparticles, such as gold and silver^{25,26}, magnetic nanoparticles^{27,28}, carbon-based nanoparticles^{29,30}, latex nanoparticles³¹, and liposome-based nanoparticles³² have been reported to enhance the surface plasmonic resonance signal for target analyte detection. AuNPs are widely used for functionalizing TOF because of their strong induction of surface plasmonic resonance and their biocompatibility. Additionally, AuNPs are stable and resistant to oxidation, making them more durable for long-term applications. Lin et al.¹⁷ demonstrated a localized surface resonance TOF sensor using AuNPs. A waist diameter of 48 μm and length of 1.25 mm was achieved and then decorated with AuNPs. A sensitivity of 380%/RIU with RI ranging from 1.333 to 1.403 was reported. Tai and Wei³³ demonstrated an intensity sensitivity of up to 8000%/RIU using a tapered fiber tip to enhance the sensitivity of optical fiber sensors further due to the nanometer tip. The tapered fiber refractive index sensors show excellent sensitivity with a fast response time in the local change of index of refraction in a real-time manner, which has enormous potential for label-free biosensing.

Research efforts with TOF-based RI sensors have resulted in devices with varied sensitivity and reproducibility. To overcome this limitation, this paper examines varying taper waist diameters with a long waist length and the coupling of AuNPs at the tapered waist region to harness the evanescent field for enhancing the efficacy of the light coupling and, thus, the sensitivity. Two TOF RI sensors with a more extended waist length ($l = 5 \text{ mm}$) and smaller waist diameters of 5 μm and 12 μm were fabricated for glucose sensing. The waist diameter is critical in enhancing the sensitivity of TOF. The TOF ($\text{O} = 5 \mu\text{m}$) exhibits significantly improved glucose sensing capability in comparison to the TOF ($\text{O} = 12 \mu\text{m}$). However, the TOF ($\text{O} = 12 \mu\text{m}$) can be easily fabricated and is highly durable. AuNPs were prepared and used to decorate the surface of the TOF ($\text{O} = 12 \mu\text{m}$) for sensitivity enhancement. A power of less than 100 nW was used for the RI detection of glucose. This ultra-low power sensing provided experimental evidence for sensitive TOF RI detection at longer wavelengths (1559 nm) with a sensitivity of 2032%/RIU toward glucose with a linear RI range of 1.328–1.393. This results from the unique characteristics of AuNPs, such as their optical absorption and specific surface area, to enhance the light-matter interaction. The detection at the longer wavelength showed excellent sensitivity in the as-fabricated LSPR and other reported SPR systems^{34,35}. TOF sensors can serve as promising candidates for non-invasive glucose monitoring in healthcare and the bioprocessing industries^{36–39}.

Results and discussion

The experimental setup is shown in Fig. 1. The TOF sensing region was suspended above the canyon floor and submerged entirely in the target analyte. The input light (Thorlabs ASE-FL7002 White Light Test Source, 1530–1610 nm) propagated through the TOF through an optical attenuator and polarization controllers. The light output was coupled to an Optical Spectrum Analyzer (Thorlabs OSA203C Fourier Transform Optical Spectrum




Figure 1. Experimental setup for glucose sensing.

Analyzer, 1.0–2.6 μm). The TOF comprised a transition region with a smooth linear taper profile and a small uniform waist diameter ($\varnothing = 5 \mu\text{m}$ or $\varnothing = 12 \mu\text{m}$). A series of aqueous glucose solutions were prepared with mass ratios ranging from 0 to 45 wt%. The corresponding RI at 1559 nm was obtained.

UV–Vis spectroscopy and structure of AuNPs. Figure 2A shows the UV–Vis absorption spectrum of the citrate-capped AuNPs in water. During the AuNPs synthesis, citrate reduces Au(III) to Au. The AuNPs exhibit a strong absorption peak at 524 nm. The absorbance peak suggests that AuNPs are approximately 24–33 nm in size and agrees with earlier reports^{40–42}. Figure 2B show the scanning electron microscopy (SEM) image of the bare TOF and AuNPs-decorated TOF RI sensor to confirm the chemisorption of AuNPs.

Glucose sensing. The comparative analysis of the optical spectra data indicates that the two TOFs with waist diameters of 5 μm (Fig. 3A) and 12 μm (Fig. 3B) impart the power output of the RI sensor. The power out-




Figure 2. (A) UV–Vis spectra of the synthesized gold nanoparticles (AuNPs). (B) Scanning electron microscopy (SEM) of bare and AuNPs-decorated TOF.




Figure 3. Optical spectra of bare TOF sensors in different mass fractions of glucose solutions. (A) TOF ($\varnothing = 5 \mu\text{m}$ diameter) and (B) TOF ($\varnothing = 12 \mu\text{m}$ diameter). (C,D) Corresponding normalized intensity spectra.

put response for the bare TOFs is observed across the near-infrared (NIR) wavelength range of 1520–1630 nm. Upon exposure to various glucose concentrations (5–45 wt%), the power output response decreases with the increase in glucose concentrations. The observed trend agrees with previous reports for other biomolecules^{40,41}. The optical spectra were processed by integrating the area under the power output curve to generate the corresponding normalized power intensity curves depicted in Fig. 3C,D. At the peak wavelength of 1559 nm, the decrease in intensity can be clearly observed for each glucose concentration. No insignificant change was observed in the spectra after changing the light's polarization. This indicates that the TOF RI sensing is polarization insensitive, making it more practical⁴³.

The TOF with a larger waist diameter exhibited better reliability with reduced sensitivity. Therefore, to improve the sensitivity, AuNPs were coated on the surface of the fiber waist. From the AuNPs-decorated TOF ($\varnothing = 12 \mu\text{m}$) spectra (Fig. 4A), it is observed that the sensor's initial output in glucose is high due to the AuNPs coating. This is attributed to the inherently high surface area of the AuNPs and the LSPR, which enhance the interaction of glucose molecules with the sensing region, thereby dramatically altering the evanescent field. The enhanced evanescent field penetration depth is modulated by the RI and enables a more intense power output than the bare TOFs. As the RI decreases, the surface evanescent wave penetration depth decreases, resulting in little interaction between the evanescent field and the glucose molecules and little power output in the detection spectrum. The corresponding normalized power intensity curves in Fig. 4B show that the AuNPs coating significantly enhances the sensitivity of the TOF RI sensor and further depicts the occurrence of glucose molecule interaction with the sensing region. Figure 5 shows the calibration curve for the bare TOF and the AuNPs-decorated TOF sensors. The smaller waist diameter TOF ($\varnothing = 5 \mu\text{m}$) exhibited a sensitivity of 1265%/RIU compared to that of 560%/RIU for TOF (waist $\varnothing = 12 \mu\text{m}$) at $\lambda = 1559 \text{ nm}$, as shown in Fig. 5A. The higher sensitivity towards glucose was observed because of the enhanced evanescent field interactions with glucose achieved with the smaller waist diameter TOF. The intensity sensitivity of the TOF sensor is given by³³:




Figure 4. (A) Optical spectra of AuNPs decorated TOF ($\varnothing = 12 \mu\text{m}$ diameter) sensor in different mass fractions of glucose solutions and (B) Corresponding normalized intensity spectra.




Figure 5. (A) Power intensity changes against refractive index of bare TOFs (red and blue) and AuNPs decorated TOF at $\lambda = 1559 \text{ nm}$. (B) Comparison of sensitivity of TOF RI sensors.

$$S = \frac{\Delta I/I_0}{\Delta n} \times 100(\%/RIU).$$

A sensitivity of 2032% RIU at $\lambda = 1559$ nm was observed for the AuNPs-decorated TOF (Fig. 5A), representing an increase by a factor of ~ 4 compared to the bare TOF. The AuNPs generate an LSPR signal with high sensitivity to detect slight changes in the reaction. However, the AuNPs-decorated TOF sensor saturated at 30 wt% glucose, wherein the peak at 1559 nm disappears. This may result from glucose molecules adsorbing onto the sensing region. Overall, the AuNPs-decorated TOF generated more considerable power output changes than bare TOF. The sensitivity enhancement spans a broad spectrum range (1540–1610 nm) and is uniform. Additionally, the optical performance of the as-fabricated TOF sensor was compared with other optical fiber-based RI sensors tested in various analytes, as summarized in Table 2.

As the measured RI changed from 1.32 to 1.40, the change in light intensity was more significant, indicating that the AuNPs can produce sensitivity enhancement for TOF RI sensors. AuNPs also enable more contact areas for glucose molecules to interact at the sensor surface. A linear RI range of 1.34–1.40 ($r^2 = 0.9254$) was observed for the TOF ($\varnothing = 12 \mu\text{m}$), and a wider linear range of 1.32–1.40 ($r^2 = 0.9940$) was observed for the TOF ($\varnothing = 5 \mu\text{m}$). A similar linear range of 1.328–1.393 ($r^2 = 0.9781$) was observed for the AuNPs decorated TOF and agrees with or exceeds recently reported fiber optic RI sensors^{17,44}. A heterocore fiber exhibited an intensity sensitivity of 57.2%/RIU within a range of 1.334–1.377, and a MXene nanosheet-based heterocore fiber demonstrated a sensitivity of 510.1%/RIU with a range of 1.334–1.377⁴⁴. Therefore, the AuNPs-decorated TOF RI sensor enhances the evanescent field to detect the RI change in the sensing environment, making the sensor ideal for glucose detection. Figure 5B compares the TOF RI sensors' sensitivity at 1559 nm and 1590 nm. These uniform sensitivities indicate that the AuNPs-decorated TOF RI sensor is a good option for biochemical detection at 1559 nm and 1590 nm.

In summary, we described and demonstrated the use of TOF as RI sensors for detecting various glucose concentrations. As the glucose concentration increased from 5 to 45 wt%, the sensor's power output intensity decreased consequently. Additionally, the bare TOF's sensitivity is highly related to its diameter and insensitive to light polarization, which is ideal for real-world applications. The bare TOF ($\varnothing = 5 \mu\text{m}$) sensor demonstrated better sensing capabilities than the bare TOF ($\varnothing = 12 \mu\text{m}$). We achieved an intensity sensitivity of 1265%/RIU over the RI range of 1.328–1.393. Upon decorating the TOF ($\varnothing = 12 \mu\text{m}$) with AuNPs, the sensor's sensitivity increased by ~ 4 times, although the linear RI range decreased slightly. The sensitivity was 2032%/RIU over the RI range of 1.328–1.379. The smaller the waist diameter TOF, the more sensitive the sensor was to changes in the biochemical environment. And the AuNPs decoration on the TOF surface resulted in sensitivity enhancement due to the high surface-to-volume ratio enabled by the AuNPs for biomolecule adsorption and the generation of localized surface plasmon resonance. The TOF ($\varnothing = 12 \mu\text{m}$) preparation method is simple, robust, reproducible, and can easily be decorated with nanomaterials to enhance sensing capability. However, the micron-sized tapered fiber sensor is exceptionally fragile, and its mechanical strength will need to be improved for real-world applications. Future work will explore ways to package the TOF RI sensor for detecting biorecognition events in biochemical processes.

Methods

Fabrication of tapered optical fiber. The TOF was fabricated by the heat and pull method⁴⁵. Briefly, the fiber coating on the two ends and the middle part of a standard commercially available optical fiber (Thorlabs1060XP, Single Mode Optical Fiber, 980–1600 nm, Extra-High Performance, $\varnothing 125 \mu\text{m}$ Cladding, $\varnothing 5.8 \mu\text{m}$ core) was carefully removed and cleaned with acetone. A propane air mixing flame was placed under bare fiber, as depicted in Fig. 6. The bare fiber ends were placed on fiber holders (Newport 125 μm Fibers, 561 Series) and pulled by two Aerotech Pro115SL linear stages controlled by A3200 Controllers to attain a sensing length of 5 mm. The transmission of the fiber was simultaneously monitored while pulling. The fabricated TOF was mounted in a custom-made canyon-shaped Teflon jig, and the sensing region was cleaned with acetone, followed by ultrapure deionized (DI) water rinse¹³. The canyon holds the sensing liquid (total volume = 2 ml).

Gold nanoparticles synthesis and functionalization of TOF. As previously reported by Ma et al.⁴⁰, the gold nanoparticles were synthesized using a chemical method, as illustrated in Fig. 7. All glassware was cleaned with freshly prepared aqua regia (HCl:HNO₃, 3:1). The precursor solution was prepared using 0.01 g

Fiber type	Analyte	Material	Sensitivity	Linear range	Wavelength	References
TOF tip	Glycerin	bare	8000%/RIU	1.3–1.4	630 nm	³³
Coreless silica fiber	Glucose	bare	1467.59 nm/RIU	1.364–1.397	1550 nm	³⁷
Heterocore fiber	NaCl	Mxene	510.1%/RIU	1.3343–1.3765	500–1000 nm	⁴⁴
Heterocore fiber	NaCl	bare	57.2%/RIU	1.3343–1.3765	500–1000 nm	⁴⁴
U shape	Glucose	AuNPs, Glucose oxidase	2.899 nm/%; 5.101 dB/%	0.1–0.5%	1400–1600 nm	⁴⁵
Unclad fiber	Sucrose	AuNPs, goat anti-rabbit IgG	13.09 AU/RIU	1.34–1.41	300–1000 nm	⁴⁶
TOF	Glucose	Bare	1265%/RIU	1.328–1.393	1530–1610 nm	This work
TOF	Glucose	AuNPs	2032%/RIU	1.328–1.379	1530–1610 nm	This work

Table 2. Comparison of the analyte characteristic of optical fiber RI sensors.




Figure 6. Schematic illustration of the heat and pull method for fabricating the tapered optical fiber.




Figure 7. Gold nanoparticle (AuNPs) synthesis by sodium citrate reduction method.

of gold (III) chloride trihydrate ($\text{HAuCl}_4 \cdot 3\text{H}_2\text{O}$) dissolved in 100 mL DI water (18.2 M Ω cm) and brought to a boil under constant stirring with a magnetic stirrer on a hot plate. Immediately following boiling, 3.2 mL 1 wt% of trisodium citrate ($\text{Na}_3\text{C}_6\text{H}_5\text{O}_7$) solution was quickly added and stirred for 10 min. The $\text{Na}_3\text{C}_6\text{H}_5\text{O}_7$ serves as the reducing and capping agent in nanoparticle synthesis. The color of the solution changed from yellow to red wine during this period signifying the fast reduction of gold. The reaction solution was allowed to cool to room temperature. The obtained AuNPs were centrifuged and resuspended in DI water and stored at 4 °C when not in use. The UV–Vis spectra of AuNPs samples were recorded using UV–Vis absorption spectroscopy (Spectra Max M5 microplate reader, Molecular Devices, LLC).

Afterward, the TOF fiber was functionalized by the salinization process method⁴⁷. The sensing region of the TOF was cleaned with acetone, followed by a DI water rinse. After drying, the sensing region was immersed in freshly prepared piranha solution ($\text{H}_2\text{SO}_4:\text{H}_2\text{O}_2$, 3:1) for 30 min and thoroughly rinsed with DI water before drying at 90 °C for 30 min. Upon drying, the TOF was treated with 1% (v/v) solution of 3-aminopropyltriethoxysilane (APTES) in ethanol for 24 h at room temperature. The APTES functionalized TOF was rinsed with ethanol, followed by DI water rinse to remove any unbound APTES, and air dried. The functionalized TOF was incubated in 2 mL of the synthesized AuNPs solution for 8 h at room temperature to form a coating of AuNPs on the TOF sensing region. The AuNPs-decorated TOF was rinsed with DI water and used for RI sensing. The use of an organosilanes results in a very robust monolayer of metallic nanoparticles integration over the tapered optical fiber based on the chemisorption of nanoparticles⁴⁸.

Data availability

The datasets used and/or analyzed during the current study available from the corresponding author on reasonable request.

Received: 26 January 2023; Accepted: 7 March 2023

Published online: 18 March 2023

References

- Lai, M. & Slaughter, G. Label-free microRNA optical biosensors. *Nanomaterials* **9**(11), 1573 (2019).
- Andryukov, B. G., Besednova, N. N., Romashko, R. V., Zaporozhets, T. S. & Efimov, T. A. Label-free biosensors for laboratory-based diagnostics of infections: Current achievements and new trends. *Biosensors* **10**(2), 11 (2020).
- Armani, A. M., Kulkarni, R. P., Fraser, S. E., Flagan, R. C. & Vahala, K. J. Label-free, single-molecule detection with optical microcavities. *Science* **317**(5839), 783–787 (2007).
- Tazawa, H., Kanie, T. & Katayama, M. Fiber-optic coupler based refractive index sensor and its application to biosensing. *Appl. Phys. Lett.* **91**(11), 113901 (2007).
- Homola, J. Surface plasmon resonance sensors for detection of chemical and biological species. *Chem. Rev.* **108**(2), 462–493 (2008).
- Ahlfeldt, A. A. S. S. H. & Edwall, B. S. R. S. G. Fiber optical Bragg grating refractometer. *Fiber Integr. Opt.* **17**(1), 51–62 (1998).
- Vengsarkar, A. M. *et al.* Long-period fiber gratings as band-rejection filters. *J. Lightwave Technol.* **14**(1), 58–65 (1996).
- Gong, Y., Guo, Y., Rao, Y. J., Zhao, T. & Wu, Y. Fiber-optic Fabry–Pérot sensor based on periodic focusing effect of graded-index multimode fibers. *IEEE Photon. Technol. Lett.* **22**(23), 1708–1710 (2010).
- Ran, Z. L., Rao, Y. J., Liu, W. J., Liao, X. & Chiang, K. S. Laser-micromachined Fabry–Perot optical fiber tip sensor for high-resolution temperature-independent measurement of refractive index. *Opt. Express* **16**(3), 2252–2263 (2008).
- Deverall, J. E., Chesnoy, D. J. & Payne, D. N. Tapered single mode fiber optic directional couplers. *J. Light Wave Technol.* **10**, 639–643 (1992).

11. Tong, L., Zi, F., Guo, X. & Lou, J. Optical microfibers and nanofibers: A tutorial. *Opt. Commun.* **285**(23), 4641–4647 (2012).
12. Liyanage, T., Lai, M. & Slaughter, G. Label-free tapered optical fiber plasmonic biosensor. *Anal. Chim. Acta* **1169**, 338629 (2021).
13. Liyanage, T., Alharbi, B., Quan, L., Esquela-Kerschner, A. & Slaughter, G. Plasmonic-based biosensor for the early diagnosis of prostate cancer. *ACS Omega* **7**(2), 2411–2418 (2022).
14. Hendrickson, S. M., Lai, M. M., Pittman, T. B. & Franson, J. D. Observation of two-photon absorption at low power levels using tapered optical fibers in rubidium vapor. *Phys. Rev. Lett.* **105**(17), 173602 (2010).
15. Irigoyen, M., Sánchez-Martin, J. A., Bernabeu, E. & Zamora, A. Tapered optical fiber sensor for chemical pollutants detection in seawater. *Meas. Sci. Technol.* **28**(4), 045802 (2017).
16. King, B. J., Idehenre, I., Powers, P. E., Haus, J. W., & Hansen, K. M. Biosensing platform with tapered optical microfibers: new results. In *Frontiers in Biological Detection: From Nanosensors to Systems VI*, vol. 8933, 16–22. (SPIE, 2014).
17. Lin, H. Y., Huang, C. H., Cheng, G. L., Chen, N. K. & Chui, H. C. Tapered optical fiber sensor based on localized surface plasmon resonance. *Opt. Express* **20**(19), 21693–21701 (2012).
18. Kumar, S. *et al.* LSPR-based cholesterol biosensor using a tapered optical fiber structure. *Biomed. Opt. Express* **10**(5), 2150–2160 (2019).
19. Wieduwilt, T. *et al.* Gold-reinforced silver nanoprisms on optical fiber tapers—A new base for high precision sensing. *APL Photon.* **1**(6), 066102 (2016).
20. Kim, H. M., Park, J. H. & Lee, S. K. Fiber optic sensor based on ZnO nanowires decorated by Au nanoparticles for improved plasmonic biosensor. *Sci. Rep.* **9**(1), 1–9 (2019).
21. Korposh, S., James, S. W., Lee, S. W. & Tatam, R. P. Tapered optical fibre sensors: Current trends and future perspectives. *Sensors* **19**(10), 2294 (2019).
22. Rodríguez-Schwendtner, E., González-Cano, A., Díaz-Herrera, N., Navarrete, M. C. & Esteban, Ó. Signal processing in SPR fiber sensors: Some remarks and a new method. *Sens. Actuators B Chem.* **268**, 150–156 (2018).
23. Zhu, G. *et al.* Tapered optical fiber-based LSPR biosensor for ascorbic acid detection. *Photon. Sens.* **11**, 418–434 (2021).
24. Chen, C. & Wang, J. Optical biosensors: An exhaustive and comprehensive review. *Analyst* **145**(5), 1605–1628 (2020).
25. Cathcart, N., Chen, J. I. & Kitaev, V. LSPR tuning from 470 to 800 nm and improved stability of Au–Ag nanoparticles formed by gold deposition and rebuilding in the presence of poly(styrenesulfonate). *Langmuir* **34**(2), 612–621 (2018).
26. Jia, S., Bian, C., Sun, J., Tong, J. & Xia, S. A wavelength-modulated localized surface plasmon resonance (LSPR) optical fiber sensor for sensitive detection of mercury (II) ion by gold nanoparticles-DNA conjugates. *Biosens. Bioelectron.* **114**, 15–21 (2018).
27. Hobbs, K., Cathcart, N. & Kitaev, V. Gold-plated silver nanoparticles engineered for sensitive plasmonic detection amplified by morphological changes. *Chem. Commun.* **52**(63), 9785–9788 (2016).
28. Liu, X. *et al.* Surface plasmon resonance immunosensor for fast, highly sensitive, and in situ detection of the magnetic nanoparticles-enriched *Salmonella enteritidis*. *Sens. Actuators B Chem.* **230**, 191–198 (2016).
29. Jia, Y. *et al.* Magnetic nanoparticle enhanced surface plasmon resonance sensor for estradiol analysis. *Sens. Actuators B Chem.* **254**, 629–635 (2018).
30. Zhang, H. *et al.* A novel graphene oxide-based surface plasmon resonance biosensor for immunoassay. *Small* **9**(15), 2537–2540 (2013).
31. Lee, E. G. *et al.* Carbon nanotube-assisted enhancement of surface plasmon resonance signal. *Anal. Biochem.* **408**(2), 206–211 (2011).
32. Severs, A. H. & Schasfoort, R. B. M. Enhanced surface plasmon resonance inhibition test (ESPRIT) using latex particles. *Biosens. Bioelectron.* **8**(7–8), 365–370 (1993).
33. Tai, Y. H. & Wei, P. K. Sensitive liquid refractive index sensors using tapered optical fiber tips. *Opt. Lett.* **35**(7), 944–946 (2010).
34. Chen, S., Liu, Y., Yu, Q. & Peng, W. Self-referencing SPR biosensing with an ultralow limit-of-detection using long-wavelength excitation. *Sens. Actuators B Chem.* **327**, 128935 (2021).
35. Homola, J. On the sensitivity of surface plasmon resonance sensors with spectral interrogation. *Sens. Actuators B Chem.* **41**(1–3), 207–211 (1997).
36. Cano Perez, J. L. *et al.* Fiber optic sensors: A review for glucose measurement. *Biosensors* **11**(3), 61 (2021).
37. Novais, S., Ferreira, C. I., Ferreira, M. S. & Pinto, J. L. Optical fiber tip sensor for the measurement of glucose aqueous solutions. *IEEE Photon. J.* **10**(5), 1–9 (2018).
38. Fang, Y. L., Wang, C. T. & Chiang, C. C. A small U-shaped bending-induced interference optical fiber sensor for the measurement of glucose solutions. *Sensors* **16**(9), 1460 (2016).
39. Birks, T. A. & Li, Y. W. The shape of fiber tapers. *J. Lightwave Technol.* **10**(4), 432–438 (1992).
40. Ma, Y. *et al.* Colorimetric sensing strategy for mercury (II) and melamine utilizing cysteamine-modified gold nanoparticles. *Analyst* **138**(18), 5338–5343 (2013).
41. Haiss, W., Thanh, N. T., Aveyard, J. & Fernig, D. G. Determination of size and concentration of gold nanoparticles from UV–Vis spectra. *Anal. Chem.* **79**(11), 4215–4221 (2007).
42. He, Y. Q., Liu, S. P., Kong, L. & Liu, Z. F. A study on the sizes and concentrations of gold nanoparticles by spectra of absorption, resonance Rayleigh scattering and resonance non-linear scattering. *Spectrochim. Acta Part A Mol. Biomol. Spectrosc.* **61**(13–14), 2861–2866 (2005).
43. Shen, Z. & Du, M. High-performance refractive index sensing system based on multiple Fano resonances in polarization-insensitive metasurface with nanorings. *Opt. Express* **29**(18), 28287–28296 (2021).
44. Chen, Y. *et al.* Refractive index sensors based on Ti3C2Tx MXene fibers. *ACS Appl. Nano Mater.* **3**(1), 303–311 (2020).
45. Chen, K. C., Li, Y. L., Wu, C. W. & Chiang, C. C. Glucose sensor using U-shaped optical fiber probe with gold nanoparticles and glucose oxidase. *Sensors* **18**(4), 1217 (2018).
46. Shao, Y., Xu, S., Zheng, X., Wang, Y. & Xu, W. Optical fiber LSPR biosensor prepared by gold nanoparticle assembly on polyelectrolyte multilayer. *Sensors* **10**(4), 3585–3596 (2010).
47. Kyaw, H. H., Al-Harathi, S. H., Sellai, A. & Dutta, J. Self-organization of gold nanoparticles on silanated surfaces. *Beilstein J. Nanotechnol.* **6**(1), 2345–2353 (2015).
48. Pisco, M. & Cusano, A. Lab-on-fiber technology: A roadmap toward multifunctional plug and play platforms. *Sensors* **20**(17), 4705 (2020).

Acknowledgements

The research presented in this article was supported by the National Science Foundation Award 1921363 and 1921364. The authors thank the Micron-NSU Nanofabrication Cleanroom personnel and Dr. Sangram Pradhan and Dr. Messaoud Bahoura at the Marie V. McDemmond Center for Applied Research (MCAR) for their assistance with the material characterization.

Author contributions

G.S. and M.L. conceived the idea; M.L. and E.U. fabricated the tapered optical fibers, performed the optical experiments, collected the results; M.L., E.U., and G.S., participated in the analysis of data and wrote the manuscript. All authors reviewed the manuscript.

Competing interests

The authors declare no competing interests.

Additional information

Correspondence and requests for materials should be addressed to G.S.

Reprints and permissions information is available at www.nature.com/reprints.

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2023