

RESEARCH

Open Access

Tool to visualize and evaluate operator proficiency in laser hair-removal treatments

Seungwoo Noh¹, Woo Seok Koh², Hyoung-woo Lim¹, Chiyul Yoon¹, Youdan Kim³, Jin Ho Chung⁴, Hee Chan Kim^{5,6} and Sungwan Kim^{5,6*}

* Correspondence:

sungwan@snu.ac.kr

⁵Department of Biomedical Engineering, Seoul National University College of Medicine, Seoul 110-799, Korea

⁶Institute of Medical and Biological Engineering, Seoul National University, Seoul 151-742, Korea
Full list of author information is available at the end of the article

Abstract

Background: The uniform delivery of laser energy is particularly important for safe and effective laser hair removal (LHR) treatment. Although it is necessary to quantitatively assess the spatial distribution of the delivered laser, laser spots are difficult to trace owing to a lack of visual cues. This study proposes a novel preclinic tool to evaluate operator proficiency in LHR treatment and applies this tool to train novice operators and compare two different treatment techniques (sliding versus spot-by-spot).

Methods: A simulation bed is constructed to visualize the irradiated laser spots. Six novice operators are recruited to perform four sessions of simulation while changing the treatment techniques and the presence of feedback (sliding without feedback, sliding with feedback, spot-by-spot without feedback, and spot-by-spot with feedback). Laser distribution maps (LDMs) are reconstructed through a series of images processed from the recorded video for each simulation session. Then, an experienced dermatologist classifies the collected LDMs into three different performance groups, which are quantitatively analyzed in terms of four performance indices.

Results: The performance groups are characterized by using a combination of four proposed indices. The best-performing group exhibited the lowest amount of randomness in laser delivery and accurate estimation of mean spot distances. The training was only effective in the sliding treatment technique. After the training, omission errors decreased by 6.32% and better estimation of the mean spot distance of the actual size of the laser-emitting window was achieved. Gels required operators to be trained when the spot-by-spot technique was used, and imposed difficulties in maintaining regular laser delivery when the sliding technique was used.

Conclusions: Because the proposed system is simple and highly affordable, it is expected to benefit many operators in clinics to train and maintain skilled performance in LHR treatment, which will eventually lead to accomplishing a uniform laser delivery for safe and effective LHR treatment.

Keywords: Laser visualization, Performance evaluation, Photomedicine, Simulation bed, Training tool

Background

Approximately two decades after the emergence of the first FDA-approved laser, laser hair removal (LHR) treatment has become one of the most successful applications of lasers in medicine [1]. According to recent statistics from the American Society for Aesthetic Plastic Surgery (ASAPS), more than 1.2 million LHR procedures were performed in the U.S. during 2012, which were the third most frequent treatments in all cosmetic procedures and the second most frequent treatments for men [2]. The prevalence of LHR is assumed to be largely attributed to increased societal concerns about aesthetics, as well as its proven safety, efficacy, and greater simplicity than conventional epilation methods [3-5].

The idea of LHR is based on the theory of selective photothermolysis. This elaborates the differences in absorption rates of light energy between the hair follicle and tissue owing to the differences in chromophores [6]. Therefore, hair follicles can be selectively destroyed by exposing the laser to the target area without aiming at each follicle. According to more recent findings, the essence of photoepilation is not just the heat generation in the hair follicle but the conduction of heat to the hair stem cells, because the hair stem cell itself lacks an appreciable amount of chromophores and is located outside the outer root sheath of the hair follicle [7]. Therefore, the delivery of an appropriate dose of laser light is paramount for safe and effective photoepilation.

Failure in choosing the correct laser intensity level often causes side effects such as pigment alteration, blistering, and erythema owing to excessive heat generated in the tissue [8,9]. Insufficient delivery of laser light can also be problematic, leading to ineffective outcomes or even paradoxical hypertrichosis [10,11]. To minimize such side effects, many studies have been performed to determine the optimal set of laser parameters to treat various body sites and skin types of patients [12-15].

Even with the right intensity of laser light, however, the actual amount of energy delivered to the target presents local variations when the laser treatment spots are not evenly applied [16-18]. The potential threat of side effects from this nonuniformly delivered laser light is serious when we consider that these treatments are often delegated to nonphysicians or even nonmedical personnel who lack sufficient training [2,19]. Physicians must also practice or train on new laser devices, new applicator tips with different dimensions, and different settings of laser parameters [20].

Therefore, it is highly desirable to develop an affordable system to evaluate the operator's procedural performance in LHR treatments. In these cases, the spatial patterns of delivered laser light energy must be quantitatively accessed, but little attention has been focused on this issue thus far.

The difficulty in implementing such a system is the visualization of the irradiated area. Because infrared is used and no prompt marks are left on the treated area after LHR treatments, the irradiated spots cannot be easily traced. Recently, a group of researchers proposed a thermovision camera-based laser-visualizing method [21]. They successfully viewed the thermal changes in the skin and quantitatively analyzed the degree of overlaps and omissions. However, the use of an infrared camera is too expensive for general use. Moreover, it is only applicable to the postoperative assessment of performance accuracy and can be potentially erroneous when any type of skin-cooling mechanism is used. Other researchers utilized photosensitive paper and a camera to visualize the laser light [22] for investigating intensity profiles across a single spot to

analyze the nonideal characteristics of an LHR device. However, this methodology cannot be directly applied to the LHR performance evaluation system, primarily because photo-sensitive papers are not reusable and extra steps are required to digitize the results. Some commercial devices incorporate an auxiliary visible laser light to guide an operator for aiming the position of the laser spot; however, these are inapplicable to contact-type LHR devices and the trajectory of the irradiated laser spots cannot be traced.

In this research, a relatively simple system is proposed to visualize and analyze the delivery patterns of laser sources during a simulated LHR procedure. The system is intended for preclinical uses to evaluate the proficiency of operators and features affordability and simplicity, based on an off-the-shelf PC camera and digital image processing methods. The purpose of this study is: 1) to demonstrate that the proposed system can quantitatively reflect the performance level of LHR treatments, and 2) to test the applicability of the system in the field by training novice operators and comparing two different treatment techniques (sliding versus spot-by-spot). For these purposes, six novice operators are recruited and their performance is evaluated according to four performance indices during the simulated procedures of LHR.

Methods

Experimental apparatus


In the experiments, commercially available laser equipment, a PC camera, and polarizers were used with a simulation bed built in-house. Table 1 summarizes the specifications of the devices.

The simulation bed was made of aluminum profiles in a cuboid shape that was open on each side. To mimic the friction of human skin during LHR simulation, a sheet of semitransparent silicon rubber was placed on transparent glass on the top of the simulation bed. Inside the simulation bed, an angled mirror was installed (at 45 degrees) to reflect the laser light coming from the top to the front of the simulation bed. Figure 1 shows the simulation bed with the positioned camera.

A PC camera (SPC-A30M, Samsung, Seoul, Korea) was placed at a distance of 200 mm from the front side of the simulation bed to capture the laser light. Its position and viewing direction were adjusted to fully view the treatment area reflected in the mirror and to not distort the scene. By attaching a series of polarizers (visible linear

Table 1 Summarized specifications of the devices used in the experiments

Devices	Manufacturer	Size	Others
Laser equipment	LightSheer XC, LUMENIS, Inc.	20 × 20 mm ² (applicator tip) 12 × 12 mm ² (laser window)	Fluence: 10–100 J/cm ² Repetition: 1–2 Hz Wavelength: 800 nm
Simulation bed		400 × 250 × 250 mm ³	Made of 15 mm × 15 mm aluminum profiles
Silicon layer	Anonymous	400 × 250 mm ²	To mimic the skin friction
Glass layer	Anonymous	400 × 250 mm ²	To mechanically support the silicon layer
Mirror	Anonymous	400 × 350 mm ²	To reflect the laser from the top to the front
Camera	SPC-A30M, Samsung, Inc.	50 × 50 × 70 mm ³ (approx.)	Sensitivity: Visible and infrared regions Frame rate: 30 Hz Resolution: 640 × 480 pixels
Polarizers	Visible linear polarizing film, Edmund Optics, Inc.	Six orthogonally aligned 15 × 15 mm ² film cuts	transmission: > 40% at 800 nm



polarizing film, Edmund optics, Barrington, NJ) in front of the camera lens and taking advantage of the PC camera's sensitivity to the infrared lights, we could make it function as an infrared camera. Additionally, the polarizers protect the camera by significantly attenuating the intensity of input laser light and increasing the signal-to-noise ratio of the video images by removing background noises. Because the intensity of the laser light is much higher than visible-band light, the captured video images only contain the laser spot. Ideally, two orthogonally aligned polarizers can block all of the incoming lights. However, owing to practical discrepancies, six layers of polarizers were used in this experiment.

Contact-type diode laser equipment (LightSheer XC, LUMENIS, Santa Clara, CA) was used in the experiments. The device fires a single pulse of laser light when the trigger button on the laser applicator is pressed or a train of continuous pulses at a predetermined rate when the button is held. The fluence and repetition frequency of the laser are configurable, but were set to 25 J/cm² and 2 Hz, the most common settings used in clinics, during all experiments. The size of the applicator tip was larger than the size of the laser window owing to the cooling area located at its perimeter (Figure 2).

Study design

Six novice operators who had no prior knowledge about the LHR procedure were recruited (all men, 24–31 years) and simulated LHR treatment on the top of the simulation bed. Operators were asked to achieve a delivery as uniform as possible of the laser on a rectangular target area designated as 140 mm × 90 mm. The usage of gel was mandatory to simulate the real conditions of an LHR procedure. Prior to the experiments, a didactic lecture was given by an experienced dermatologist about the safety issues, principles, and techniques of LHR treatment. A real demonstration followed the lecture.

Two different techniques of treatments, spot-by-spot (SBS) and sliding, which are most frequently used in clinics, were simulated. In the SBS technique, a single laser pulse was fired at a time while the laser applicator was repeatedly placed on and removed from the skin by an operator. In the sliding technique, laser pulses were continuously fired while the laser applicator was slid, maintaining contact with skin




Figure 2 Applicator tip of the laser device used in the experiments. The size of the applicator tip is larger than the laser window owing to the cooling area located at its perimeter.

throughout the whole procedure. The order of the simulated techniques was randomly assigned to avoid the possible effect of habituation to the experimental apparatus.

Each simulated technique of treatment was composed of two separate sessions; the infrared camera installed inside the simulation bed recorded all simulated sessions. To simulate the training of novices, feedback on the procedural performance was given to the operator between the two sessions. Feedback was provided by viewing the image-processed video recorded during the prefeedback session, and erroneous laser delivery patterns such as overlapping and omissions of laser spots were confirmed automatically. Before initiating each session, operators were allowed free time to test and become accustomed to their strategy of treatment. Time for the free run was unlimited and varied among operators, ranging from 1 to 3 min. Figure 3 shows the overall study design.

This experimental protocol was approved by a local institutional review board (IRB No. C-1302-075-467 at Seoul National University Hospital) and conducted in accordance with




Figure 3 Design of simulation experiments. Two different treatments techniques were simulated and each treatment technique was composed of two recorded sessions of the simulation (filled circles). Feedback on the procedural performance was given between these two successive sessions.

the principles of the 2004 version of the Declaration of Helsinki. All subjects signed the informed consent form.

Image processing

A laser energy distribution map (LDM) was synthesized through a series of image processing algorithms on recorded video frames. Figure 4 shows the block diagram of the overall image processing steps.

First, calibration between physical and image space was made by locating four corners of the target area with 10 laser exposures. The calibration process revealed that 140 mm of physical dimension was equivalent to 397 pixels in the image space. The degree of distortion in the camera view and the errors in locating the laser spots were inconsequential.

Next, the frames containing single pulses of the laser were retrieved from the video recording of a simulated session. Because a laser spot is shown as a cluster of bright pixels in the image, the laser-exposed frames can be selected from a plot of mean gray-scale of each image frame. As exemplified in Figure 5, peaks in the plot represent laser-exposed frames. Some frames were not counted as valid laser-exposed frames; for example, those with no laser spot but an overall increase in brightness level at less than 30% of normal laser-spotted frames. These frames occurred when the laser was fired in the air, mostly during directional changes of the laser applicator during the simulated SBS technique.






Figure 4 Block diagram of image processing steps to synthesize an LDM. Laser-exposed frames were extracted from the video recorded during a simulation session. Each laser-exposed frame was binarized to clearly detect the laser spot, then its centroid position was computed. The LDM was reconstructed by overlaying the template image of the laser window to each centroid position of the detected laser spot.



Next, the position of the laser spot in each laser-exposed frame was located (Figure 6). Images were converted to black and white (binarization) with a threshold of 30% of maximum value in the frame. Then, the coordinates of the centroid of a white cluster were computed. Normally, the converted images contained only a single cluster of white pixels. However, in some cases, the reflected laser light from the aluminum profile appeared as additional white clusters. In such cases, the cluster with the largest area was regarded as the true laser spot.

Finally, a square template that represents the actual shape and size of a laser spot was overlaid at the position of the detected laser spot (Figure 6). The gray value of the pixels inside the template was set to a constant of 1, which reflects a uniform distribution



of the energy delivered from the laser light. This type of reconstruction process was necessary because the captured images of laser spots were blurred owing to the scattering effect of the silicon rubber sheet as well as the point spread characteristics of the camera system. A calibration process determined the size of the template to be 34×34 pixels. The LDM was built by accumulating all of these template-represented laser spots in the retrieved frames, allowing superimposition at overlapped regions (multiple doses of laser). Each pixel in the target area was initially set to zero and accumulated its value by the gray value of the template every time it was attributed to a laser spot.

Performance indices

For the quantitative analysis of the LDM, four performance indices are proposed by using general statistics. The first two indices, δ_0 and δ_z , are measures of the errors in the LDM, defined as the percent ratio of untreated and redundantly treated areas to the total area of the target, respectively, which is shown in Eqns. 1 and 2.

$$\delta_0[\%] = \frac{(A_0 - A_1)}{A_0} \times 100 \quad (1)$$

$$\delta_z[\%] = \frac{\sum_{k=2}^{\alpha} A_k}{A_0} \times 100 \quad (2)$$

Where

$$A_k = \sum_{m=1}^M \sum_{n=1}^N L(m, n) \circ k \quad k = 1, 2, \dots, \alpha \text{ and } A_0 = M \times N \quad (3)$$

$$x \circ y = \begin{cases} 1, & \text{if } x \geq y \\ 0, & \text{otherwise} \end{cases} \quad (4)$$

Here, α is the highest pixel value found in the LDM, $L(m, n)$, which denotes the maximum redundancy in the laser delivery. M and N represent the dimensions of the image in pixels (number of rows and columns, respectively). The maximum value of δ_z may exceed 100% because the target area can be treated redundantly with more than two laser exposures (multiple doses are counted by summing each A_k).

The μ is an index that represents an operator's estimation of the spot size, which is defined by using the mean of every two consecutive spot distances, d_c , as depicted in Eqn. 5.

$$\mu [\text{mm}] = \text{mean}(d_c) \times C \quad (5)$$

where

$$d_c = \{x_i \mid x_i = |S(i) - S(i+1)|, i = 1, 2, \dots, (\beta-1)\} \quad (6)$$

The constant C is the conversion ratio between the physical and image spaces, which was found to be 0.35 mm/pixel in the calibration step of our experiments. S is the array containing the centroid of the position of each laser spot. β is the number of laser spots exposed to the target. The ideal value of μ is equal to the actual size of laser window (which is 12 mm in this study).

The v is the measure of randomness in the LDM. It is formulated as a normalized form of distance variations measured from each laser spot to its nearest one:

$$v [\%] = \frac{\text{std}(d_n)}{\text{mean}(d_n)} \times 100 \quad (7)$$

where

$$d_n = \{y_i \mid y_i = \min(|S(i) - S(j)|), \quad \forall j \in \{1, 2, \dots, \beta\}\} \quad (8)$$

An increase in v may result in increases of both or either δ_0 and δ_z ; however, the reverse is not always true.

To illustrate the computation of performance indices, a synthetic LDM with three shots of lasers is shown in Figure 7. In this example, the size of the LDM and the laser spot were set to 13×12 and 5×5 pixels and the conversion ratio, C , was set to 1 (i.e. 1 mm = 1 pixel). The numbers in the pixels indicate the amount of laser exposure at the site, and the centroid position of each laser spot is marked in red. The computation process is as follows:

1. Maximum redundancy in the laser delivery $\alpha = 3$.




Figure 7 Synthetic LDM to illustrate the computation process of performance indices. The LDM has three overlapped laser spots and the pixel values indicate the amount of laser exposure at the site. Here, the maximum redundancy in laser delivery is 3 and the centroid position of each laser spot was marked as red. The indices δ_0 and δ_z are computed based on the pixel values, and μ and v are computed based on the centroid position of laser spots.

2. $A_0, A_1, A_2,$ and A_3 are 156, 58, 15, and 2, according to Eqn. 3.
3. $\delta_0 = 62.82\%$ and $\delta_z = 10.90\%$, according to Eqns. 1 and 2.
4. $S = \{(7, 5), (5, 8), (9, 7)\}$.
5. $d_c = \{3.61, 4.12\}$ and $d_n = \{2.83, 3.61, 2.83\}$, according to Eqns. 6 and 8.
6. $\mu = 3.87$ mm, $v = 0.15\%$, according to Eqns. 5 and 7.

Statistical comparisons

A total of 24 LDMs (six subjects, two techniques for treatments, and pre/postfeedback) were collected for the statistical comparison.

To validate the efficacy of proposed indices in assessing the performance of the LHR, an experienced dermatologist sorted the collected LDMs into three groups according to performance. Eight LDMs that presented the highest level of performance were selected first and designated as group G. The other 16 LDMs presented relatively poor performance compared to group G, but were sorted again into groups P_+ and P_- because there were alternative reasons for poor performance. Therefore, LDMs having too many overlaps were assigned to group P_+ and those with too many omissions to group P_- , with eight LDMs each.

The effect of the training was examined by comparing data between prefeedback and postfeedback sessions. The comparison was made for each treatment technique by comparing D1 and D2 for SBS technique and D3 and D4 for the sliding technique. Because we also hypothesized that there is a difference between treatment techniques, D1 was separately compared to D3 and D2 to D4, with the presence of feedback.


Statistical analysis of the experimental data was performed by using the Student's t-test for the matched paired, one-sample t-test, one-way ANOVA, and a Turkey post-test at a significance level of 0.05. Prior to the t-test and ANOVA, the normality of the data was assessed by using Shapiro-Wilk's method.

Results

Validation of performance indices

The results of ANOVA indicated that the mean values of performance indices are significantly different among performance groups (Figure 8). Group G showed lower values of δ_z and v than group P_+ . The v value of group G was also lower than that of group P_- . Group P_+ and P_- could be distinguished by using any single performance index except for v . The one-sample t-test result for μ showed that only group P_+ presented a significantly different value of μ from the actual laser window size of 12 mm (Table 2).

To exemplify different levels of performance, nine LDMs selected from three performance groups are shown in Figure 9. It is clear that the LDMs from group G presented better laser distribution than the rest. Specifically, groups P_+ and P_- exhibited more overlapping and omission, respectively, than group G. In these cases, the mean value of δ_z was 29.08 for group P_+ and 6.05 for group G. The value of δ_0 was 23.77 for group P_- and 13.12 for group G. Therefore, the quantitative analysis results are consistent with the results of visual inspection.



Training of operators


Four categories of data described in Table 3 were considered to study the effects of training (prefeedback versus postfeedback). The mean values in Table 4 represent the differences between D1 and D2 for the sliding technique and between D3 and D4 for the SBS technique.

The paired t-test results of the training indicate that the feedback was effective when the sliding technique was used, as indicated by the reduction in δ_0 and μ . The value of μ decreased from 13.49 mm to 11.70 mm, approaching the ideal value of 12 mm. The increase in δ_z was not desirable, but the degree of its change was not statistically significant. In contrast to the sliding technique, none of the performance indices changed meaningfully when the SBS technique was used. The values in Table 4 represent the differences between the indices in the postfeedback session computed relative to the prefeedback session.

Table 2 Deviation of μ from actual laser window size

	μ : 12 mm (mean \pm SD)	P
Group G	+0.50 \pm 2.17	0.537
Group P ₊	-1.65 \pm 2.1	0.002**
Group P ₋	1.59 \pm 2.74	0.144

**P < 0.01 Negative value indicates shorter spacing between consecutive laser spots.



Comparison of treatment techniques

To compare the two different treatment techniques, the differences between D1 and D3 were computed for the case of prefeedback. The same procedure was followed between D2 and D4 for the case of postfeedback. The results are summarized in Table 5. During prefeedback sessions, operators showed a higher value of v when the sliding technique was used than when the SBS technique was used. The remaining indices did not show differences between the treatment techniques. The differences were measured relative to the SBS technique.

Discussion

As pointed out in a previous study [23], methods to learn clinical skills are changing as the opportunities for learning through work with actual patients has diminished. The same holds true for LHR treatment because current residency programs in dermatology place insufficient emphasis on photodermatology and laser therapy [24], and even non-physician treatments have become prevalent [25]. The absence of validation methods is a problem in LHR treatment; because of this, no operators, even licensed ones, can initiate treatment with great confidence. The use of models or simulators has been the common practice to tackle this shortage of experience [23].

The proposed LHR evaluation tool successfully visualized simulated patterns of laser delivery and evaluated them in terms of four indices. Significant differences in indices among performance groups were found; however, none of these indices were sufficient

Table 3 Categories of collected data for statistical comparison

SBS technique		Sliding technique	
Prefeedback	Postfeedback	Prefeedback	Postfeedback
D1	D2	D3	D4

Table 4 Effects of feedback depending on treatment technique

	Sliding mode		SBS mode	
	(mean ± SD)	P	(mean ± SD)	P
δ_z [%]	+6.00 ± 8.03	0.127	+1.07 ± 5.27	0.641
δ_0 [%]	-6.32 ± 5.12	0.029*	-4.68 ± 6.14	0.121
μ [mm]	-1.79 ± 1.67	0.047*	-0.32 ± 1.08	0.504
v [%]	-0.77 ± 3.83	0.645	+0.75 ± 1.22	0.188

Prefeedback μ values were >12 mm, regardless of techniques.

(13.49 in sliding, and 13.86 in SBS).

Positive signs indicate a higher index value in postfeedback session.

* $P < 0.05$.

to characterize all of the performance groups at once. Therefore, the proposed indices should be used in combination to assess the level of performance.

The ideal situation of group G, the most well-performing group, should have a μ value of 12 mm and the lowest values in the rest of the indices. However, only the two indices, μ and v , corresponded to this expectation. Therefore, the quantitative description of satisfactory patterns in laser delivered LHR treatment aims to achieve minimum randomness in the spatial distribution of laser spots and accurate estimation of the size of laser spots. A moderate level of omissions and overlapping appears to be allowed; however, it has been deduced that the degree of overlapping is a more important factor than that of omissions when describing satisfactory performance, because group G could be distinguished from one of the poorly performing groups in terms of δ_z but not δ_0 .


The training was effective only when the sliding technique was used, as indicated by the improvements in δ_0 and μ values (Table 4). A careful examination of the changes in μ values is given in Figure 10. The more deviation an operator exhibited in μ from its ideal value of 12 mm, the more improvements were observed after training. The reduced δ_0 after the training is believed to be the result of a more accurate μ . In contrast to the sliding technique, operators' performance did not significantly improve when using the SBS technique. The difference of training effects between treatment techniques is hypothesized to be largely attributable to the characteristics of the gel. In the SBS technique, gels bear the marks of previous contact with the applicator tip at every departure of the applicator tip from the target surface. Operators may have been misled by the illusion of marks that inhibited them to reflect on the feedback from the system. This was confirmed by interviewing the operators as a group after finishing the experiments. Five of six operators replied that they took advantage of the marks on the gel to easily determine the position of the laser spot during the simulation of the SBS technique.

Table 5 Results of LDM analysis on the effects of treatment technique

	Prefeedback		Postfeedback	
	(mean ± SD)	P	(mean ± SD)	P
δ_z [%]	-2.87 ± 8.50	0.445	+2.06 ± 12.80	0.710
δ_0 [%]	+5.40 ± 6.78	0.108	+3.76 ± 13.31	0.520
μ [mm]	+1.63 ± 1.58	0.053	+0.16 ± 1.45	0.794
v [%]	+7.21 ± 5.79	0.028*	+5.69 ± 5.79	0.146

Positive signs indicate higher index values in the sliding technique.

* $P < 0.05$.



In the sliding mode, however, operators were not deceived by marks on the gel, because an operator had to adaptively control the amount of force applied to the laser applicator tip to cope with the varying frictions while maintaining contact with the surface. The results listed in Table 5, which compare treatment techniques, support this hypothesis. The higher value of v in the sliding technique indicates that the operators experienced more difficulties maintaining constant distances between successive laser spots while using this technique, particularly before feedback was provided. This result also supports the reason for recommending the SBS technique to novice operators in clinics. Therefore, operators are required to be well-informed about the effects of the gels on performance, depending on the technique used for the LHR treatment.

It was interesting to note the trend of elevated prefeedback μ values, regardless of treatment techniques (Table 4). This indicates that the operators overestimated the size of the laser spot, which is assumed to be attributable to the mismatch in size between the applicator tip and the laser window. Because the actual size of the laser-emitting window was smaller than that of the applicator tip, and only the top view of the applicator tip was visible during the treatment, operators were susceptible to overestimating the size of the laser spot.

The presented study has three limitations, described as follows. First, the aforementioned criterion of satisfactory delivery of the laser in LHR treatment can be prematurely generalized, because the performance groups presented in this paper reflect only a single professional dermatologist's opinion. More general wisdom—for example, a composite score of LHR proficiency—will be found when additional tests are conducted with a group of dermatologists by comparing their simulated results against those of novices. Second, owing to the practical limitations in utilizing clinical settings, the system was tested by a small number of operators; therefore, the statistical analysis

performed in this study may have a high probability for type II error. In other words, the performance indices might be more significant for characterizing performance groups and observing the effects of training. Additionally, a prospective study might be conducted to confirm the long-term efficacy of the system in operator training; however, the usage of simulators has proven effectiveness in other medical fields for the acquisition of particularly junior levels of maneuverability [23]. Third, the system might not be applicable to noncontact types of LHR devices, so the simulation of treatments on curved areas of the body would not be feasible. Because a mirror was used to reflect the laser beam to the camera, the location of the laser spot can be erroneous as the laser applicator deviates from the vertical line to its target surface. However, this limitation does not seriously degrade the fidelity of the simulation. Referring to the results from a research group that recently analyzed patients' thermal deposit images during cosmetic laser procedures [21], the patterns in experts' laser delivery (Figure 7 in the study [21]) resembles those of well-performing groups in our experiments (Figure 9 in this experiment). In the previous study, operators also tried to cover the target area with uniform but not overlapping laser spots, as in our experiments. Although the percent errors of the omitted and overlapped areas were not directly comparable owing to some differences in the methods used for computing indices, the similarities between the simulated and clinical trials ensure the applicability of the proposed system in operator training. The pulse rate of the laser is another aspect of hardware limitation. A normal webcam can capture up to 30 frames of images per second; therefore, lasers pulsed higher than this rate may not be correctly detected. However, such settings are uncommon in clinical practice.

Even with these limitations, it is expected that the proposed system could be effectively utilized in clinics, because it is a cost-effective and intuitive solution to visualize and evaluate the proficiency of LHR treatment. The system might be improved to be used in other areas of photomedicine. For example, laser applications in the treatment of pigmented lesions and facial rejuvenation require different degrees of laser overlaps and patterns, depending on the need of the patient and the treatment settings [26,27]. Therefore, the performance indices might be adjusted to reflect the general guidelines or an expert's strategy for various treatments. An improvement to the software is also expected. Accounting for thermal relaxation time, rather than simply counting the number of redundant exposures, will be more advantageous in predicting actual thermal damage to the tissue.

Conclusions

A uniform laser delivery during LHR treatment is significant for safe and effective treatment. A highly affordable system, which is also easy to operate, has been developed in this research. The proposed system was able to visualize and evaluate laser patterns during preclinical trials without using an expensive infrared camera. For the study, four useful performance indices were proposed for assessing the proficiency of operators during LHR treatment. With these indices, the developed system could quantitatively analyze an operator's proficiency in LHR treatment. A performance analysis of the proposed, affordable system has shown that operators reduced omission errors by 6.32% and accurately estimated the spot distances to match the actual size of the laser-emitting window. Further, the proposed system was used as a scientific tool for the comparison of

two different treatment techniques (sliding versus spot-by-spot) by observing the performances. Therefore, the proposed training system is expected to benefit many operators in clinical practice and to maintain skilled performance in LHR treatment, which may result in eventually accomplishing a uniform laser delivery treatment. The proposed system may also be applicable to other areas of photomedicine.

Competing interests

The authors declare that they have no competing interests.

Authors' contributions

SN proposed an idea to visualize and analyze dermatologic laser treatments and played a leading role in preparing the manuscript. WSK, as a medical professional, has established a niche in the quantification of laser treatment and played a consultant role in designing the study. HL participated in designing the experimental apparatus and was involved in drafting the manuscript. CY participated in the study design and provided useful tips for performing statistical analysis. YK actively participated in drafting the manuscript to increase the integrity of the data. JHC played a consultant role in designing the study as a medical professional. HCK and SK conceived of the study, participated in its design and coordination, and helped to draft the manuscript. All authors read and approved the final manuscript.

Acknowledgements

This work was supported in part by the Seoul National University (SNU) Foundation Research Expense (Grant No. 800-20100525), by the SNU College of Medicine (Grant No. 800-20120023), and by Seoul Ocean Aquarium (No. 0411-20120068).

Author details

¹Interdisciplinary Program for Bioengineering, Seoul National University, Seoul 110-744, Korea. ²JMO Dermatology, Seoul 135-887, Korea. ³Department of Mechanical and Aerospace Engineering, Seoul National University College of Engineering, Seoul 151-742, Korea. ⁴Department of Dermatology, Seoul National University Hospital, Seoul 110-744, Korea. ⁵Department of Biomedical Engineering, Seoul National University College of Medicine, Seoul 110-799, Korea. ⁶Institute of Medical and Biological Engineering, Seoul National University, Seoul 151-742, Korea.

Received: 8 January 2014 Accepted: 1 April 2014

Published: 8 April 2014

References

1. Nouri K, Vejjabhinanta V, Patel SS, Singh A: **Photoepilation: a growing trend in laser-assisted cosmetic dermatology.** *J Cosmet Dermatol* 2008, **7**(1):61-67.
2. *ASAPS 2012 Cosmetic Surgery National Data Bank Statistics.* 2012 [http://www.surgery.org/sites/default/files/ASAPS-2012-Stats.pdf]
3. Gan SD, Graber EM: **Laser hair removal: a review.** *Dermatol Surg* 2013, **39**(6):823-838.
4. Ibrahim OA, Avram MM, Hanke CW, Kilmer SL, Anderson RR: **Laser hair removal.** *Dermatol Ther* 2011, **24**(1):94-107.
5. Lapidoth M, Dierickx C, Lanigan S, Paasch U, Campo-Voegel A, Dahan S, Marini L, Adatto M: **Best practice options for hair removal in patients with unwanted facial hair using combination therapy with laser: guidelines drawn up by an expert working group.** *Dermatol* 2010, **221**(1):34-42.
6. Anderson RR, Parrish JA: **Selective photothermolysis: precise microsurgery by selective absorption of pulsed radiation.** *Science* 1983, **220**(4596):524-527.
7. Altshuler GB, Anderson RR, Manstein D, Zenzie HH, Smirnov MZ: **Extended theory of selective photothermolysis.** *Lasers Surg Med* 2001, **29**(5):416-432.
8. Lim SPR, Lanigan SW: **A review of the adverse effects of laser hair removal.** *Lasers Med Sci* 2006, **21**(3):121-125.
9. Lanigan SW: **Incidence of side effects after laser hair removal.** *J Am Acad Dermatol* 2003, **49**(5):882-886.
10. Willey A, Torrontegui J, Azpiazu J, Landa N: **Hair stimulation following laser and intense pulsed light photo-epilation: review of 543 cases and ways to manage it.** *Lasers Surg Med* 2007, **39**(4):297-301.
11. Desai S, Mahmoud BH, Bhatia AC, Hamzavi IH: **Paradoxical hypertrichosis after laser therapy: a review.** *Dermatol Surg* 2010, **36**(3):291-298.
12. Liew SH: **Unwanted body hair and its removal: a review.** *Dermatol Surg* 1999, **25**(6):431-439.
13. Battle EF: **Advances in laser hair removal in skin of color.** *J Drugs Dermatol* 2011, **10**(11):1235-1239.
14. Vachiramon V, McMichael AJ: **Laser hair removal in ethnic skin: principles and practical aspects.** *J Drugs Dermatol* 2011, **10**(Suppl 12):s17-s19.
15. Ataie-Fashtami L, Shirkavand A, Sarkar S, Alinaghizadeh M, Hejazi M, Fateh M, Djavid GE, Zand N, Mohammadreza H: **Simulation of heat distribution and thermal damage patterns of diode hair-removal lasers: an applicable method for optimizing treatment parameters.** *Photomed Laser Surg* 2011, **29**(7):509-515.
16. Town G, Ash C, Dierickx C, Fritz K, Bjerring P, Haedersdal M: **Guidelines on the safety of light-based home-use hair removal devices from the European Society for Laser Dermatology.** *J Eur Acad Dermatol Venereol* 2012, **26**(7):799-811.
17. Town G, Ash C: **Are home-use intense pulsed light (IPL) devices safe?** *Lasers Med Sci* 2010, **25**(6):773-780.
18. Town G, Ash C: **Measurement of home-use laser and intense pulsed light systems for hair removal: preliminary report.** *J Cosmet Laser Ther* 2009, **11**(3):157-168.
19. Alam M: **Who is qualified to perform laser surgery and in what setting?** *Semin Plast Surg* 2007, **21**(3):193-200.
20. Greve B, Raulin C: **Professional errors caused by lasers and intense pulsed light technology in dermatology and aesthetic medicine: preventive strategies and case studies.** *Dermatol Surg* 2002, **28**(2):156-161.

21. Koproński R, Wilczyński S, Samojedny A, Wróbel Z, Deda A: **Image analysis and processing methods in verifying the correctness of performing low-invasive esthetic medical procedures.** *Biomed Eng Online* 2013, **12**(1):51.
22. Thomas G, Ash C, Hugtenburg R, Kiernan M, Town G, Clement M: **Investigation and development of a measurement technique for the spatial energy distribution of home-use intense pulsed light (IPL) systems.** *J Med End Technol* 2011, **35**(3-4):191-196.
23. Reznick RK, MacRae H: **Teaching surgical skills—changes in the wind.** *N Engl J Med* 2006, **355**(25):2664-2669.
24. Freiman A, Barzilai DA, Barankin B, Natsheh A, Shear NH: **National appraisal of dermatology residency training: a Canadian study.** *Arch Dermatol* 2005, **141**(9):1100-1104.
25. Kelsall D: **Laser hair removal: no training required?** *CMAJ* 2010, **182**(8):743.
26. Polder KD, Landau JM, Vergilis-Kalner IJ, Goldberg LH, Friedman PM, Bruce S: **Laser eradication of pigmented lesions: a review.** *Dermatol Surg* 2011, **37**(5):572-595.
27. Chan CS, Saedi N, Mickle C, Dover JS: **Combined treatment for facial rejuvenation using an optimized pulsed light source followed by a fractional non-ablative laser.** *Lasers Surg Med* 2013, **45**(7):405-409.

doi:10.1186/1475-925X-13-40

Cite this article as: Noh et al.: Tool to visualize and evaluate operator proficiency in laser hair-removal treatments. *BioMedical Engineering OnLine* 2014 **13**:40.

**Submit your next manuscript to BioMed Central
and take full advantage of:**

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit

