MEETING ABSTRACT



Open Access

The potential of TOF PET-MRI for reducing artifacts in PET images

Andrei lagaru^{*}, Ryogo Minamimoto, Craig Levin, Amid Barkhodari, Mehran Jamali, Dawn Holley, Zaharchuk Greg

From PSMR 2015: 4th Conference on PET/MR and SPECT/MR La Biodola, Isola d'Elba, Italy. 17-21 May 2015

Stanford University, USA

Here we evaluated the potential of TOF PET/MRI to reduce various PET image artifacts, by comparing the images to non-TOF PET/MRI, TOF PET/CT and non-TOF PET/CT.

Methods and materials

All patients underwent a single-injection of FDG, followed first by PET-CT and subsequently by PET-MRI scan. The PET/CT exams were performed using a GE Discovery 690 PET/CT scanner. The PET/MRI images were acquired on a GE Signa PET/MRI scanner. All PET images were reconstructed with and without the TOF data. Visual analysis of these series was performed for dental metal / breathing artifcats and 3) artifacts caused by high excretion of FDG in the bladder. PET image quality was evaluated using a 3-point scale (1 - clinically significant artifact; 2 - non clinically-significant artifact; and 3 - no artifact). Data from 18 oncologic patients (mean \pm SD age: 55 \pm 10 years; female 7, male 11) were used. The average scores of TOF PET/MRI, non-TOF PET/MRI, TOF PET/CT and non-TOF PET/CT for dental artifacts were 3.0, 2.8, 2.4 and 2.3, respectively; for breathing artifacts were 3.0, 2.5, 2.5 and 2.3, respectively; and for pelvic artifacts were 2.9, 1.6, 2.1 and 1.4, respectively. TOF PET/MRI had the highest image quality scores among the 4 series of PET data analyzed for these types of artifacts. TOF PET/MRI showed promising results in reduction of various PET artifacts in this cohort, when compared to non-TOF PET/MRI, TOF PET/CT and non-TOF PET/CT. This may be due to the better timing resolution (<400 ps) for the PET/MR system compared to the PET/CT system (>600 ps).

Published: 18 May 2015

doi:10.1186/2197-7364-2-S1-A77 **Cite this article as:** lagaru *et al*:: The potential of TOF PET-MRI for reducing artifacts in PET images. *EJNMMI Physics* 2015 **2**(Suppl 1):A77.

SpringerOpen[⊍]

© 2015 lagaru et al; licensee Springer. This is an Open Access article distributed under the terms of the Creative Commons Attribution License (http://creativecommons.org/licenses/by/4.0), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.