



P123 Beyond Diameters, Flow Hemodynamics Quantified by Magnetic Resonance Imaging to Help Characterizing Aneurysmal Aorta

Sophia Houriez—Gombaudo-Saintonge^{1,2,*}, Ariel Pascaner³, Gilles Soulat⁴, Umit Gencer⁵, Thomas Dietenbeck⁶, Damian Craiem⁷, Emilie Bollache⁶, Yasmina Chenoune², Elie Mousseaux⁴, Nadja Kachenoura¹

¹Laboratoire d'Imagerie Biomédicale

²ESME Research Lab

³IMETTYB - Universidad Favaloro - CONICET, Buenos Aires, Argentina

⁴Hôpital Européen Georges Pompidou, Paris, France

⁵Hôpital Européen Georges Pompidou, Paris, France

⁶Laboratoire d'Imagerie Biomédicale, Paris, France

⁷IMETTYB - Universidad Favaloro - CONICET, Buenos Aires, France

ABSTRACT

Purpose: Referral to surgery in thoracic aortic aneurysms (TAA) is based on maximal diameter (Dmax) measured from imaging, which is known to have a high diagnosis failure rate. In addition to geometry, 4D flow MRI provides a comprehensive time-resolved flow imaging. Thus, our aim was to evaluate the ability of 4D flow MRI-derived quantitative flow indices to characterize TAA.

Methods: We studied 20 patients with TAA and tricuspid valve (TAVd, Dmax = 43 ± 5 mm, Age = 66 ± 14 years) and 56 healthy controls (YC: 30 subjects, Age = 36 ± 9 years ≤ 50 years, OC: 26 subjects, Age = 65 ± 9 years > 50 years). All underwent 4D flow MRI. After aortic segmentation, ascending aorta (AA) backward flow volume (VBF) was calculated in addition to maximal velocity jet angle and eccentricity (Ecc). Receiver operating characteristic analysis was performed to assess the ability of flow indices to characterize AA = dilation.

Results: While AA Dmax was 1.4-fold higher in TAVd than OC, VBF increased by 6.5 folds and Ecc and Angle varied by 1.3 to 1.7 folds between the two groups. Moreover, VBF changed by 12.7 folds between the aneurysmal AA as compared to TAVd descending aorta. Finally, VBF increased consistently with age in all controls and was able to detect AA dilation with a 0.98 accuracy.

Conclusion: AA backward flow quantified from 4D flow MRI outperformed the previously described indices such as flow eccentricity and angle in the characterization of thoracic aortic aneurysms.

© 2019 Association for Research into Arterial Structure and Physiology. Publishing services by Atlantis Press International B.V. This is an open access article distributed under the CC BY-NC 4.0 license (<http://creativecommons.org/licenses/by-nc/4.0/>).