Figure 2. Schematic diagram of the timer circuit. Power is supplied by four 1.5-V AA alkaline batteries (Mallory MN1500 or equivalent). Average current drain: 50 mA.

place. Due to its proximity to the film, the blip it exposes is about the same size as the LED itself. Connecting leads to the LED can be either small gauge round wire or ribbon wire. The latter may be preferable if the leads are to pass around the edge of the door to the camera's film compartment. For some cameras, however, it may be more convenient to drill a hole, or modify an existing one, and use a miniature phone jack connection between the timer output leads and the LED input leads.

A schematic circuit for pulsing the LED is shown in Figure 2. In addition to the switch for the power supply, a switch is needed for the two oscillator ICs (No. 74121) to prevent occasional "lock-ups." Adjustment of the pulse rate and pulse duration is made through the potentiometers for the two oscillator units. Since pulse frequency and duration tend to interact in this design, adjustment must be made by trial and error, using an oscilloscope, until the desired values are obtained. Experience has shown that a pulse duration on the order of 5-10 msec provides acceptable exposure of Kodachrome-II, Plus-X, and Tri-X movie films. A pulse rate of 2/sec has been found satisfactory for filming at speeds from 18 to 70 fps, but a faster frequency might be preferable if high filming speeds are to be used exclusively. The precision of the settings seems to be quite stable, not varying detectably over a minimal period of 5 months.

REFERENCES

- Davis, J. M. Socially induced flight reactions in pigeons. *Animal Behaviour*, in press.
- Larsson, K., & Komisaruk, B. R. Abnormally fast vibrissae movements induced by tetrabenazine in rats. *Psychopharmacologia (Berl.)*, 1972, 23, 300-304.
- Rothschild, M., Schlein, Y., Parker, K., Neville, C., & Sternberg, S. The flying leap of the flea. *Scientific American*, 1973, 229, 92-100.
- Welker, W. I. Analysis of sniffing of the albino rat. *Behaviour*, 1964, 22, 223-244.

ERRATUM

Johnson, J. H., Cole, E. B., & Williams, T. A. PROSE: A simple user-oriented program for computer constructed narratives. *Behavior Research Methods & Instrumentation*, 1975, 7 (3), 309-310. The part of the paper containing Page 10 was inadvertently omitted in the printing of the issue. It is now produced below. We apologize for the error.

310 JOHNSON, COLE, AND WILLIAMS

The patient is a 24 year-old single male who was seen for evaluation on April 1, 1974 by Barbara Hoff. At the time of interview his appearance and behavior were not appropriate to the situation. Posturing, silliness, and agitation were noted. The patient was oriented for time, place, person. He was not cooperative during the evaluation.

Figure 5. Output report; observations.

Several other features should be mentioned. Text insertion points need not be in numerical order or in the same order as the list of insertion points in the narrative. Text insertion points must have an associated number that is greater than 0. A 0 identifying number is considered a dummy insertion point and is skipped by PROSE. A dummy insertion point is useful in

delimiting text within the scope of a Skip command. Using Skip, the user can conditionally include or exclude text.

Availability. A program listing and documentation developed by Mr. Cook are available without charge from James H. Johnson, Psychiatric Service, VA Hospital, Salt Lake City, Utah 84113.

REFERENCES

- Craig, L., Golenzer, F., & Laska, E. Computer constructed narratives. In N. S. Kline and E. Laska (Eds.), *Computers and electronic devices in psychiatry*. New York: Grune and Stratton, 1968, 59-80.
- Lanyon, R. I. Technological approach to the improvement of decision making in mental health services. *Journal of Consulting Clinical Psychology*, 1972, 39, 43-48.
- Laska, E., Morrill, D., Kline, N. S., et al. SCRIBE—a method for producing automated narrative psychiatric case histories. *American Journal of Psychiatry*, 1967, 124, 120-121.