- LEUKEL, F., & QUINTON, E. Carbon dioxide effects on acquisition and extinction of avoidance behavior. *Journal of Comparative and Physiological Psychology*. 1964, 57, 267-270.
- MURKLAND, R. T. An examination of the effects of magnesium pemoline on Sidman-avoidance and T-maze acquisition and memory. Master's thesis, Memphis State University Library, 1967.
- QUINTON, F. E. Retrograde amnesia induced by carbon dioxide inhalation. *Psychonomic Science*, 1966, 5, 417-418.

- WEINSTEIN, S. A. The effect of hypoxia upon learned escape from carbon-dioxide. *Psychonomic Science*, 1966, 6, 91-92.
- WILLIAMS, C. M. Continuous anesthesia for insects. *Science*, 1946, 103, 57.

NOTE

1. We are indebted to Mr. Robert Murkland for providing pilot data which led to the present investigation.

A wide range timing circuit with two independent controls

NORMAN T. WELFORD, MEDICAL ELECTRONICS LABORATORY, UNIVERSITY OF TEXAS MEDICAL BRANCH, Galveston, Texas 77550

The timer described in this note is an example of how a small increase in engineering complexity can lead to a marked increase in user convenience and overall system accuracy.

The timer was designed for an experiment by Owens (1969) where the S dropped marbles into a box, competing against a machine dropping marbles into another box. The timing circuit had to produce pulses at a base rate which could be set for each individual S. Each pulse caused the machine to drop one marble. During the experiment, trials were run at this base rate and at fixed percentages above and below the S's base rate. While this can be done with any timer that has a calibrated dial, such an apparently simple system gives considerable work for the E. It also suffers from the errors of any system where repeatability depends on the accuracy with which an E can reset a dial. With only a small increase in engineering complexity, these difficulties are overcome, resulting in a device that is ergonomically sound, and has worked without trouble for some 200 Ss in Owens's experiment.

The circuit is shown in Fig. 1. It is based on a recently available device, the GE programmable unijunction transistor, D13T2, and circuits for the same, described by Spofford (1967).

The D13T2 used in this oscillator has the advantage over the more common unijunction transistor of providing two independent places in the circuit where, by simple resistive potentiometers, the time of the oscillator may be varied.

Q1, C1, R1-R12, and D1 form a free-running relaxation oscillator. The base rate is determined by the setting of R1 in conjunction with R2 and C1. The fixed percentage increase and decrease from base rate is determined by the potential at R4. This potential is dependent on the setting of Switch S1. In the mid-position (base rate), R4 is connected to a fixed voltage determined by R8 and R9. In the up position (faster rate), the voltage at R4 is lower and determined by the setting of R6 and the values of R5 and R7. In the down position (slower rate), the potential at R4 is higher, being determined in a like manner by R10, R11, and R12.

The percentage increase or decrease of the oscillator frequency obtained from the three positions of the fast-normal-slow switch, S1, is independent of the setting of the base rate control, R1.

The oscillator is coupled via C2 to a one-shot circuit, Q2 and Q3 and associated components, that drives the solenoid, L1, through an inverter, Q5. An operating pulse of 40 msec was delivered to the solenoid at each cycle of the oscillator, and this

Fig. 1. Circuit diagram and parts list for timing circuit: R1—1M variable; R2—330K; R3—100 ohms; R4—1M; R5, 8, 9, 12, and 14—4.7K; R6 and 11—2.5K variable; R7, 10, 22, and 23—2.2K; R13—33K; R15, 19, and 21—6.8K; R16, 17, 18, and 20—1K; C1—2 mfd mylar or better; C2—0.02 mfd mylar; C3—6 mfd 15 V electrolytic; D1, 2, 3, and 4—1N4154 or similar; L1—500 ohms solenoid; Q1—D13T2; Q2, 3, and 4—2N1305 or similar; Q5—2N3440 or similar; S1—3-position single-pole switch. All fixed resistors $\frac{1}{2}$ W $\pm 10\%$ carbon composition, variable resistors. Ohmite Type AB or equal. For greater stability use metal film, deposited carbon, cermet, or wirewound precision resistors, and a polycarbonate or polystyrene capacitor for C1.

allowed one marble to be dropped by the apparatus. The power supply is unregulated for the solenoid supply and Zener-diode regulated for the 12-V supply.

The oscillator timing stability depends primarily on the quality of C1 and R1 and R2. For C2, a mylar capacitor is suitable as it has low leakage and low temperature coefficient. Polystyrene and polycarbonate capacitors are better but more expensive. The resistors can all be carbon composition for most applications. Precision metal film, deposited carbon, or cermet resistors all have lower temperature coefficients and higher long-term stability. The timing stability of the oscillator is almost independent of the supply voltage (a 10% charge in supply voltage gives less than $\frac{1}{4}\%$ charge in timing). Using a mylar capacitor for C1 and carbon resistors in a normal room temperature, the short-term stability can be expected to be within $\pm 1\%$.

REFERENCES

- OWENS, K. L. Competition in children as a function of age, race, sex, and socio-economic status. Unpublished doctoral dissertation, Texas Technological College, 1969.
- SPOFFORD, W. J., JR. The D13T-A programmable unijunction transistor. G.E. Application Note 90-70, 1967.